

Surrogate Summary

SW-846

SDG No.: P3467

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P3467-01	919-J-WS-080224	2-Methylnaphthalene-d10	0.4	0.36	90		30 (30)	150 (150)
		Fluoranthene-d10	0.4	0.46	115		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.34	85		30 (11)	130 (175)
		2-Fluorobiphenyl	0.4	0.36	91		30 (10)	130 (175)
		Terphenyl-d14	0.4	0.55	138	*	30 (54)	130 (171)
PB162490BL	PB162490BL	2-Methylnaphthalene-d10	0.4	0.38	95		30 (30)	150 (150)
		Fluoranthene-d10	0.4	0.38	94		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.40	100		30 (11)	130 (175)
		2-Fluorobiphenyl	0.4	0.17	43		30 (10)	130 (175)
		Terphenyl-d14	0.4	0.43	107		30 (54)	130 (171)
PB162490BS	PB162490BS	2-Methylnaphthalene-d10	0.4	0.42	104		30 (30)	150 (150)
		Fluoranthene-d10	0.4	0.38	95		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.35	86		30 (11)	130 (175)
		2-Fluorobiphenyl	0.4	0.40	99		30 (10)	130 (175)
		Terphenyl-d14	0.4	0.41	102		30 (54)	130 (171)
PB162490BSD	PB162490BSD	2-Methylnaphthalene-d10	0.4	0.37	92		30 (30)	150 (150)
		Fluoranthene-d10	0.4	0.39	98		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.35	87		30 (11)	130 (175)
		2-Fluorobiphenyl	0.4	0.39	98		30 (10)	130 (175)
		Terphenyl-d14	0.4	0.41	101		30 (54)	130 (171)