

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P3467

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270-Modified

DataFile: BN033272.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB162490BS	Naphthalene	0.4	0.40	ug/L	100				70 (67)	130 (120)	
	2-Methylnaphthalene	0.4	0.33	ug/L	83				70 (50)	130 (122)	
	Acenaphthylene	0.4	0.39	ug/L	98				70 (60)	130 (119)	
	Acenaphthene	0.4	0.41	ug/L	103				70 (65)	130 (119)	
	Fluorene	0.4	0.41	ug/L	103				70 (58)	130 (122)	
	Phenanthrene	0.4	0.37	ug/L	93				70 (65)	130 (117)	
	Anthracene	0.4	0.33	ug/L	83				70 (57)	130 (118)	
	Fluoranthene	0.4	0.37	ug/L	93				70 (53)	130 (126)	
	Pyrene	0.4	0.41	ug/L	103				70 (54)	130 (124)	
	Benzo(a)anthracene	0.4	0.31	ug/L	78				70 (54)	130 (130)	
	Chrysene	0.4	0.47	ug/L	117				70 (64)	130 (126)	
	Benzo(b)fluoranthene	0.4	0.36	ug/L	90				70 (65)	130 (121)	
	Benzo(k)fluoranthene	0.4	0.40	ug/L	100				70 (72)	130 (119)	
	Benzo(a)pyrene	0.4	0.39	ug/L	98				70 (68)	130 (120)	
	Indeno(1,2,3-cd)pyrene	0.4	0.36	ug/L	90				70 (70)	130 (127)	
	Dibenz(a,h)anthracene	0.4	0.35	ug/L	88				70 (65)	130 (121)	
	Benzo(g,h,i)perylene	0.4	0.37	ug/L	93				70 (76)	130 (117)	
	1,4-Dioxane	0.4	0.40	ug/L	100				20 (60)	160 (132)	