

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3467 SAS No.: P3467 SDG NO.: P3467

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 08/06/2024

Lab File ID: BN033270.D Time Analyzed: 13:04

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	699	7.553	2661	10.33	1835	14.14
UPPER LIMIT	1398	8.053	5322	10.829	3670	14.637
LOWER LIMIT	349.5	7.053	1330.5	9.829	917.5	13.637
EPA SAMPLE NO.						
01 PB162490BL	736	7.69	2380	10.53	1701	14.22
02 PB162490BS	693	7.56	2705	10.34	1884	14.14
03 PB162490BSD	699	7.56	2758	10.34	1926	14.14
04 919-J-WS-080224	968	7.55	3824	10.29	2488	14.11

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.