

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3467 SAS No.: P3467 SDG NO.: P3467

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/10/2024

Lab File ID: BF138901.D Time Analyzed: 10:41

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	63234	6.845	258541	8.13	146261	9.88
UPPER LIMIT	126468	7.345	517082	8.628	292522	10.38
LOWER LIMIT	31617	6.345	129271	7.628	73130.5	9.38
EPA SAMPLE NO.						
01 919-J-WS-080224	37654	6.84	147641	8.12	79728	9.87

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.