

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3467 SAS No.: P3467 SDG NO.: P3467

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/09/2024

Lab File ID: BF138879.D Time Analyzed: 09:48

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	254940	11.363	120643	14.004	111360	15.462
UPPER LIMIT	509880	11.863	241286	14.504	222720	15.962
LOWER LIMIT	127470	10.863	60321.5	13.504	55680	14.962
EPA SAMPLE NO.						
01 PB162489BL	189104	11.36	112088	14.00	75976	15.47
02 PB162489BS	232367	11.36	113862	14.00	96864	15.46

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.