

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF073024

Review By	yogesh	Review On	7/31/2024 6:39:45 AM
Supervise By	mohammad	Supervise On	7/31/2024 8:08:34 AM
SubDirectory	BF073024	HP Acquire Method	BNA_F
		HP Processing Method	bf073024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12036 10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF138679.D	30 Jul 2024 12:24		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF138680.D	30 Jul 2024 12:54		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF138681.D	30 Jul 2024 13:25	Compound#32,41,54,56,65,70 removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF138682.D	30 Jul 2024 13:56		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BF138683.D	30 Jul 2024 14:25	Comopund#41 Kept on LR	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF138684.D	30 Jul 2024 14:56	The calibration is good for 8270 DOD & 625.1	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF138685.D	30 Jul 2024 15:27		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF138686.D	30 Jul 2024 15:58	Compound#9 removed from 60 ppm	RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF138687.D	30 Jul 2024 16:29	Compound#9,77 removed from 80 ppm	RC/JU	Ok
10	SSTDICV040	ICVBF073024	BF138688.D	30 Jul 2024 17:55		RC/JU	Ok
11	PB162188BL	PB162188BL	BF138689.D	30 Jul 2024 18:26		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF080924

Review By	yogesh	Review On	8/10/2024 12:55:52 AM
Supervise By	mohammad	Supervise On	8/12/2024 1:39:31 AM
SubDirectory	BF080924	HP Acquire Method	BNA_F
		HP Processing Method	bf073024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12039,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF138878.D	09 Aug 2024 09:17		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF138879.D	09 Aug 2024 09:48		RC/JU	Ok,M
3	PB162421BL	PB162421BL	BF138880.D	09 Aug 2024 10:19		RC/JU	Ok
4	PB162358BL	PB162358BL	BF138881.D	09 Aug 2024 10:50	Surrogate Fail	RC/JU	Not Ok
5	PB162358BS	PB162358BS	BF138882.D	09 Aug 2024 11:21		RC/JU	Ok,M
6	PB162423BL	PB162423BL	BF138883.D	09 Aug 2024 11:52		RC/JU	Ok
7	PB162423BS	PB162423BS	BF138884.D	09 Aug 2024 12:23		RC/JU	Ok,M
8	PB162489BL	PB162489BL	BF138885.D	09 Aug 2024 12:55		RC/JU	Ok
9	PB162489BS	PB162489BS	BF138886.D	09 Aug 2024 13:25		RC/JU	Ok,M
10	P3464-02	TOPSOIL	BF138887.D	09 Aug 2024 14:10		RC/JU	Ok
11	P3468-01	IDW-S1-01	BF138888.D	09 Aug 2024 14:59		RC/JU	Ok
12	P3468-02	MLS-15-682-697	BF138889.D	09 Aug 2024 15:31		RC/JU	Ok
13	P3469-01	MLS-15-513-528	BF138890.D	09 Aug 2024 16:02		RC/JU	Ok
14	P3469-02	MLS-15-625-640	BF138891.D	09 Aug 2024 16:33		RC/JU	Ok
15	P3460-01	72-11929	BF138892.D	09 Aug 2024 17:04		RC/JU	Ok
16	P3460-01MS	72-11929MS	BF138893.D	09 Aug 2024 17:35		RC/JU	Ok,M
17	P3460-01MSD	72-11929MSD	BF138894.D	09 Aug 2024 18:05		RC/JU	Ok,M
18	P3507-01	EO-03-080724	BF138895.D	09 Aug 2024 18:35	Internal Standard Fail	RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF080924

Review By	yogesh	Review On	8/10/2024 12:55:52 AM		
Supervise By	mohammad	Supervise On	8/12/2024 1:39:31 AM		
SubDirectory	BF080924	HP Acquire Method	BNA_F	HP Processing Method	bf073024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12039,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	P3509-01	OR-02-080724	BF138896.D	09 Aug 2024 19:06	Internal Standard Fail	RC/JU	Ok,M
20	P3509-01MS	OR-02-080724MS	BF138897.D	09 Aug 2024 19:36	Internal Standard Fail	RC/JU	Ok,M
21	P3509-01MSD	OR-02-080724MSD	BF138898.D	09 Aug 2024 20:06	Internal Standard Fail	RC/JU	Ok,M
22	P3394-02	B-105-SB02	BF138899.D	09 Aug 2024 20:36	Internal Standard Fail	RC/JU	ReRun

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF081024

Review By	Jagrut	Review On	8/12/2024 5:10:00 PM
Supervise By	mohammad	Supervise On	8/13/2024 5:17:21 AM
SubDirectory	BF081024	HP Acquire Method	BNA_F
		HP Processing Method	bf073024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12039,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF138900.D	10 Aug 2024 10:12		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF138901.D	10 Aug 2024 10:41		RC/JU	Ok,M
3	PB162577TB	PB162577TB	BF138902.D	10 Aug 2024 11:11		RC/JU	Ok
4	P3479-03	TS09	BF138903.D	10 Aug 2024 11:46		RC/JU	Ok
5	P3479-05	TS10	BF138904.D	10 Aug 2024 12:16		RC/JU	Ok
6	P3479-01	TS08	BF138905.D	10 Aug 2024 12:46		RC/JU	Ok
7	P3479-07	TS11	BF138906.D	10 Aug 2024 13:16		RC/JU	Ok
8	P3464-02MS	TOPSOILMS	BF138907.D	10 Aug 2024 13:47	Internal Standard Fail	RC/JU	Ok,M
9	P3464-02MSD	TOPSOILMSD	BF138908.D	10 Aug 2024 14:17	Internal Standard Fail,	RC/JU	Ok,M
10	P3415-04MS	MLS-15-70-85MS	BF138909.D	10 Aug 2024 14:48		RC/JU	Ok
11	P3415-05MSD	MLS-15-70-85MSD	BF138910.D	10 Aug 2024 15:19		RC/JU	Ok
12	P3476-01	MUL04	BF138911.D	10 Aug 2024 15:49		RC/JU	Ok
13	P3466-04	932-K1-WP0-0.25-0802	BF138912.D	10 Aug 2024 16:20		RC/JU	Ok
14	P3466-05	919-J-WP0-0.25-08022	BF138913.D	10 Aug 2024 16:51		RC/JU	Ok
15	P3476-07	MUL07	BF138914.D	10 Aug 2024 17:21	Internal Standard Fail	RC/JU	ReRun
16	P3457-01	924-K1-WS-080224	BF138915.D	10 Aug 2024 17:52		RC/JU	Ok
17	P3457-02	932-K1-WS-080224	BF138916.D	10 Aug 2024 18:22		RC/JU	Ok
18	P3466-01	924-K1-WP0-0.25-0802	BF138917.D	10 Aug 2024 18:53		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF081024

Review By	Jagrut	Review On	8/12/2024 5:10:00 PM		
Supervise By	mohammad	Supervise On	8/13/2024 5:17:21 AM		
SubDirectory	BF081024	HP Acquire Method	BNA_F	HP Processing Method	bf073024
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557				
CCC	SP6553				
Internal Standard/PEM	S12039,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P3466-02MS	924-K1-WP0-0.25-0802	BF138918.D	10 Aug 2024 19:23	MSD not ok	RC/JU	Not Ok
20	P3466-03MSD	924-K1-WP0-0.25-0802	BF138919.D	10 Aug 2024 19:54	Internal Standard Fail	RC/JU	Not Ok
21	P3476-03	MUL05	BF138920.D	10 Aug 2024 20:24		RC/JU	Ok
22	P3476-05	MUL06	BF138921.D	10 Aug 2024 20:54	Internal Standard Fail	RC/JU	ReRun
23	P3467-01	919-J-WS-080224	BF138922.D	10 Aug 2024 21:25		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF081424

Review By	Jagrut	Review On	8/15/2024 10:26:10 AM
Supervise By	mohammad	Supervise On	8/16/2024 3:02:17 AM
SubDirectory	BF081424	HP Acquire Method	BNA_F
		HP Processing Method	bf073024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF138987.D	14 Aug 2024 11:33		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF138988.D	14 Aug 2024 12:03		RC/JU	Ok,M
3	PB162577BL	PB162577BL	BF138989.D	14 Aug 2024 12:32		RC/JU	Ok
4	PB162577BS	PB162577BS	BF138990.D	14 Aug 2024 13:02		RC/JU	Ok,M
5	P3474-01	WC-A1-COMP	BF138991.D	14 Aug 2024 13:37		RC/JU	Ok
6	P3474-10	WC-B2-COMP	BF138992.D	14 Aug 2024 14:08		RC/JU	Ok
7	P3512-05	931-K1-SD-0.5-1.0-080	BF138993.D	14 Aug 2024 14:38		RC/JU	Ok
8	P3512-06	931-K1-SD-0.5-1.0-080	BF138994.D	14 Aug 2024 15:09		RC/JU	Ok
9	P3543-01	ARS20-0019	BF138995.D	14 Aug 2024 15:40	Surrogate Fail	RC/JU	Ok
10	P3548-01	72-12017-72-11934	BF138996.D	14 Aug 2024 16:10		RC/JU	Ok
11	P3548-03	RBR200036	BF138997.D	14 Aug 2024 16:42		RC/JU	Ok,M
12	P3548-06	343	BF138998.D	14 Aug 2024 17:12	Internal standard fail	RC/JU	ReRun
13	P3548-12	IN-SITU-SOIL	BF138999.D	14 Aug 2024 17:43	Internal standard fail	RC/JU	ReRun
14	P3548-09	289	BF139000.D	14 Aug 2024 18:14		RC/JU	Ok
15	P3549-01	SU-04-080924	BF139001.D	14 Aug 2024 18:46	Internal standard fail	RC/JU	Ok,M
16	P3476-07RE	MUL07RE	BF139002.D	14 Aug 2024 19:16	Internal Standard Fail	RC/JU	Confirms
17	P3476-05RE	MUL06RE	BF139003.D	14 Aug 2024 19:46	Fax is already given	RC/JU	Confirms
18	P3466-02MS	924-K1-WP0-0.25-0802	BF139004.D	14 Aug 2024 20:16		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF081424

Review By	Jagrut	Review On	8/15/2024 10:26:10 AM
Supervise By	mohammad	Supervise On	8/16/2024 3:02:17 AM
SubDirectory	BF081424	HP Acquire Method	BNA_F
		HP Processing Method	bf073024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

19	P3466-03MSD	924-K1-WP0-0.25-0802	BF139005.D	14 Aug 2024 20:47		RC/JU	Ok
20	P3475-01	S-866-N-S0-0-0.5-0805	BF139006.D	14 Aug 2024 21:17	Internal Standard Fail	RC/JU	ReRun
21	P3475-02	S-866-N-S0-1.0-1.5-0806	BF139007.D	14 Aug 2024 21:47	Internal Standard Fail	RC/JU	Ok,M
22	P3475-03MS	S-866-N-S0-1.0-1.5-0806	BF139008.D	14 Aug 2024 22:17	Internal Standard Fail,	RC/JU	Ok,M
23	P3475-04MSD	S-866-N-S0-1.0-1.5-0806	BF139009.D	14 Aug 2024 22:47	Internal Standard Fail,	RC/JU	Ok,M
24	P3526-01	SP-1-COMP	BF139010.D	14 Aug 2024 23:18	Internal Standard Fail	RC/JU	ReRun

M : Manual Integration