

Surrogate Summary

SW-846

SDG No.: P3467

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P3466-02MS	924-K1-WP0-0.25-080224MS	2-Fluorophenol	150	77.1	51		15 (10)	110 (139)
		Phenol-d6	150	45.0	30		15 (10)	110 (134)
		Nitrobenzene-d5	100	104	104		30 (49)	130 (133)
		2-Fluorobiphenyl	100	103	103		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	157	105		15 (44)	110 (137)
		Terphenyl-d14	100	115	115		30 (48)	130 (125)
P3466-03MSD	924-K1-WP0-0.25-080224MSD	2-Fluorophenol	150	75.1	50		15 (10)	110 (139)
		Phenol-d6	150	43.2	29		15 (10)	110 (134)
		Nitrobenzene-d5	100	104	104		30 (49)	130 (133)
		2-Fluorobiphenyl	100	100	100		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	156	104		15 (44)	110 (137)
		Terphenyl-d14	100	113	113		30 (48)	130 (125)
P3467-01	919-J-WS-080224	2-Fluorophenol	150	74.5	50		15 (10)	110 (139)
		Phenol-d6	150	47.2	31		15 (10)	110 (134)
		Nitrobenzene-d5	100	101	101		30 (49)	130 (133)
		2-Fluorobiphenyl	100	99.3	99		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	148	99		15 (32)	110 (145)
		Terphenyl-d14	100	99.1	99		30 (36)	130 (145)
PB162489BL	PB162489BL	2-Fluorophenol	150	166	110		15 (10)	110 (139)
		Phenol-d6	150	167	111	*	15 (10)	110 (134)
		Nitrobenzene-d5	100	112	112		30 (49)	130 (133)
		2-Fluorobiphenyl	100	110	110		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	182	122	*	15 (32)	110 (145)
		Terphenyl-d14	100	136	136	*	30 (36)	130 (145)
PB162489BS	PB162489BS	2-Fluorophenol	150	130	87		15 (10)	110 (139)
		Phenol-d6	150	129	86		15 (10)	110 (134)
		Nitrobenzene-d5	100	85.0	85		30 (49)	130 (133)
		2-Fluorobiphenyl	100	86.9	87		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	139	92		15 (32)	110 (145)
		Terphenyl-d14	100	116	116		30 (36)	130 (145)