

5B

SEMIVOLATILE ORGANIC INSTRUMENT PERFORMANCE CHECK
DECAFLUOROTRIPHENYLPHOSPHINE (DFTPP)

Lab Name: CHEMTECH

Contract: JAC005

Lab Code: CHEM

SAS No.: P3467

SDG NO.: P3467

Lab File ID: BF138900.D

DFTPP Injection Date: 08/10/2024

Instrument ID: BNA_F

DFTPP Injection Time: 10:12

m/e	ION ABUNDANCE CRITERIA	% RELATIVE ABUNDANCE
51	10.0 - 80.0% of mass 198	47.8
68	Less than 2.0% of mass 69	0.3 (0.8) 1
69	Mass 69 relative abundance	40.9
70	Less than 2.0% of mass 69	0.3 (0.6) 1
127	10.0 - 80.0% of mass 198	48.7
197	Less than 2.0% of mass 198	0.0
198	Base Peak, 100% relative abundance	100
199	5.0 to 9.0% of mass 198	6.8
275	10.0 - 60.0% of mass 198	25.6
365	Greater than 1% of mass 198	3.1
441	Present, but less than mass 443	13.1
442	Greater than 50% of mass 198	81.3
443	15.0 - 24.0% of mass 442	15.6 (19.2) 2

1-Value is % mass 69

2-Value is % mass 442

THIS CHECK APPLIES TO THE FOLLOWING SAMPLES, MS, MSD, BLANKS, AND STANDARDS:

EPA SAMPLE NO.	LAB SAMPLE ID	LAB FILE ID	DATE ANALYZED	TIME ANALYZED
SSTDCCC040	SSTDCCC040	BF138901.D	08/10/2024	10:41
919-J-WS-080224	P3467-01	BF138922.D	08/10/2024	21:25