

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: JAC005  
Lab Code: CHEM Case No.: P3467 SAS No.: P3467 SDG No.: P3467  
Instrument ID: MSVOA\_N Calibration Date(s): 08/07/2024 08/07/2024  
Heated Purge: (Y/N) N Calibration Time(s): 10:33 12:34  
GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID: RRF100 = VN083135.D RRF050 = VN083136.D RRF020 = VN083137.D<br>RRF010 = VN083138.D RRF005 = VN083139.D RRF001 = VN083140.D |        |        |        |        |        |        |       |       |
|-----------------------------------------------------------------------------------------------------------------------------------------|--------|--------|--------|--------|--------|--------|-------|-------|
| COMPOUND                                                                                                                                | RRF100 | RRF050 | RRF020 | RRF010 | RRF005 | RRF001 | RRF   | % RSD |
| Dichlorodifluoromethane                                                                                                                 | 0.548  | 0.571  | 0.566  | 0.579  | 0.597  | 0.542  | 0.567 | 3.6   |
| Chloromethane                                                                                                                           | 0.547  | 0.591  | 0.584  | 0.589  | 0.595  | 0.576  | 0.581 | 3.1   |
| Vinyl Chloride                                                                                                                          | 0.562  | 0.598  | 0.588  | 0.612  | 0.586  | 0.607  | 0.592 | 3     |
| Bromomethane                                                                                                                            | 0.313  | 0.357  | 0.370  | 0.385  | 0.413  |        | 0.368 | 10.1  |
| 1,1,2-Trichlorotrifluoroethane                                                                                                          | 0.524  | 0.548  | 0.548  | 0.550  | 0.558  | 0.510  | 0.539 | 3.4   |
| Acetone                                                                                                                                 | 0.302  | 0.286  | 0.260  | 0.284  | 0.274  | 0.265  | 0.278 | 5.5   |
| Carbon Disulfide                                                                                                                        | 1.466  | 1.560  | 1.596  | 1.599  | 1.630  | 1.879  | 1.622 | 8.5   |
| Methyl tert-butyl Ether                                                                                                                 | 1.927  | 2.028  | 2.028  | 2.015  | 2.053  | 1.953  | 2.001 | 2.5   |
| Methylene Chloride                                                                                                                      | 0.569  | 0.599  | 0.619  | 0.614  | 0.640  | 0.805  | 0.641 | 13    |
| trans-1,2-Dichloroethene                                                                                                                | 0.533  | 0.574  | 0.577  | 0.586  | 0.568  | 0.599  | 0.573 | 3.9   |
| Cyclohexane                                                                                                                             | 0.933  | 0.982  | 1.010  | 1.110  | 1.238  |        | 1.055 | 11.5  |
| 2-Butanone                                                                                                                              | 0.418  | 0.427  | 0.424  | 0.430  | 0.414  | 0.453  | 0.428 | 3.2   |
| Carbon Tetrachloride                                                                                                                    | 0.524  | 0.546  | 0.548  | 0.541  | 0.539  | 0.493  | 0.532 | 3.9   |
| cis-1,2-Dichloroethene                                                                                                                  | 0.658  | 0.689  | 0.691  | 0.701  | 0.687  | 0.721  | 0.691 | 3     |
| Chloroform                                                                                                                              | 1.087  | 1.137  | 1.110  | 1.132  | 1.122  | 1.102  | 1.115 | 1.7   |
| 1,1,1-Trichloroethane                                                                                                                   | 1.027  | 1.075  | 1.066  | 1.076  | 1.081  | 1.007  | 1.055 | 2.9   |
| Methylcyclohexane                                                                                                                       | 0.565  | 0.597  | 0.587  | 0.583  | 0.575  | 0.572  | 0.580 | 2     |
| Benzene                                                                                                                                 | 1.386  | 1.447  | 1.431  | 1.429  | 1.403  | 1.343  | 1.406 | 2.7   |
| 1,2-Dichloroethane                                                                                                                      | 0.494  | 0.520  | 0.514  | 0.524  | 0.524  | 0.498  | 0.512 | 2.6   |
| Trichloroethene                                                                                                                         | 0.331  | 0.344  | 0.343  | 0.341  | 0.340  | 0.310  | 0.335 | 3.9   |
| Bromodichloromethane                                                                                                                    | 0.521  | 0.539  | 0.540  | 0.533  | 0.541  | 0.544  | 0.537 | 1.6   |
| Toluene                                                                                                                                 | 0.892  | 0.926  | 0.905  | 0.900  | 0.885  | 0.824  | 0.889 | 3.9   |
| 1,1,2-Trichloroethane                                                                                                                   | 0.322  | 0.334  | 0.331  | 0.327  | 0.314  | 0.283  | 0.318 | 6     |
| Dibromochloromethane                                                                                                                    | 0.400  | 0.413  | 0.397  | 0.388  | 0.383  | 0.329  | 0.385 | 7.7   |
| Tetrachloroethene                                                                                                                       | 0.319  | 0.343  | 0.338  | 0.343  | 0.324  | 0.320  | 0.331 | 3.5   |
| Chlorobenzene                                                                                                                           | 1.070  | 1.142  | 1.101  | 1.153  | 1.094  | 1.069  | 1.105 | 3.2   |
| Ethyl Benzene                                                                                                                           | 1.967  | 2.075  | 2.061  | 2.070  | 2.032  | 1.957  | 2.027 | 2.6   |
| m/p-Xylenes                                                                                                                             | 0.747  | 0.787  | 0.771  | 0.785  | 0.759  | 0.706  | 0.759 | 4     |
| o-Xylene                                                                                                                                | 0.735  | 0.773  | 0.759  | 0.773  | 0.750  | 0.703  | 0.749 | 3.6   |
| Isopropylbenzene                                                                                                                        | 3.822  | 4.249  | 4.133  | 4.386  | 4.249  | 4.254  | 4.182 | 4.6   |

\* Compounds with required minimum RRF and maximum %RSD values.  
All other compounds must meet a minimum RRF of 0.010.  
RRF of 1,4-Dioxane = Value should be divide by 1000.

### VOLATILE ORGANICS INITIAL CALIBRATION DATA

Lab Name: CHEMTECH Contract: JAC005  
 Lab Code: CHEM Case No.: P3467 SAS No.: P3467 SDG No.: P3467  
 Instrument ID: MSVOA\_N Calibration Date(s): 08/07/2024 08/07/2024  
 Heated Purge: (Y/N) N Calibration Time(s): 10:33 12:34  
 GC Column: RXI-624 ID: 0.25 (mm)

| LAB FILE ID:          |        | RRF100 = VN083135.D |        | RRF050 = VN083136.D |        | RRF020 = VN083137.D |       |       |  |
|-----------------------|--------|---------------------|--------|---------------------|--------|---------------------|-------|-------|--|
|                       |        | RRF010 = VN083138.D |        | RRF005 = VN083139.D |        | RRF001 = VN083140.D |       |       |  |
| COMPOUND              | RRF100 | RRF050              | RRF020 | RRF010              | RRF005 | RRF001              | RRF   | % RSD |  |
| 1,4-Dichlorobenzene   | 1.655  | 1.783               | 1.728  | 1.751               | 1.795  | 1.862               | 1.762 | 4     |  |
| 1,2-Dichlorobenzene   | 1.571  | 1.704               | 1.664  | 1.739               | 1.731  | 1.741               | 1.692 | 3.9   |  |
| 1,2-Dichloroethane-d4 | 0.717  | 0.745               | 0.601  | 0.710               | 0.786  |                     | 0.712 | 9.7   |  |
| Dibromofluoromethane  | 0.311  | 0.332               | 0.264  | 0.307               | 0.346  |                     | 0.312 | 10    |  |
| Toluene-d8            | 1.201  | 1.261               | 0.985  | 1.133               | 1.240  |                     | 1.164 | 9.6   |  |
| 4-Bromofluorobenzene  | 0.473  | 0.492               | 0.390  | 0.432               | 0.483  |                     | 0.454 | 9.4   |  |

\* Compounds with required minimum RRF and maximum %RSD values.  
 All other compounds must meet a minimum RRF of 0.010.  
 RRF of 1,4-Dioxane = Value should be divide by 1000.