

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3475
Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3475 **SAS No.:** P3475
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
ICV01	Silver	255	250	102	90 - 110	P	08/08/2024	13:22	LB131909

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3475
Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3475 **SAS No.:** P3475
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
LLICV01	Silver	11.5	10.0	115	80 - 120	P	08/08/2024	13:32	LB131909

Metals

- 2a -

INITIAL AND CONTINUING CALIBRATION VERIFICATION

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3475
Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3475 **SAS No.:** P3475
Initial Calibration Source: EPA
Continuing Calibration Source: Inorganic Ventures

Sample ID	Analyte	Result ug/L	True Value	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CCV01	Silver	1250	1250	100	90 - 110	P	08/08/2024	14:04	LB131909
CCV02	Silver	1200	1250	96	90 - 110	P	08/08/2024	14:53	LB131909
CCV03	Silver	1250	1250	100	90 - 110	P	08/08/2024	15:49	LB131909
CCV04	Silver	1190	1250	96	90 - 110	P	08/08/2024	16:41	LB131909
CCV05	Silver	1200	1250	96	90 - 110	P	08/08/2024	17:33	LB131909
CCV06	Silver	1240	1250	99	90 - 110	P	08/08/2024	18:26	LB131909