



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: JACOBS Engineering Group, Inc.

SDG No.: P3475

Contract: JACO05

Lab Code: CHEM

Case No.: P3475

SAS No.: P3475

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Silver	12.0	10.0	120	40 - 160	P	08/08/2024	13:42	LB131909