

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3475
Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3475 **SAS No.:** P3475
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Silver	1.87			-10	10	08/08/2024	13:46	LB131909
ICSAB01	Silver	220	201	110	170	232	08/08/2024	13:51	LB131909