

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: P3475 SAS No.: P3475 SDG NO.: P3475

Continuing Calib Date: 08/06/2024 Initial Calibration Date(s): 07/29/2024 07/30/2024

Continuing Calib Time: 12:37 Initial Calibration Time(s): 18:07 01:54

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	5.60	5.60	5.50	5.70	0.00
Aroclor-1016-2	(2)	5.62	5.62	5.52	5.72	0.00
Aroclor-1016-3	(3)	5.69	5.69	5.59	5.79	0.00
Aroclor-1016-4	(4)	5.79	5.79	5.69	5.89	0.00
Aroclor-1016-5	(5)	6.08	6.08	5.98	6.18	0.00
Aroclor-1260-1	(1)	7.22	7.22	7.12	7.32	0.00
Aroclor-1260-2	(2)	7.48	7.47	7.37	7.57	-0.01
Aroclor-1260-3	(3)	7.84	7.84	7.74	7.94	0.00
Aroclor-1260-4	(4)	8.07	8.06	7.96	8.16	-0.01
Aroclor-1260-5	(5)	8.39	8.39	8.29	8.49	0.00
Tetrachloro-m-xylene		4.43	4.43	4.33	4.53	0.00
Decachlorobiphenyl		10.22	10.22	10.12	10.32	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: P3475 SAS No.: P3475 SDG NO.: P3475

Continuing Calib Date: 08/06/2024 Initial Calibration Date(s): 07/29/2024 07/30/2024

Continuing Calib Time: 12:37 Initial Calibration Time(s): 18:07 01:54

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW FROM TO		DIFF RT
Aroclor-1016-1	(1)	4.59	4.60	4.50	4.70	0.01
Aroclor-1016-2	(2)	4.61	4.61	4.51	4.71	0.00
Aroclor-1016-3	(3)	4.78	4.79	4.69	4.89	0.01
Aroclor-1016-4	(4)	4.83	4.83	4.73	4.93	0.01
Aroclor-1016-5	(5)	5.03	5.04	4.94	5.14	0.01
Aroclor-1260-1	(1)	6.05	6.05	5.95	6.15	0.00
Aroclor-1260-2	(2)	6.23	6.24	6.14	6.34	0.01
Aroclor-1260-3	(3)	6.38	6.39	6.29	6.49	0.01
Aroclor-1260-4	(4)	6.85	6.86	6.76	6.96	0.01
Aroclor-1260-5	(5)	7.09	7.10	7.00	7.20	0.01
Tetrachloro-m-xylene		3.54	3.54	3.44	3.64	0.00
Decachlorobiphenyl		8.43	8.44	8.34	8.54	0.01