

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: P3475 SAS No.: P3475 SDG NO.: P3475

Continuing Calib Date: 08/06/2024 Initial Calibration Date(s): 07/29/2024 07/30/2024

Continuing Calib Time: 13:12 Initial Calibration Time(s): 18:07 01:54

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1248-1 (1)	5.60	5.60	5.50	5.70	0.00
Aroclor-1248-2 (2)	5.88	5.87	5.77	5.97	-0.01
Aroclor-1248-3 (3)	6.08	6.08	5.98	6.18	0.00
Aroclor-1248-4 (4)	6.49	6.49	6.39	6.59	0.00
Aroclor-1248-5 (5)	6.53	6.53	6.43	6.63	0.00
Tetrachloro-m-xylene	4.43	4.43	4.33	4.53	0.00
Decachlorobiphenyl	10.22	10.22	10.12	10.32	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: P3475 SAS No.: P3475 SDG NO.: P3475

Continuing Calib Date: 08/06/2024 Initial Calibration Date(s): 07/29/2024 07/30/2024

Continuing Calib Time: 13:12 Initial Calibration Time(s): 18:07 01:54

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND		CCAL RT	AVG RT	RT WINDOW		DIFF RT
				FROM	TO	
Aroclor-1248-1	(1)	4.59	4.60	4.50	4.70	0.01
Aroclor-1248-2	(2)	4.83	4.83	4.73	4.93	0.01
Aroclor-1248-3	(3)	4.87	4.87	4.77	4.97	0.00
Aroclor-1248-4	(4)	5.03	5.04	4.94	5.14	0.01
Aroclor-1248-5	(5)	5.42	5.42	5.32	5.52	0.00
Tetrachloro-m-xylene		3.54	3.54	3.44	3.64	0.00
Decachlorobiphenyl		8.43	8.44	8.34	8.54	0.01