

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: P3475 SAS No.: P3475 SDG NO.: P3475

Continuing Calib Date: 08/06/2024 Initial Calibration Date(s): 07/29/2024 07/30/2024

Continuing Calib Time: 18:54 Initial Calibration Time(s): 18:07 01:54

GC Column: ZB-MR1 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	5.60	5.60	5.50	5.70	0.00
Aroclor-1242-2 (2)	5.63	5.62	5.52	5.72	-0.01
Aroclor-1242-3 (3)	5.69	5.68	5.58	5.78	-0.01
Aroclor-1242-4 (4)	5.79	5.78	5.68	5.88	-0.01
Aroclor-1242-5 (5)	6.53	6.53	6.43	6.63	0.00
Tetrachloro-m-xylene	4.43	4.43	4.33	4.53	0.00
Decachlorobiphenyl	10.22	10.22	10.12	10.32	0.00

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: P3475 SAS No.: P3475 SDG NO.: P3475

Continuing Calib Date: 08/06/2024 Initial Calibration Date(s): 07/29/2024 07/30/2024

Continuing Calib Time: 18:54 Initial Calibration Time(s): 18:07 01:54

GC Column: ZB-MR2 ID: 0.32 (mm)

COMPOUND	CCAL RT	AVG RT	RT WINDOW		DIFF RT
			FROM	TO	
Aroclor-1242-1 (1)	4.59	4.60	4.50	4.70	0.01
Aroclor-1242-2 (2)	4.61	4.61	4.51	4.71	0.00
Aroclor-1242-3 (3)	4.78	4.79	4.69	4.89	0.01
Aroclor-1242-4 (4)	4.87	4.87	4.77	4.97	0.00
Aroclor-1242-5 (5)	5.38	5.38	5.28	5.48	0.00
Tetrachloro-m-xylene	3.54	3.54	3.44	3.64	0.00
Decachlorobiphenyl	8.43	8.44	8.34	8.54	0.01