

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: P3475 SAS No.: P3475 SDG NO.: P3475

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/29/2024 07/29/2024

Client Sample No.: CCAL01 Date Analyzed: 08/06/2024

Lab Sample No.: AR1660CCC500 Data File : PO105223.D Time Analyzed: 12:37

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.602	5.500	5.700	530.420	500.000	6.1
Aroclor-1016-2	5.624	5.522	5.722	530.880	500.000	6.2
Aroclor-1016-3	5.687	5.585	5.785	519.380	500.000	3.9
Aroclor-1016-4	5.786	5.685	5.885	498.320	500.000	-0.3
Aroclor-1016-5	6.083	5.981	6.181	536.700	500.000	7.3
Aroclor-1260-1	7.218	7.116	7.316	482.650	500.000	-3.5
Aroclor-1260-2	7.477	7.374	7.574	473.520	500.000	-5.3
Aroclor-1260-3	7.839	7.737	7.937	439.460	500.000	-12.1
Aroclor-1260-4	8.067	7.964	8.164	467.430	500.000	-6.5
Aroclor-1260-5	8.391	8.287	8.487	454.050	500.000	-9.2
Decachlorobiphenyl	10.224	10.121	10.321	47.340	50.000	-5.3
Tetrachloro-m-xylene	4.431	4.329	4.529	54.010	50.000	8.0

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: P3475 SAS No.: P3475 SDG NO.: P3475

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/29/2024 07/29/2024

Client Sample No.: CCAL01 Date Analyzed: 08/06/2024

Lab Sample No.: AR1660CCC500 Data File : PO105223.D Time Analyzed: 12:37

COMPOUND	RT	RT WINDOW FROM TO		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
Aroclor-1016-1	4.593	4.496	4.696	513.270	500.000	2.7
Aroclor-1016-2	4.611	4.514	4.714	513.160	500.000	2.6
Aroclor-1016-3	4.783	4.686	4.886	513.670	500.000	2.7
Aroclor-1016-4	4.825	4.728	4.928	511.080	500.000	2.2
Aroclor-1016-5	5.033	4.937	5.137	513.420	500.000	2.7
Aroclor-1260-1	6.048	5.952	6.152	485.230	500.000	-3.0
Aroclor-1260-2	6.234	6.138	6.338	470.290	500.000	-5.9
Aroclor-1260-3	6.384	6.289	6.489	439.130	500.000	-12.2
Aroclor-1260-4	6.850	6.755	6.955	480.230	500.000	-4.0
Aroclor-1260-5	7.090	6.995	7.195	459.530	500.000	-8.1
Decachlorobiphenyl	8.431	8.337	8.537	46.030	50.000	-7.9
Tetrachloro-m-xylene	3.539	3.440	3.640	52.310	50.000	4.6