

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: P3475 SAS No.: P3475 SDG NO.: P3475

GC Column: ZB-MR1 ID: 0.32 (mm) Initi. Calib. Date(s): 07/29/2024 07/29/2024

Client Sample No.: CCAL02 Date Analyzed: 08/06/2024

Lab Sample No.: AR1660CCC500 Data File : PO105238.D Time Analyzed: 18:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	5.601	5.500	5.700	525.960	500.000	5.2
Aroclor-1016-2	5.624	5.522	5.722	534.610	500.000	6.9
Aroclor-1016-3	5.687	5.585	5.785	522.110	500.000	4.4
Aroclor-1016-4	5.786	5.685	5.885	499.040	500.000	-0.2
Aroclor-1016-5	6.083	5.981	6.181	536.420	500.000	7.3
Aroclor-1260-1	7.218	7.116	7.316	490.910	500.000	-1.8
Aroclor-1260-2	7.476	7.374	7.574	492.850	500.000	-1.4
Aroclor-1260-3	7.838	7.737	7.937	450.390	500.000	-9.9
Aroclor-1260-4	8.065	7.964	8.164	465.510	500.000	-6.9
Aroclor-1260-5	8.388	8.287	8.487	449.940	500.000	-10.0
Decachlorobiphenyl	10.220	10.121	10.321	48.050	50.000	-3.9
Tetrachloro-m-xylene	4.430	4.329	4.529	53.860	50.000	7.7

CALIBRATION VERIFICATION SUMMARY

Contract: JACO05

Lab Code: CHEM Case No.: P3475 SAS No.: P3475 SDG NO.: P3475

GC Column: ZB-MR2 ID: 0.32 (mm) Initi. Calib. Date(s): 07/29/2024 07/29/2024

Client Sample No.: CCAL02 Date Analyzed: 08/06/2024

Lab Sample No.: AR1660CCC500 Data File : PO105238.D Time Analyzed: 18:36

COMPOUND	RT	RT WINDOW		CALC AMOUNT(ng)	NOM AMOUNT(ng)	%D
		FROM	TO			
Aroclor-1016-1	4.594	4.496	4.696	519.960	500.000	4.0
Aroclor-1016-2	4.611	4.514	4.714	519.310	500.000	3.9
Aroclor-1016-3	4.784	4.686	4.886	525.090	500.000	5.0
Aroclor-1016-4	4.825	4.728	4.928	522.780	500.000	4.6
Aroclor-1016-5	5.034	4.937	5.137	533.400	500.000	6.7
Aroclor-1260-1	6.049	5.952	6.152	493.810	500.000	-1.2
Aroclor-1260-2	6.234	6.138	6.338	474.510	500.000	-5.1
Aroclor-1260-3	6.385	6.289	6.489	446.110	500.000	-10.8
Aroclor-1260-4	6.850	6.755	6.955	483.660	500.000	-3.3
Aroclor-1260-5	7.090	6.995	7.195	464.770	500.000	-7.0
Decachlorobiphenyl	8.430	8.337	8.537	47.640	50.000	-4.7
Tetrachloro-m-xylene	3.539	3.440	3.640	52.650	50.000	5.3