

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072924\
 Data File : P0105014.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 29 Jul 2024 21:00
 Operator : YP/AJ
 Sample : AR1242ICC250
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1242ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 30 05:19:21 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 30 05:19:13 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.429	3.540	190.5E6	70162767	25.464	25.409
2)	SA Decachlor...	10.220	8.435	140.7E6	44322098	25.821	26.305
Target Compounds							
16)	L4 AR-1242-1	5.599	4.596	52690567	20549378	263.308	263.830
17)	L4 AR-1242-2	5.622	4.614	72725879	27677668	258.476	258.278
18)	L4 AR-1242-3	5.684	4.786	45108031	14871613	263.182	260.238
19)	L4 AR-1242-4	5.784	4.868	40536122	15231210	271.940	264.660
20)	L4 AR-1242-5	6.525	5.381	46089663	19021004	278.707	260.125



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072924\
Data File : PO105014.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 29 Jul 2024 21:00
Operator : YP/AJ
Sample : AR1242ICC250
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 30 05:19:21 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 30 05:19:13 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.429 min
Delta R.T.: 0.000 min
Response: 190502769
Conc: 25.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.540 min
Delta R.T.: 0.000 min
Response: 70162767
Conc: 25.41 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.220 min
Delta R.T.: 0.000 min
Response: 140731241
Conc: 25.82 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.435 min
Delta R.T.: 0.000 min
Response: 44322098
Conc: 26.30 ng/ml



#16 AR-1242-1


R.T.: 5.599 min
Delta R.T.: 0.000 min
Response: 52690567
Conc: 263.31 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#16 AR-1242-1

R.T.: 4.596 min
Delta R.T.: 0.000 min
Response: 20549378
Conc: 263.83 ng/ml




#17 AR-1242-2

R.T.: 5.622 min
Delta R.T.: 0.000 min
Response: 72725879
Conc: 258.48 ng/ml



#17 AR-1242-2


R.T.: 4.614 min
Delta R.T.: 0.000 min
Response: 27677668
Conc: 258.28 ng/ml



#18 AR-1242-3


R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 45108031
Conc: 263.18 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250




#18 AR-1242-3

R.T.: 4.786 min
Delta R.T.: 0.000 min
Response: 14871613
Conc: 260.24 ng/ml




#19 AR-1242-4

R.T.: 5.784 min
Delta R.T.: 0.000 min
Response: 40536122
Conc: 271.94 ng/ml



#19 AR-1242-4


R.T.: 4.868 min
Delta R.T.: 0.000 min
Response: 15231210
Conc: 264.66 ng/ml



#20 AR-1242-5

R.T.: 6.525 min
Delta R.T.: 0.000 min
Response: 46089663
Conc: 278.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1242ICC250



#20 AR-1242-5

R.T.: 5.381 min
Delta R.T.: 0.000 min
Response: 19021004
Conc: 260.12 ng/ml