

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072924\
 Data File : P0105030.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jul 2024 01:37
 Operator : YP/AJ
 Sample : AR1268ICC250
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 AR1268ICC250

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 30 06:55:56 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 30 06:55:47 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.428	3.539	195.4E6	71991625	25.455	25.177
2)	SA Decachlor...	10.220	8.435	249.3E6	79602498	26.008	26.224
Target Compounds							
41)	L9 AR-1268-1	8.702	7.373	206.6E6	90279026	259.439	261.626
42)	L9 AR-1268-2	8.797	7.437	181.3E6	80058660	257.435	259.597
43)	L9 AR-1268-3	9.029	7.640	153.8E6	65039534	255.653	260.087
44)	L9 AR-1268-4	9.458	7.925	65650146	27556423	251.149	262.670
45)	L9 AR-1268-5	9.878	8.199	508.1E6	173.2E6	253.999	254.314



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072924\
Data File : P0105030.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jul 2024 01:37
Operator : YP/AJ
Sample : AR1268ICC250
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 30 06:55:56 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 30 06:55:47 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.428 min
Delta R.T.: 0.000 min
Response: 195445365
Conc: 25.46 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#1 Tetrachloro-m-xylene

R.T.: 3.539 min
Delta R.T.: 0.000 min
Response: 71991625
Conc: 25.18 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.220 min
Delta R.T.: 0.000 min
Response: 249284090
Conc: 26.01 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.435 min
Delta R.T.: 0.000 min
Response: 79602498
Conc: 26.22 ng/ml



#41 AR-1268-1


R.T.: 8.702 min
Delta R.T.: 0.000 min
Response: 206563031
Conc: 259.44 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#41 AR-1268-1

R.T.: 7.373 min
Delta R.T.: 0.000 min
Response: 90279026
Conc: 261.63 ng/ml




#42 AR-1268-2

R.T.: 8.797 min
Delta R.T.: 0.000 min
Response: 181299736
Conc: 257.44 ng/ml



#42 AR-1268-2


R.T.: 7.437 min
Delta R.T.: 0.000 min
Response: 80058660
Conc: 259.60 ng/ml



#43 AR-1268-3


R.T.: 9.029 min
Delta R.T.: 0.000 min
Response: 153785699
Conc: 255.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250




#43 AR-1268-3

R.T.: 7.640 min
Delta R.T.: 0.000 min
Response: 65039534
Conc: 260.09 ng/ml




#44 AR-1268-4

R.T.: 9.458 min
Delta R.T.: 0.000 min
Response: 65650146
Conc: 251.15 ng/ml



#44 AR-1268-4


R.T.: 7.925 min
Delta R.T.: 0.000 min
Response: 27556423
Conc: 262.67 ng/ml



#45 AR-1268-5

R.T.: 9.878 min
Delta R.T.: 0.000 min
Response: 508056578
Conc: 254.00 ng/ml

Instrument :
ECD_O
ClientSampleId :
AR1268ICC250



#45 AR-1268-5

R.T.: 8.199 min
Delta R.T.: 0.000 min
Response: 173244924
Conc: 254.31 ng/ml