

Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\P0072924\
 Data File : P0105033.D
 Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
 Acq On : 30 Jul 2024 02:29
 Operator : YP/AJ
 Sample : AR1242ICV500
 Misc :
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 ECD_O
 ClientSampleId :
 ICVPO072924AR1242

Integration File signal 1: autoint1.e
 Integration File signal 2: autoint2.e
 Quant Time: Jul 30 05:23:34 2024
 Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\P0072924.M
 Quant Title : GC EXTRACTABLES
 QLast Update : Tue Jul 30 05:21:29 2024
 Response via : Initial Calibration
 Integrator: ChemStation

Volume Inj. : 2 µl
 Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
 Signal #1 Info : 30Mx0.32mmx 0.50µ Signal #2 Info : 30M x 0.32mm x 0.25µm

	Compound	RT#1	RT#2	Resp#1	Resp#2	ng/ml	ng/ml

System Monitoring Compounds							
1)	SA Tetrachlo...	4.428	3.539	408.6E6	148.0E6	56.347	55.095
2)	SA Decachlor...	10.222	8.435	291.4E6	92974008	54.943	55.597
Target Compounds							
16)	L4 AR-1242-1	5.599	4.596	109.1E6	41664725	558.713	544.219
17)	L4 AR-1242-2	5.622	4.614	152.0E6	57586476	555.501	548.865
18)	L4 AR-1242-3	5.684	4.786	93266330	30835164	552.645	555.578
19)	L4 AR-1242-4	5.784	4.868	79461277	31086075	542.192	555.497
20)	L4 AR-1242-5	6.527	5.380	85565000	39396331	510.909	555.523



(f)=RT Delta > 1/2 Window (#)=Amounts differ by > 25% (m)=manual int.


Data Path : Z:\pestpcbsrv\HPCHEM1\ECD_O\Data\PO072924\
Data File : PO105033.D
Signal(s) : Signal #1: ECD1A.ch Signal #2: ECD2B.ch
Acq On : 30 Jul 2024 02:29
Operator : YP/AJ
Sample : AR1242ICV500
Misc :
ALS Vial : 32 Sample Multiplier: 1

Instrument :
ECD_O
ClientSampleId :
ICVPO072924AR1242

Integration File signal 1: autoint1.e
Integration File signal 2: autoint2.e
Quant Time: Jul 30 05:23:34 2024
Quant Method : Z:\pestpcbsrv\HPCHEM1\ECD_O\methods\PO072924.M
Quant Title : GC EXTRACTABLES
QLast Update : Tue Jul 30 05:21:29 2024
Response via : Initial Calibration
Integrator: ChemStation

Volume Inj. : 2 µl
Signal #1 Phase : ZB-MR1 Signal #2 Phase: ZB-MR2
Signal #1 Info : 30Mx0.32mmx 0.50µm Signal #2 Info : 30M x 0.32mm x 0.25µm






#1 Tetrachloro-m-xylene


R.T.: 4.428 min
Delta R.T.: 0.000 min
Response: 408608864
Conc: 56.35 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO072924AR1242




#1 Tetrachloro-m-xylene

R.T.: 3.539 min
Delta R.T.: 0.000 min
Response: 147971750
Conc: 55.10 ng/ml




#2 Decachlorobiphenyl

R.T.: 10.222 min
Delta R.T.: 0.003 min
Response: 291363415
Conc: 54.94 ng/ml



#2 Decachlorobiphenyl


R.T.: 8.435 min
Delta R.T.: 0.000 min
Response: 92974008
Conc: 55.60 ng/ml



#16 AR-1242-1


R.T.: 5.599 min
Delta R.T.: 0.000 min
Response: 109071232
Conc: 558.71 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO072924AR1242




#16 AR-1242-1

R.T.: 4.596 min
Delta R.T.: 0.000 min
Response: 41664725
Conc: 544.22 ng/ml




#17 AR-1242-2

R.T.: 5.622 min
Delta R.T.: 0.000 min
Response: 152048836
Conc: 555.50 ng/ml



#17 AR-1242-2


R.T.: 4.614 min
Delta R.T.: 0.000 min
Response: 57586476
Conc: 548.87 ng/ml



#18 AR-1242-3


R.T.: 5.684 min
Delta R.T.: 0.000 min
Response: 93266330
Conc: 552.65 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO072924AR1242




#18 AR-1242-3

R.T.: 4.786 min
Delta R.T.: 0.000 min
Response: 30835164
Conc: 555.58 ng/ml




#19 AR-1242-4

R.T.: 5.784 min
Delta R.T.: 0.002 min
Response: 79461277
Conc: 542.19 ng/ml



#19 AR-1242-4


R.T.: 4.868 min
Delta R.T.: 0.000 min
Response: 31086075
Conc: 555.50 ng/ml



#20 AR-1242-5

R.T.: 6.527 min
Delta R.T.: 0.002 min
Response: 85565000
Conc: 510.91 ng/ml

Instrument :
ECD_O
ClientSampleId :
ICVPO072924AR1242



#20 AR-1242-5

R.T.: 5.380 min
Delta R.T.: 0.000 min
Response: 39396331
Conc: 555.52 ng/ml