

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072924

Review By	yogesh	Review On	7/30/2024 8:25:12 AM
Supervise By	Ankita	Supervise On	7/30/2024 1:22:01 PM
SubDirectory	PO072924	HP Acquire Method	HP Processing Method PO072924

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23216,PP23217,PP23218,PP23219,PP23220,PP23221,PP23222,PP23223,PP23224,PP23225,PP23226,PP23227,PP23228,PP23229,PP23230,PP23231,PP23232,PP23233,PP23234,PP23235,PP23236,PP23237,PP23238,PP23239,PP23240,PP23241,PP23242,PP23243,PP23244,PP23245,PP23246,PP23247,PP23248,PP23249,PP23250,PP23251,PP23252,PP23253,PP23254,PP23255
CCC	PP23218,PP23223,PP23228,PP23233,PP23238,PP23243,PP23248,PP23253
Internal Standard/PEM	
ICV/I.BLK	PP23259,PP23261,PP23263,PP23265,PP23267,PP23269,PP23271,PP23273
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO105002.D	29 Jul 2024 17:32		YP/AJ	Ok
2	I.BLK	I.BLK	PO105003.D	29 Jul 2024 17:50		YP/AJ	Ok
3	AR1660ICC1000	AR1660ICC1000	PO105004.D	29 Jul 2024 18:07		YP/AJ	Ok
4	AR1660ICC750	AR1660ICC750	PO105005.D	29 Jul 2024 18:24		YP/AJ	Ok
5	AR1660ICC500	AR1660ICC500	PO105006.D	29 Jul 2024 18:42		YP/AJ	Ok
6	AR1660ICC250	AR1660ICC250	PO105007.D	29 Jul 2024 18:59		YP/AJ	Ok
7	AR1660ICC050	AR1660ICC050	PO105008.D	29 Jul 2024 19:16		YP/AJ	Ok,M
8	AR1221ICC500	AR1221ICC500	PO105009.D	29 Jul 2024 19:34		YP/AJ	Ok
9	AR1232ICC500	AR1232ICC500	PO105010.D	29 Jul 2024 19:51		YP/AJ	Ok
10	AR1242ICC1000	AR1242ICC1000	PO105011.D	29 Jul 2024 20:08		YP/AJ	Ok
11	AR1242ICC750	AR1242ICC750	PO105012.D	29 Jul 2024 20:26		YP/AJ	Ok
12	AR1242ICC500	AR1242ICC500	PO105013.D	29 Jul 2024 20:43		YP/AJ	Ok
13	AR1242ICC250	AR1242ICC250	PO105014.D	29 Jul 2024 21:00		YP/AJ	Ok
14	AR1242ICC050	AR1242ICC050	PO105015.D	29 Jul 2024 21:18		YP/AJ	Ok,M
15	AR1248ICC1000	AR1248ICC1000	PO105016.D	29 Jul 2024 21:35		YP/AJ	Ok
16	AR1248ICC750	AR1248ICC750	PO105017.D	29 Jul 2024 21:52		YP/AJ	Ok
17	AR1248ICC500	AR1248ICC500	PO105018.D	29 Jul 2024 22:10		YP/AJ	Ok
18	AR1248ICC250	AR1248ICC250	PO105019.D	29 Jul 2024 22:27		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO072924

Review By	yogesh	Review On	7/30/2024 8:25:12 AM
Supervise By	Ankita	Supervise On	7/30/2024 1:22:01 PM
SubDirectory	PO072924	HP Acquire Method	HP Processing Method PO072924
STD. NAME	STD REF.#		
Tune/Reschk	PP23216,PP23217,PP23218,PP23219,PP23220,PP23221,PP23222,PP23223,PP23224,PP23225,PP23226,PP23227,PP23228,PP23229,PP23230,P P23231,PP23232,PP23233,PP23234,PP23235,PP23236,PP23237,PP23238,PP23239,PP23240,PP23241,PP23242,PP23243,PP23244,PP23245,PP 23246,PP23247,PP23248,PP23249,PP23250,PP23251,PP23252,PP23253,PP23254,PP23255 PP23218,PP23223,PP23228,PP23233,PP23238,PP23243,PP23248,PP23253 PP23259,PP23261,PP23263,PP23265,PP23267,PP23269,PP23271,PP23273		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	AR1248ICC050	AR1248ICC050	PO105020.D	29 Jul 2024 22:44		YP/AJ	Ok,M
20	AR1254ICC1000	AR1254ICC1000	PO105021.D	29 Jul 2024 23:01		YP/AJ	Ok
21	AR1254ICC750	AR1254ICC750	PO105022.D	29 Jul 2024 23:19		YP/AJ	Ok
22	AR1254ICC500	AR1254ICC500	PO105023.D	29 Jul 2024 23:36		YP/AJ	Ok
23	AR1254ICC250	AR1254ICC250	PO105024.D	29 Jul 2024 23:53		YP/AJ	Ok
24	AR1254ICC050	AR1254ICC050	PO105025.D	30 Jul 2024 00:11		YP/AJ	Ok,M
25	AR1262ICC500	AR1262ICC500	PO105026.D	30 Jul 2024 00:28		YP/AJ	Ok
26	AR1268ICC1000	AR1268ICC1000	PO105027.D	30 Jul 2024 00:45		YP/AJ	Ok
27	AR1268ICC750	AR1268ICC750	PO105028.D	30 Jul 2024 01:02		YP/AJ	Ok
28	AR1268ICC500	AR1268ICC500	PO105029.D	30 Jul 2024 01:20		YP/AJ	Ok
29	AR1268ICC250	AR1268ICC250	PO105030.D	30 Jul 2024 01:37		YP/AJ	Ok
30	AR1268ICC050	AR1268ICC050	PO105031.D	30 Jul 2024 01:54		YP/AJ	Ok,M
31	PO072924ICV500	ICVPO072924	PO105032.D	30 Jul 2024 02:11	icv failed	YP/AJ	Not Ok
32	AR1242ICV500	ICVPO072924AR1242	PO105033.D	30 Jul 2024 02:29		YP/AJ	Ok
33	AR1248ICV500	ICVPO072924AR1248	PO105034.D	30 Jul 2024 02:46		YP/AJ	Ok
34	AR1254ICV500	ICVPO072924AR1254	PO105035.D	30 Jul 2024 03:03		YP/AJ	Ok
35	AR1268ICV500	ICVPO072924AR1268	PO105036.D	30 Jul 2024 03:20		YP/AJ	Ok
36	PO072924ICV500	ICVPO072924	PO105037.D	30 Jul 2024 10:30		YP/AJ	Ok

M : Manual Integration

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO080624

Review By	yogesh	Review On	8/7/2024 8:21:34 AM
Supervise By	Ankita	Supervise On	8/7/2024 10:00:34 AM
SubDirectory	PO080624	HP Acquire Method	HP Processing Method PO072924

STD. NAME	STD REF.#
Tune/Reschk	
Initial Calibration Stds	PP23216,PP23217,PP23218,PP23219,PP23220,PP23221,PP23222,PP23223,PP23224,PP23225,PP23226,PP23227,PP23228,PP23229,PP23230,PP23231,PP23232,PP23233,PP23234,PP23235,PP23236,PP23237,PP23238,PP23239,PP23240,PP23241,PP23242,PP23243,PP23244,PP23245,PP23246,PP23247,PP23248,PP23249,PP23250,PP23251,PP23252,PP23253,PP23254,PP23255
CCC	PP23218,PP23223,PP23228,PP23233,PP23238,PP23243,PP23248,PP23253
Internal Standard/PEM	
ICV/I.BLK	PP23259,PP23261,PP23263,PP23265,PP23267,PP23269,PP23271,PP23273
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	HEXANE	HEXANE	PO105214.D	06 Aug 2024 08:59		YP/AJ	Ok
2	AR1660CCC500	AR1660CCC500	PO105215.D	06 Aug 2024 09:38		YP/AJ	Ok
3	AR1242CCC500	AR1242CCC500	PO105216.D	06 Aug 2024 09:55		YP/AJ	Ok,M
4	AR1248CCC500	AR1248CCC500	PO105217.D	06 Aug 2024 10:13		YP/AJ	Ok
5	AR1254CCC500	AR1254CCC500	PO105218.D	06 Aug 2024 10:30		YP/AJ	Ok
6	I.BLK	I.BLK	PO105219.D	06 Aug 2024 10:48		YP/AJ	Ok
7	P3462-11DL	JC-3-6-7DL	PO105220.D	06 Aug 2024 11:05	AR1242 Hit	YP/AJ	Ok
8	P3462-03DL	JC-1-4-5DL	PO105221.D	06 Aug 2024 11:22	AR1242 Hit	YP/AJ	Ok
9	P3461-01DL2	CHRT-28607DL2	PO105222.D	06 Aug 2024 11:40	AR1254 Hit	YP/AJ	Ok
10	AR1660CCC500	AR1660CCC500	PO105223.D	06 Aug 2024 12:37		YP/AJ	Ok
11	AR1242CCC500	AR1242CCC500	PO105224.D	06 Aug 2024 12:54		YP/AJ	Ok
12	AR1248CCC500	AR1248CCC500	PO105225.D	06 Aug 2024 13:12		YP/AJ	Ok
13	AR1254CCC500	AR1254CCC500	PO105226.D	06 Aug 2024 13:29		YP/AJ	Ok
14	I.BLK	I.BLK	PO105227.D	06 Aug 2024 13:47		YP/AJ	Ok
15	PB162521BL	PB162521BL	PO105228.D	06 Aug 2024 15:14		YP/AJ	Ok
16	PB162521BS	PB162521BS	PO105229.D	06 Aug 2024 15:31		YP/AJ	Ok
17	P3468-01	IDW-S1-01	PO105230.D	06 Aug 2024 15:49		YP/AJ	Ok
18	P3474-01	WC-A1-COMP	PO105231.D	06 Aug 2024 16:06		YP/AJ	Ok

Instrument ID: ECD_O

Daily Analysis Runlog For Sequence/QC Batch ID # PO080624

Review By	yogesh	Review On	8/7/2024 8:21:34 AM
Supervise By	Ankita	Supervise On	8/7/2024 10:00:34 AM
SubDirectory	PO080624	HP Acquire Method	HP Processing Method PO072924
STD. NAME	STD REF.#		
Tune/Reschk	PP23216,PP23217,PP23218,PP23219,PP23220,PP23221,PP23222,PP23223,PP23224,PP23225,PP23226,PP23227,PP23228,PP23229,PP23230,P P23231,PP23232,PP23233,PP23234,PP23235,PP23236,PP23237,PP23238,PP23239,PP23240,PP23241,PP23242,PP23243,PP23244,PP23245,PP 23246,PP23247,PP23248,PP23249,PP23250,PP23251,PP23252,PP23253,PP23254,PP23255 PP23218,PP23223,PP23228,PP23233,PP23238,PP23243,PP23248,PP23253 PP23259,PP23261,PP23263,PP23265,PP23267,PP23269,PP23271,PP23273		
Initial Calibration Stds			
CCC			
Internal Standard/PEM			
ICV/I.BLK			
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	P3474-04	WC-B1-COMP	PO105232.D	06 Aug 2024 16:23		YP/AJ	Ok
20	P3474-07	WC-A2-COMP	PO105233.D	06 Aug 2024 16:41	AR1254 hit	YP/AJ	Ok,M
21	P3474-10	WC-B2-COMP	PO105234.D	06 Aug 2024 16:58		YP/AJ	Ok
22	P3475-02	S-866-N-SO-1.0-1.5-08	PO105235.D	06 Aug 2024 17:16		YP/AJ	Ok,M
23	P3475-03MS	S-866-N-SO-1.0-1.5-08	PO105236.D	06 Aug 2024 17:33		YP/AJ	Ok
24	P3475-04MSD	S-866-N-SO-1.0-1.5-08	PO105237.D	06 Aug 2024 17:50		YP/AJ	Ok
25	AR1660CCC500	AR1660CCC500	PO105238.D	06 Aug 2024 18:36		YP/AJ	Ok
26	AR1242CCC500	AR1242CCC500	PO105239.D	06 Aug 2024 18:54		YP/AJ	Ok
27	AR1248CCC500	AR1248CCC500	PO105240.D	06 Aug 2024 19:11		YP/AJ	Ok
28	AR1254CCC500	AR1254CCC500	PO105241.D	06 Aug 2024 19:28		YP/AJ	Ok
29	I.BLK	I.BLK	PO105242.D	06 Aug 2024 19:46		YP/AJ	Ok

M : Manual Integration