



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P3475

Analytical Method: 8082A

Client: JACOBS Engineering Group, Inc.

Datafile : PO105229.D

Lab Sample ID	Parameter	Spike	Result	Units	Rec	RPD	Qual	RPD	Limits		RPD
								Qual	Low	High	
PB162521BS (Column 1)	AR1016	166.6	193	ug/kg	116				40 (71)	140 (120)	
	AR1260	166.6	169	ug/kg	101				40 (65)	140 (130)	
PB162521BS (Column 2)	AR1016	166.6	184	ug/kg	110				40 (71)	140 (120)	
	AR1260	166.6	169	ug/kg	101				40 (65)	140 (130)	