

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger Site Princeton NJ	Date Received:	
Client Sample ID:	PB162520BS	SDG No.:	P3475
Lab Sample ID:	PB162520BS	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	100
Sample Wt/Vol:	30.03 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group2
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF138967.D	1	08/06/24 08:26	08/13/24 11:08	PB162520

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
56-55-3	Benzo(a)anthracene	1600		80.6	170	ug/Kg
50-32-8	Benzo(a)pyrene	1800		92.9	170	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	1500		81.1	170	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	87.2		30 (18) - 130 (107)	87%	SPK: 100
321-60-8	2-Fluorobiphenyl	86.9		30 (20) - 130 (109)	87%	SPK: 100
1718-51-0	Terphenyl-d14	111		30 (10) - 130 (105)	111%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	49400	6.84			
1146-65-2	Naphthalene-d8	210000	8.122			
15067-26-2	Acenaphthene-d10	120000	9.875			
1517-22-2	Phenanthrene-d10	210000	11.363			
1719-03-5	Chrysene-d12	104000	14.004			
1520-96-3	Perylene-d12	91900	15.463			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products