

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF073024\
 Data File : BF138682.D
 Acq On : 30 Jul 2024 13:56
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010

Quant Time: Jul 30 17:42:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 30 17:38:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	6.845	152	73349	20.000	ng	0.00
21) Naphthalene-d8	8.128	136	297904	20.000	ng	0.00
39) Acenaphthene-d10	9.880	164	162245	20.000	ng	0.00
64) Phenanthrene-d10	11.369	188	269225	20.000	ng	0.00
76) Chrysene-d12	14.004	240	146258	20.000	ng	0.00
86) Perylene-d12	15.468	264	137026	20.000	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.469	112	100297	21.108	ng	0.00
7) Phenol-d6	6.475	99	133232	20.884	ng	-0.01
23) Nitrobenzene-d5	7.404	82	123997	20.350	ng	0.00
42) 2,4,6-Tribromophenol	10.669	330	26595	20.011	ng	0.00
45) 2-Fluorobiphenyl	9.198	172	231000	21.392	ng	0.00
79) Terphenyl-d14	12.945	244	185079	21.187	ng	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	2.593	88	20339	9.777	ng	95
3) Pyridine	3.357	79	50057	9.933	ng	98
4) n-Nitrosodimethylamine	3.287	42	29962	9.983	ng	# 98
6) Aniline	6.504	93	59629	10.481	ng	94
8) 2-Chlorophenol	6.628	128	51997	10.401	ng	97
9) Benzaldehyde	6.398	77	41047	10.733	ng	99
10) Phenol	6.492	94	71371	10.626	ng	92
11) bis(2-Chloroethyl)ether	6.575	93	54121	10.471	ng	99
12) 1,3-Dichlorobenzene	6.787	146	59006	10.544	ng	98
13) 1,4-Dichlorobenzene	6.863	146	59226	10.487	ng	97
14) 1,2-Dichlorobenzene	7.016	146	55425	10.501	ng	98
15) Benzyl Alcohol	6.987	79	47206	10.267	ng	99
16) 2,2'-oxybis(1-Chloropr...	7.116	45	94469	10.620	ng	97
17) 2-Methylphenol	7.098	107	42323	10.252	ng	97
18) Hexachloroethane	7.351	117	22530	10.598	ng	97
19) n-Nitroso-di-n-propyla...	7.245	70	39862	10.345	ng	99
20) 3+4-Methylphenols	7.251	107	56989	10.760	ng	97
22) Acetophenone	7.251	105	75892	10.404	ng	99
24) Nitrobenzene	7.422	77	63020	10.164	ng	98
25) Isophorone	7.663	82	105784	10.167	ng	99
26) 2-Nitrophenol	7.739	139	26149	9.803	ng	96
27) 2,4-Dimethylphenol	7.781	122	32386	10.147	ng	98
28) bis(2-Chloroethoxy)met...	7.875	93	64569	10.191	ng	99
29) 2,4-Dichlorophenol	7.987	162	41138	10.031	ng	99
30) 1,2,4-Trichlorobenzene	8.063	180	48272	10.199	ng	98
31) Naphthalene	8.145	128	164167	10.469	ng	99
32) Benzoic acid	7.875	122	18580	7.409	ng	98
33) 4-Chloroaniline	8.198	127	54003	10.260	ng	98
34) Hexachlorobutadiene	8.263	225	29309	10.224	ng	98
35) Caprolactam	8.539	113	12474	10.193	ng	93
36) 4-Chloro-3-methylphenol	8.681	107	47584	10.152	ng	96
37) 2-Methylnaphthalene	8.839	142	104097	10.511	ng	99
38) 1-Methylnaphthalene	8.934	142	102100	10.521	ng	100
40) 1,2,4,5-Tetrachloroben...	9.004	216	45439	10.082	ng	98
41) Hexachlorocyclopentadiene	8.986	237	5161	10.915	ng	96
43) 2,4,6-Trichlorophenol	9.116	196	27439	9.985	ng	98

Data Path : Z:\svoasrv\HPCHEM1\BNA_F\Data\BF073024\
 Data File : BF138682.D
 Acq On : 30 Jul 2024 13:56
 Operator : RC/JU
 Sample : SSTDICC010
 Misc :
 ALS Vial : 4 Sample Multiplier: 1


Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Quant Time: Jul 30 17:42:53 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Jul 30 17:38:59 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
44) 2,4,5-Trichlorophenol	9.163	196	30782	10.247	ng	97
46) 1,1'-Biphenyl	9.298	154	132668	10.441	ng	99
47) 2-Chloronaphthalene	9.328	162	97850	10.354	ng	97
48) 2-Nitroaniline	9.422	65	32731	10.216	ng	97
49) Acenaphthylene	9.739	152	141438	10.552	ng	99
50) Dimethylphthalate	9.598	163	105187	10.139	ng	100
51) 2,6-Dinitrotoluene	9.663	165	23351	9.974	ng	95
52) Acenaphthene	9.910	154	95143	10.560	ng	98
53) 3-Nitroaniline	9.833	138	24353	10.062	ng	99
54) 2,4-Dinitrophenol	9.951	184	7908	7.337	ng	88
55) Dibenzofuran	10.080	168	136302	10.717	ng	99
56) 4-Nitrophenol	10.004	139	12643	8.687	ng	95
57) 2,4-Dinitrotoluene	10.069	165	30462	10.198	ng	99
58) Fluorene	10.428	166	107585	10.622	ng	100
59) 2,3,4,6-Tetrachlorophenol	10.204	232	22672	9.872	ng	# 94
60) Diethylphthalate	10.298	149	102108	10.381	ng	99
61) 4-Chlorophenyl-phenyle...	10.416	204	53870	10.814	ng	99
62) 4-Nitroaniline	10.439	138	23653	10.284	ng	98
63) Azobenzene	10.575	77	112880	10.347	ng	99
65) 4,6-Dinitro-2-methylph...	10.475	198	14344	8.733	ng	96
66) n-Nitrosodiphenylamine	10.533	169	86813	10.316	ng	100
67) 4-Bromophenyl-phenylether	10.910	248	29510	10.124	ng	98
68) Hexachlorobenzene	10.975	284	31049	10.317	ng	96
69) Atrazine	11.063	200	23189	10.680	ng	96
70) Pentachlorophenol	11.175	266	10392	7.661	ng	93
71) Phenanthrene	11.392	178	149253	10.766	ng	99
72) Anthracene	11.439	178	145043	10.620	ng	99
73) Carbazole	11.598	167	126321	10.721	ng	100
74) Di-n-butylphthalate	11.922	149	133246	10.060	ng	99
75) Fluoranthene	12.580	202	140690	10.871	ng	99
77) Benzidine	12.704	184	35263	10.080	ng	100
78) Pyrene	12.810	202	143221	10.400	ng	99
80) Butylbenzylphthalate	13.421	149	41771	9.472	ng	99
81) Benzo(a)anthracene	13.998	228	105185	10.444	ng	97
82) 3,3'-Dichlorobenzidine	13.957	252	26584	10.314	ng	98
83) Chrysene	14.033	228	89780	9.881	ng	99
84) Bis(2-ethylhexyl)phtha...	13.980	149	61314	9.495	ng	98
85) Di-n-octyl phthalate	14.592	149	111460	9.329	ng	100
87) Indeno(1,2,3-cd)pyrene	16.933	276	100310	10.215	ng	99
88) Benzo(b)fluoranthene	15.039	252	93263	10.980	ng	99
89) Benzo(k)fluoranthene	15.068	252	71464	9.717	ng	99
90) Benzo(a)pyrene	15.404	252	71724	10.038	ng	98
91) Dibenzo(a,h)anthracene	16.945	278	84356	10.465	ng	97
92) Benzo(g,h,i)perylene	17.374	276	84564	10.110	ng	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Jul 30 17:42:53 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_F\Methods\8270-BF073024.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Jul 30 17:38:59 2024
Response via : Initial Calibration









#1
1,4-Dichlorobenzene-d4
Concen: 20.000 ng
RT: 6.845 min Scan# 80
Delta R.T. 0.000 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:152 Resp: 73349
Ion Ratio Lower Upper
152 100
150 157.4 126.0 189.0
115 66.0 51.7 77.5



#2
1,4-Dioxane
Concen: 9.777 ng
RT: 2.593 min Scan# 85
Delta R.T. 0.018 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion: 88 Resp: 20339
Ion Ratio Lower Upper
88 100
58 94.3 71.6 107.4
43 38.8 28.7 43.1





#3

Pyridine

Concen: 9.933 ng

RT: 3.357 min Scan# 211

Delta R.T. 0.024 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion: 79 Resp: 50057

Ion Ratio Lower Upper

79 100

52 92.6 74.7 112.1

51 46.4 34.6 51.8



#4

n-Nitrosodimethylamine

Concen: 9.983 ng

RT: 3.287 min Scan# 203

Delta R.T. -0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion: 42 Resp: 29962


Ion Ratio Lower Upper

42 100

74 107.5 84.2 126.4

44 8.8 4.9 7.3#









#5
2-Fluorophenol
Concen: 21.108 ng
RT: 5.469 min Scan# 51
Delta R.T. 0.000 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:112 Resp: 100297
Ion Ratio Lower Upper
112 100
64 66.6 54.2 81.4
63 35.2 28.7 43.1



#6
Aniline
Concen: 10.481 ng
RT: 6.504 min Scan# 750
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion: 93 Resp: 59629
Ion Ratio Lower Upper
93 100
66 54.0 46.9 70.3
65 29.9 26.5 39.7









#7
Phenol-d6
Concen: 20.884 ng
RT: 6.475 min Scan# 747
Delta R.T. -0.012 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion: 99 Resp: 133232
Ion Ratio Lower Upper
99 100
42 22.2 17.4 26.0
71 36.0 28.1 42.1



#8
2-Chlorophenol
Concen: 10.401 ng
RT: 6.628 min Scan# 771
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion: 128 Resp: 51997
Ion Ratio Lower Upper
128 100
130 31.2 12.0 52.0
64 58.9 36.3 76.3





#9

Benzaldehyde

Concen: 10.733 ng

RT: 6.398 min Scan# 748

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion: 77 Resp: 41047

Ion Ratio Lower Upper

77 100

105 94.8 72.9 112.9

106 88.9 68.4 108.4



#10

Phenol

Concen: 10.626 ng

RT: 6.492 min Scan# 748

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion: 94 Resp: 71371


Ion Ratio Lower Upper

94 100

65 33.6 16.9 56.9

66 50.1 36.5 76.5









#11
bis(2-Chloroethyl)ether
Concen: 10.471 ng
RT: 6.575 min Scan# 762
Delta R.T. -0.012 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion: 93 Resp: 54121
Ion Ratio Lower Upper
93 100
63 84.5 65.3 105.3
95 31.8 12.4 52.4



#12
1,3-Dichlorobenzene
Concen: 10.544 ng
RT: 6.787 min Scan# 798
Delta R.T. 0.000 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion:146 Resp: 59006
Ion Ratio Lower Upper
146 100
148 62.2 51.2 76.8
75 32.6 27.4 41.2





#13

1,4-Dichlorobenzene

Concen: 10.487 ng

RT: 6.863 min Scan# 811

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:146 Resp: 59226

Ion Ratio Lower Upper

146 100

148 65.4 50.2 75.2

111 43.4 35.9 53.9



#14

1,2-Dichlorobenzene

Concen: 10.501 ng

RT: 7.016 min Scan# 837

Delta R.T. -0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:146 Resp: 55425


Ion Ratio Lower Upper

146 100

148 65.7 50.8 76.2

111 45.7 37.4 56.2









#15
Benzyl Alcohol
Concen: 10.267 ng
RT: 6.987 min Scan# 832
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion: 79 Resp: 47206
Ion Ratio Lower Upper
79 100
108 71.4 56.6 85.0
77 64.6 50.3 75.5



#16
2,2'-oxybis(1-Chloropropane)
Concen: 10.620 ng
RT: 7.116 min Scan# 854
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion: 45 Resp: 94469
Ion Ratio Lower Upper
45 100
77 13.0 0.0 34.9
79 11.7 0.0 32.2









#17
2-Methylphenol
Concen: 10.252 ng
RT: 7.098 min Scan# 851
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDIC010

Tgt Ion:	107	Resp:	42323
Ion	Ratio	Lower	Upper
107	100		
108	112.9	89.2	133.8
77	49.1	43.0	64.4
79	49.7	42.2	63.2



#18
Hexachloroethane
Concen: 10.598 ng
RT: 7.351 min Scan# 894
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion:	117	Resp:	22530
Ion	Ratio	Lower	Upper
117	100		
119	95.1	74.6	111.8
201	91.9	77.2	115.8





#19

n-Nitroso-di-n-propylamine

Concen: 10.345 ng

RT: 7.245 min Scan# 81

Delta R.T. -0.012 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010

Tgt Ion: 70 Resp: 39862





Ion Ratio Lower Upper

70 100

42 72.3 57.4 86.0

101 9.5 7.5 11.3

130 19.1 16.4 24.6



#20

3+4-Methylphenols

Concen: 10.760 ng

RT: 7.251 min Scan# 877

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Tgt Ion:107 Resp: 56989




Ion Ratio Lower Upper


107 100

108 87.2 68.2 108.2

77 158.6 132.1 172.1

79 30.6 11.5 51.5









#21
Naphthalene-d8
Concen: 20.000 ng
RT: 8.128 min Scan# 1026
Delta R.T. 0.000 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:136	Resp:	297904	
Ion	Ratio	Lower	Upper
136	100		
137	10.8	8.9	13.3
54	10.1	8.6	12.8
68	6.0	4.8	7.2



#22
Acetophenone
Concen: 10.404 ng
RT: 7.251 min Scan# 877
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt	Ion:105	Resp:	75892
Ion	Ratio	Lower	Upper
105	100		
71	8.8	7.2	10.8
51	45.2	35.9	53.9
120	22.7	17.6	26.4





#23

Nitrobenzene-d5

Concen: 20.350 ng

RT: 7.404 min Scan# 904

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion: 82 Resp: 123997

Ion Ratio Lower Upper

82 100

128 41.4 32.8 49.2

54 59.1 48.3 72.5



#24

Nitrobenzene

Concen: 10.164 ng

RT: 7.422 min Scan# 906

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion: 77 Resp: 63020


Ion Ratio Lower Upper

77 100

123 42.9 33.3 49.9

65 14.0 11.9 17.9









#25
Isophorone
Concen: 10.167 ng
RT: 7.663 min Scan# 948
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion: 82 Resp: 105784
Ion Ratio Lower Upper
82 100
95 7.0 5.7 8.5
138 16.4 13.7 20.5



#26
2-Nitrophenol
Concen: 9.803 ng
RT: 7.739 min Scan# 960
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion:139 Resp: 26149
Ion Ratio Lower Upper
139 100
109 34.4 25.9 38.9
65 62.5 47.0 70.6









#27
2,4-Dimethylphenol
Concen: 10.147 ng
RT: 7.781 min Scan# 967
Delta R.T. 0.000 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:122 Resp: 32386
Ion Ratio Lower Upper
122 100
107 117.6 95.0 142.6
121 57.1 47.3 70.9



#28
bis(2-Chloroethoxy)methane
Concen: 10.191 ng
RT: 7.875 min Scan# 983
Delta R.T. -0.000 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion: 93 Resp: 64569
Ion Ratio Lower Upper
93 100
95 32.0 25.8 38.8
123 12.3 9.4 14.0









#29
2,4-Dichlorophenol
Concen: 10.031 ng
RT: 7.987 min Scan# 1002
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:162 Resp: 41138
Ion Ratio Lower Upper
162 100
164 65.4 44.7 84.7
98 41.8 20.8 60.8



#30
1,2,4-Trichlorobenzene
Concen: 10.199 ng
RT: 8.063 min Scan# 1015
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion:180 Resp: 48272
Ion Ratio Lower Upper
180 100
182 93.7 76.9 115.3
145 30.4 25.0 37.4





#31

Naphthalene

Concen: 10.469 ng

RT: 8.145 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:128 Resp: 164167

Ion Ratio Lower Upper

128 100

129 10.7 8.7 13.1

127 13.0 10.6 16.0



#32

Benzoic acid

Concen: 7.409 ng

RT: 7.875 min Scan# 983

Delta R.T. -0.029 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:122 Resp: 18580


Ion Ratio Lower Upper

122 100

105 125.2 106.7 146.7

77 102.9 81.1 121.1





#33

4-Chloroaniline

Concen: 10.260 ng

RT: 8.198 min Scan# 1

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010

Tgt Ion:127 Resp: 54003





Ion Ratio Lower Upper

127 100

129 29.9 25.9 38.9

65 33.9 27.6 41.4

92 20.3 16.8 25.2



#34

Hexachlorobutadiene

Concen: 10.224 ng

RT: 8.263 min Scan# 1049

Delta R.T. -0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:225 Resp: 29309


Ion Ratio Lower Upper

225 100

223 62.3 51.2 76.8

227 62.4 51.1 76.7





#35

Caprolactam

Concen: 10.193 ng

RT: 8.539 min Scan# 1101

Delta R.T. -0.029 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:113 Resp: 12474

Ion Ratio Lower Upper

113 100

55 198.5 186.7 226.7

56 148.2 138.9 178.9



#36

4-Chloro-3-methylphenol

Concen: 10.152 ng

RT: 8.681 min Scan# 1120

Delta R.T. -0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:107 Resp: 47584


Ion Ratio Lower Upper

107 100

144 25.1 18.2 27.2

142 74.5 57.4 86.2





#37

2-Methylnaphthalene

Concen: 10.511 ng

RT: 8.839 min Scan# 1147

Delta R.T. -0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :





SSTDICC010

Tgt Ion:142 Resp: 104097

Ion Ratio Lower Upper

142 100

141 87.8 70.8 106.2



#38

1-Methylnaphthalene

Concen: 10.521 ng

RT: 8.934 min Scan# 1163

Delta R.T. -0.006 min

Lab File: BF138682.D




Acq: 30 Jul 2024 13:56


Tgt Ion:142 Resp: 102100

Ion Ratio Lower Upper

142 100

141 91.3 73.1 109.7





#39

Acenaphthene-d10

Concen: 20.000 ng

RT: 9.880 min Scan# 1175

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:164 Resp: 162245

Ion Ratio Lower Upper

164 100

162 101.3 79.4 119.0

160 44.3 35.1 52.7



#40

1,2,4,5-Tetrachlorobenzene

Concen: 10.082 ng

RT: 9.004 min Scan# 1175

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Tgt Ion:216 Resp: 45439




Ion Ratio Lower Upper


216 100

214 78.4 63.9 95.9

179 23.9 17.8 26.6

108 19.1 16.0 24.0





#41

Hexachlorocyclopentadiene

Concen: 10.915 ng

RT: 8.986 min Scan# 1172

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:237 Resp: 5161

Ion Ratio Lower Upper

237 100

235 65.1 41.8 81.8

272 12.6 0.0 32.2



#42

2,4,6-Tribromophenol

Concen: 20.011 ng

RT: 10.669 min Scan# 1458

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:330 Resp: 26595


Ion Ratio Lower Upper

330 100

332 95.6 76.4 114.6

141 41.0 31.1 46.7





#43

2,4,6-Trichlorophenol

Concen: 9.985 ng

RT: 9.116 min Scan# 1194

Delta R.T. -0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:196 Resp: 27439

Ion Ratio Lower Upper

196 100

198 102.7 80.5 120.7

200 33.2 25.9 38.9



#44

2,4,5-Trichlorophenol

Concen: 10.247 ng

RT: 9.163 min Scan# 1202

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Tgt Ion:196 Resp: 30782




Ion Ratio Lower Upper


196 100

198 97.4 81.2 121.8

97 57.6 47.8 71.6

132 30.8 25.3 37.9





#45

2-Fluorobiphenyl

Concen: 21.392 ng

RT: 9.198 min Scan# 1208

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:172 Resp: 231000

Ion Ratio Lower Upper

172 100

171 35.3 28.3 42.5

170 23.2 18.8 28.2



#46

1,1'-Biphenyl

Concen: 10.441 ng

RT: 9.298 min Scan# 1225

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:154 Resp: 132668


Ion Ratio Lower Upper

154 100

153 40.2 20.8 60.8

76 12.8 0.0 32.8





#47

2-Chloronaphthalene

Concen: 10.354 ng

RT: 9.328 min Scan# 1230

Delta R.T. -0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDIC010





Tgt Ion:162 Resp: 97850

Ion Ratio Lower Upper

162 100

127 41.4 35.4 53.2

164 33.1 25.6 38.4



#48

2-Nitroaniline

Concen: 10.216 ng

RT: 9.422 min Scan# 1246

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion: 65 Resp: 32731


Ion Ratio Lower Upper

65 100

92 63.9 52.0 78.0

138 90.5 76.2 114.4





#49

Acenaphthylene

Concen: 10.552 ng

RT: 9.739 min Scan# 14

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:152 Resp: 141438

Ion Ratio Lower Upper

152 100

151 20.0 16.0 24.0

153 13.2 11.0 16.4



#50

Dimethylphthalate

Concen: 10.139 ng

RT: 9.598 min Scan# 1276

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:163 Resp: 105187


Ion Ratio Lower Upper

163 100

194 3.6 3.1 4.7

164 9.8 7.8 11.8





#51

2,6-Dinitrotoluene

Concen: 9.974 ng

RT: 9.663 min Scan# 1287

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDIC010





Tgt Ion:165 Resp: 23351

Ion Ratio Lower Upper

165 100

63 67.4 52.0 78.0

89 63.4 47.0 70.6



#52

Acenaphthene

Concen: 10.560 ng

RT: 9.910 min Scan# 1329

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:154 Resp: 95143


Ion Ratio Lower Upper

154 100

153 111.6 89.9 134.9

152 48.3 40.6 60.8





#53

3-Nitroaniline

Concen: 10.062 ng

RT: 9.833 min Scan# 1317

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:138 Resp: 24353

Ion Ratio Lower Upper

138 100

108 11.9 9.1 13.7

92 122.5 98.7 148.1



#54

2,4-Dinitrophenol

Concen: 7.337 ng

RT: 9.951 min Scan# 1336

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:184 Resp: 7908


Ion Ratio Lower Upper

184 100

63 79.4 57.5 86.3

154 75.8 51.7 77.5





#55

Dibenzenofuran

Concen: 10.717 ng

RT: 10.080 min Scan# 1359

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:168 Resp: 136302

Ion Ratio Lower Upper

168 100

139 39.8 32.6 49.0

169 13.0 10.7 16.1



#56

4-Nitrophenol

Concen: 8.687 ng

RT: 10.004 min Scan# 1345

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:139 Resp: 12643


Ion Ratio Lower Upper

139 100

109 78.0 55.5 95.5

65 123.5 96.7 136.7









#57
2,4-Dinitrotoluene
Concen: 10.198 ng
RT: 10.069 min Scan# 1417
Delta R.T. 0.000 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:165	Resp:	30462
Ion Ratio	Lower	Upper
165	100	
63	56.8	46.3 69.5
89	81.7	64.2 96.4
182	2.9	2.5 3.7



#58
Fluorene
Concen: 10.622 ng
RT: 10.428 min Scan# 1417
Delta R.T. 0.000 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion:166	Resp:	107585
Ion Ratio	Lower	Upper
166	100	
165	97.6	78.4 117.6
167	13.1	10.6 16.0





#59

2,3,4,6-Tetrachlorophenol

Concen: 9.872 ng

RT: 10.204 min Scan# 1379

Delta R.T. -0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010

Tgt Ion:232 Resp: 22672





Ion Ratio Lower Upper

232 100

131 52.2 37.0 55.4

130 1.6 2.0 3.0#

166 31.5 24.7 37.1



#60

Diethylphthalate

Concen: 10.381 ng

RT: 10.298 min Scan# 1395

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:149 Resp: 102108


Ion Ratio Lower Upper

149 100

177 22.6 17.8 26.8

150 13.1 10.1 15.1









#61
4-Chlorophenyl-phenylether
Concen: 10.814 ng
RT: 10.416 min Scan# 1415
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:204 Resp: 53870
Ion Ratio Lower Upper
204 100
206 31.3 26.1 39.1
141 64.8 51.4 77.0



#62
4-Nitroaniline
Concen: 10.284 ng
RT: 10.439 min Scan# 1419
Delta R.T. -0.012 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion:138 Resp: 23653
Ion Ratio Lower Upper
138 100
92 55.6 34.2 74.2
108 75.1 56.2 96.2





#63

Azobenzene

Concen: 10.347 ng

RT: 10.575 min Scan# 1442

Delta R.T. -0.006 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010

Tgt Ion: 77 Resp: 112880





Ion Ratio Lower Upper

77 100

182 22.7 3.4 43.4

105 19.7 0.2 40.2

51 34.9 14.6 54.6



#64

Phenanthrene-d10

Concen: 20.000 ng

RT: 11.369 min Scan# 1577

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:188 Resp: 269225


Ion Ratio Lower Upper

188 100

94 10.1 7.6 11.4

80 10.7 8.6 12.8









#65
4,6-Dinitro-2-methylphenol
Concen: 8.733 ng
RT: 10.475 min Scan# 1425
Delta R.T. -0.005 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:198 Resp: 14344
Ion Ratio Lower Upper
198 100
51 62.9 39.9 79.9
105 43.9 26.1 66.1



#66
n-Nitrosodiphenylamine
Concen: 10.316 ng
RT: 10.533 min Scan# 1435
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion:169 Resp: 86813
Ion Ratio Lower Upper
169 100
168 66.7 53.0 79.6
167 36.2 29.0 43.6









#67
4-Bromophenyl-phenylether
Concen: 10.124 ng
RT: 10.910 min Scan# 1499
Delta R.T. -0.000 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDIC010


Tgt Ion:248 Resp: 29510
Ion Ratio Lower Upper
248 100
250 94.2 77.7 116.5
141 84.6 68.0 102.0



#68
Hexachlorobenzene
Concen: 10.317 ng
RT: 10.975 min Scan# 1510
Delta R.T. -0.000 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion:284 Resp: 31049
Ion Ratio Lower Upper
284 100
142 35.9 31.3 46.9
249 32.7 27.2 40.8





#69

Atrazine

Concen: 10.680 ng

RT: 11.063 min Scan# 1525

Delta R.T. -0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:200 Resp: 23189

Ion Ratio Lower Upper

200 100

173 25.2 6.0 46.0

215 49.3 26.1 66.1



#70

Pentachlorophenol

Concen: 7.661 ng

RT: 11.175 min Scan# 1544

Delta R.T. -0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:266 Resp: 10392


Ion Ratio Lower Upper

266 100

268 62.8 49.2 73.8

264 71.1 49.8 74.6









#71
Phenanthrene
Concen: 10.766 ng
RT: 11.392 min Scan# 1581
Delta R.T. 0.000 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:178 Resp: 149253
Ion Ratio Lower Upper
178 100
176 18.6 15.4 23.0
179 15.1 12.2 18.2



#72
Anthracene
Concen: 10.620 ng
RT: 11.439 min Scan# 1589
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56


Tgt Ion:178 Resp: 145043
Ion Ratio Lower Upper
178 100
176 18.5 14.9 22.3
179 15.0 12.4 18.6








#73
 Carbazole
 Concen: 10.721 ng
 RT: 11.598 min Scan# 1616
 Delta R.T. 0.000 min
 Lab File: BF138682.D
 Acq: 30 Jul 2024 13:56


Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010





Tgt Ion:167 Resp: 126321
 Ion Ratio Lower Upper
 167 100
 166 21.3 17.2 25.8
 139 13.5 10.6 16.0




#74
 Di-n-butylphthalate
 Concen: 10.060 ng
 RT: 11.922 min Scan# 1671
 Delta R.T. -0.000 min
 Lab File: BF138682.D
 Acq: 30 Jul 2024 13:56



Tgt Ion:149 Resp: 133246
 Ion Ratio Lower Upper
 149 100
 150 8.8 7.4 11.0
 104 5.0 4.1 6.1









#75
Fluoranthene
Concen: 10.871 ng
RT: 12.580 min Scan# 1783
Delta R.T. -0.000 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:	202	Resp:	140690
Ion Ratio	Lower	Upper	
202	100		
101	10.7	0.0	31.2
203	16.8	0.0	37.3



#76
Chrysene-d12
Concen: 20.000 ng
RT: 14.004 min Scan# 2025
Delta R.T. -0.000 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion:	240	Resp:	146258
Ion Ratio	Lower	Upper	
240	100		
120	13.0	10.2	15.4
236	26.0	19.8	29.8





#77

Benzidine

Concen: 10.080 ng

RT: 12.704 min Scan# 1804

Delta R.T. -0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:184 Resp: 35263

Ion Ratio Lower Upper

184 100

185 13.9 11.1 16.7

183 11.9 9.6 14.4



#78

Pyrene

Concen: 10.400 ng

RT: 12.810 min Scan# 1822

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:202 Resp: 143221


Ion Ratio Lower Upper

202 100

200 20.6 16.8 25.2

203 17.2 13.8 20.6






#79
 Terphenyl-d14
 Concen: 21.187 ng
 RT: 12.945 min Scan# 1845
 Delta R.T. -0.000 min
 Lab File: BF138682.D
 Acq: 30 Jul 2024 13:56

Instrument :




BNA_F

ClientSampleId :

SSTDIC010







Tgt Ion:244 Resp: 185079
 Ion Ratio Lower Upper
 244 100
 212 6.5 5.4 8.2
 122 11.8 9.6 14.4



#80
 Butylbenzylphthalate
 Concen: 9.472 ng
 RT: 13.421 min Scan# 1926
 Delta R.T. -0.000 min
 Lab File: BF138682.D
 Acq: 30 Jul 2024 13:56

Tgt Ion:149 Resp: 41771
 Ion Ratio Lower Upper
 149 100
 91 80.4 63.7 95.5
 206 19.6 16.2 24.2









#81
Benzo(a)anthracene
Concen: 10.444 ng
RT: 13.998 min Scan# 2024
Delta R.T. 0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56




Instrument :
BNA_F
ClientSampleId :
SSTDICC010


Tgt Ion:228 Resp: 105185
Ion Ratio Lower Upper
228 100
226 25.8 22.1 33.1
229 18.3 15.4 23.0



#82
3,3'-Dichlorobenzidine
Concen: 10.314 ng
RT: 13.957 min Scan# 2017
Delta R.T. 0.000 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion:252 Resp: 26584
Ion Ratio Lower Upper
252 100
254 64.9 50.8 76.2
126 13.3 10.2 15.2





#83

Chrysene

Concen: 9.881 ng

RT: 14.033 min Scan# 2030

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:228 Resp: 89780

Ion Ratio Lower Upper

228 100

226 30.4 23.7 35.5

229 19.2 15.0 22.6



#84

Bis(2-ethylhexyl)phthalate

Concen: 9.495 ng

RT: 13.980 min Scan# 2021

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:149 Resp: 61314


Ion Ratio Lower Upper

149 100

167 26.6 22.2 33.4

279 4.3 3.4 5.0





#85

Di-n-octyl phthalate

Concen: 9.329 ng

RT: 14.592 min Scan# 2125

Delta R.T. -0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:149 Resp: 111460

Ion Ratio Lower Upper

149 100

167 1.6 1.4 2.0

43 12.8 10.4 15.6



#86

Perylene-d12

Concen: 20.000 ng

RT: 15.468 min Scan# 2274

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:264 Resp: 137026


Ion Ratio Lower Upper

264 100

260 22.9 19.0 28.6

265 21.5 17.0 25.6









#87
 Indeno(1,2,3-cd)pyrene
 Concen: 10.215 ng
 RT: 16.933 min Scan# 2523
 Delta R.T. -0.006 min
 Lab File: BF138682.D
 Acq: 30 Jul 2024 13:56




Instrument :
 BNA_F
 ClientSampleId :
 SSTDICC010


Tgt Ion:276 Resp: 100310
 Ion Ratio Lower Upper
 276 100
 138 27.7 21.8 32.8
 277 25.5 20.6 30.8



#88
 Benzo(b)fluoranthene
 Concen: 10.980 ng
 RT: 15.039 min Scan# 2201
 Delta R.T. 0.000 min
 Lab File: BF138682.D
 Acq: 30 Jul 2024 13:56

Tgt Ion:252 Resp: 93263
 Ion Ratio Lower Upper
 252 100
 253 21.2 17.5 26.3
 125 10.7 8.9 13.3





#89

Benzo(k)fluoranthene

Concen: 9.717 ng

RT: 15.068 min Scan# 2206

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56

Instrument :

BNA_F

ClientSampleId :

SSTDICC010





Tgt Ion:252 Resp: 71464

Ion Ratio Lower Upper

252 100

253 20.8 17.4 26.0

125 10.7 8.6 13.0



#90

Benzo(a)pyrene

Concen: 10.038 ng

RT: 15.404 min Scan# 2263

Delta R.T. 0.000 min

Lab File: BF138682.D

Acq: 30 Jul 2024 13:56




Tgt Ion:252 Resp: 71724


Ion Ratio Lower Upper

252 100

253 22.0 17.3 25.9

125 13.1 9.5 14.3









#91
Dibenzo(a,h)anthracene
Concen: 10.465 ng
RT: 16.945 min Scan# 2526
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Instrument :
BNA_F
ClientSampleId :
SSTDICC010

Tgt Ion:278 Resp: 84356
Ion Ratio Lower Upper
278 100
139 18.9 14.0 21.0
279 22.6 19.0 28.4



#92
Benzo(g,h,i)perylene
Concen: 10.110 ng
RT: 17.374 min Scan# 2598
Delta R.T. -0.006 min
Lab File: BF138682.D
Acq: 30 Jul 2024 13:56

Tgt Ion:276 Resp: 84564
Ion Ratio Lower Upper
276 100
277 24.7 19.0 28.4
138 22.3 18.5 27.7

