

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3475 SAS No.: P3475 SDG NO.: P3475

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/13/2024

Lab File ID: BF138965.D Time Analyzed: 10:08

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	46987	6.839	189694	8.12	108534	9.88
UPPER LIMIT	93974	7.339	379388	8.622	217068	10.375
LOWER LIMIT	23493.5	6.339	94847	7.622	54267	9.375
EPA SAMPLE NO.						
01 PB162520BS	49442	6.84	209908	8.12	120228	9.88

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.