

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3475 SAS No.: P3475 SDG NO.: P3475

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/16/2024

Lab File ID: BF139049.D Time Analyzed: 10:14

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	62968	6.84	249711	8.12	137022	9.88
UPPER LIMIT	125936	7.34	499422	8.622	274044	10.375
LOWER LIMIT	31484	6.34	124856	7.622	68511	9.375
EPA SAMPLE NO.						
01 PB162520BL	46870	6.83	190657	8.12	110837	9.87

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.