

8C

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3475 SAS No.: P3475 SDG NO.: P3475

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/16/2024

Lab File ID: BF139049.D Time Analyzed: 10:14

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS4 (PHN) AREA #	RT #	IS5 (CRY) AREA #	RT #	IS6 (PRY) AREA #	RT #
12 HOUR STD	227673	11.357	105031	14.004	114377	15.468
UPPER LIMIT	455346	11.857	210062	14.504	228754	15.968
LOWER LIMIT	113837	10.857	52515.5	13.504	57188.5	14.968
EPA SAMPLE NO.						
01 PB162520BL	209233	11.36	121458	14.00	88021	15.47

IS4 (PHN) = Phenanthrene-d10

IS5 (CRY) = Chrysene-d12

IS6 (PRY) = Perylene-d12

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.