

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132039

Review By	jaswal	Review On	8/22/2024 9:20:07 AM
Supervise By	Mohan	Supervise On	8/22/2024 9:20:54 AM
STD. NAME	STD REF.#		
ICAL Standard	MP81838,MP81845,MP81843,MP81842,MP81840,MP81839,MP81851		
ICV Standard	MP81846		
CCV Standard	MP81849		
ICSA Standard	MP81847,MP81848		
CRI Standard			
LCS Standard			
Chk Standard	MP81852,MP81853		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	08/15/24 11:23		Kareem	OK
2	S1	S1	CAL2	08/15/24 11:27		Kareem	OK
3	S2	S2	CAL3	08/15/24 11:32		Kareem	OK
4	S3	S3	CAL4	08/15/24 11:36		Kareem	OK
5	S4	S4	CAL5	08/15/24 11:40		Kareem	OK
6	S5	S5	CAL6	08/15/24 11:44		Kareem	OK
7	ICV01	ICV01	ICV	08/15/24 12:24	ICV fail for As,Cd,Mg,Na (200.7)	Kareem	OK
8	LLICV01	LLICV01	LLICV	08/15/24 12:29		Kareem	OK
9	ICB01	ICB01	ICB	08/15/24 12:33		Kareem	OK
10	CRI01	CRI01	CRDL	08/15/24 12:38		Kareem	OK
11	ICSA01	ICSA01	ICSA	08/15/24 12:42		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	08/15/24 12:56		Kareem	OK
13	ICSADL	ICSADL	ICSA	08/15/24 13:00		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	08/15/24 13:11		Kareem	OK
15	CCV01	CCV01	CCV	08/15/24 13:16		Kareem	OK
16	CCB01	CCB01	CCB	08/15/24 13:22		Kareem	OK
17	LR1	LR1	HIGH STD	08/15/24 13:27		Kareem	OK
18	LR2	LR2	HIGH STD	08/15/24 13:32		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132039

Review By	jaswal	Review On	8/22/2024 9:20:07 AM
Supervise By	Mohan	Supervise On	8/22/2024 9:20:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP81838,MP81845,MP81843,MP81842,MP81840,MP81839,MP81851
ICV Standard	MP81846
CCV Standard	MP81849
ICSA Standard	MP81847,MP81848
CRI Standard	
LCS Standard	
Chk Standard	MP81852,MP81853

19	P3519-21	OU4-CF-03-080624	SAM	08/15/24 13:37		Kareem	OK
20	P3519-23	OU4-CF-04-080624	SAM	08/15/24 13:41		Kareem	OK
21	P3519-25	OU4-CF-05-080624	SAM	08/15/24 13:45		Kareem	OK
22	P3519-27	OU4-CF-06-080624	SAM	08/15/24 13:50		Kareem	OK
23	P3519-29	OU4-CF-07-080624	SAM	08/15/24 13:54		Kareem	OK
24	P3548-12	IN-SITU-SOIL	SAM	08/15/24 13:58		Kareem	OK
25	P3548-12DUP	IN-SITU-SOILDUP	DUP	08/15/24 14:02		Kareem	OK
26	P3548-12L	IN-SITU-SOILL	SD	08/15/24 14:06		Kareem	OK
27	CCV02	CCV02	CCV	08/15/24 14:10		Kareem	OK
28	CCB02	CCB02	CCB	08/15/24 14:15		Kareem	OK
29	P3548-12MS	IN-SITU-SOILMS	MS	08/15/24 14:19		Kareem	OK
30	P3548-12MSD	IN-SITU-SOILMSD	MSD	08/15/24 14:23		Kareem	OK
31	P3548-12A	IN-SITU-SOILA	PS	08/15/24 14:27	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
32	P3520-02	OU4-CF-08-080624	SAM	08/15/24 14:31	NOT USE	Kareem	Not Ok
33	P3520-04	OU4-CF-09-080624	SAM	08/15/24 14:36	NOT USE	Kareem	Not Ok
34	P3520-06	OU4-CF-10-080624	SAM	08/15/24 14:40	NOT USE	Kareem	Not Ok
35	P3520-08	OU4-CF-11-080624	SAM	08/15/24 14:44	NOT USE	Kareem	Not Ok
36	P3520-10	OU4-CF-12-080624	SAM	08/15/24 14:48	NOT USE	Kareem	Not Ok
37	P3520-12	OU4-CF-13-080624	SAM	08/15/24 14:52	NOT USE	Kareem	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132039

Review By	jaswal	Review On	8/22/2024 9:20:07 AM
Supervise By	Mohan	Supervise On	8/22/2024 9:20:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP81838,MP81845,MP81843,MP81842,MP81840,MP81839,MP81851
ICV Standard	MP81846
CCV Standard	MP81849
ICSA Standard	MP81847,MP81848
CRI Standard	
LCS Standard	
Chk Standard	MP81852,MP81853

38	P3520-14	OU4-CF-14-080624	SAM	08/15/24 14:57	NOT USE	Kareem	Not Ok
39	CCV03	CCV03	CCV	08/15/24 15:01		Kareem	OK
40	CCB03	CCB03	CCB	08/15/24 15:05		Kareem	OK
41	P3520-16	OU4-VSL-01-080624	SAM	08/15/24 15:10	NOT USE	Kareem	Not Ok
42	P3520-18	OU4-57S-01-080624	SAM	08/15/24 15:14	NOT USE	Kareem	Not Ok
43	P3574-02	WC-W-06-202408	SAM	08/15/24 15:18		Kareem	OK
44	CCV04	CCV04	CCV	08/15/24 15:23		Kareem	OK
45	CCB04	CCB04	CCB	08/15/24 15:27		Kareem	OK
46	P3518-01	OU4-TS-01-080724	SAM	08/15/24 15:31		Kareem	OK
47	P3518-03	OU4-TS-02-080724	SAM	08/15/24 15:36		Kareem	OK
48	P3518-05	OU4-TS-03-080724	SAM	08/15/24 15:40		Kareem	OK
49	P3518-07	OU4-TS-04-080724	SAM	08/15/24 15:44		Kareem	OK
50	P3518-09	OU4-TS-05-080724	SAM	08/15/24 15:48		Kareem	OK
51	PB162626BS	PB162626BS	LCS	08/15/24 15:57	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
52	P3518-11	OU4-TS-06-080724	SAM	08/15/24 16:01		Kareem	OK
53	P3518-13	OU4-TS-07-080724	SAM	08/15/24 16:05		Kareem	OK
54	PB162626BL	PB162626BL	MB	08/15/24 16:15		Kareem	OK
55	P3518-15	OU4-TS-08-080724	SAM	08/15/24 16:19		Kareem	OK
56	CCV05	CCV05	CCV	08/15/24 16:23		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132039

Review By	jaswal	Review On	8/22/2024 9:20:07 AM
Supervise By	Mohan	Supervise On	8/22/2024 9:20:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP81838,MP81845,MP81843,MP81842,MP81840,MP81839,MP81851
ICV Standard	MP81846
CCV Standard	MP81849
ICSA Standard	MP81847,MP81848
CRI Standard	
LCS Standard	
Chk Standard	MP81852,MP81853

57	CCB05	CCB05	CCB	08/15/24 16:27		Kareem	OK
58	P3518-17	OU4-TS-09-080724	SAM	08/15/24 16:32		Kareem	OK
59	P3518-19	OU4-TS-10-080724	SAM	08/15/24 16:36		Kareem	OK
60	P3518-21	OU4-TS-11-080724	SAM	08/15/24 16:40		Kareem	OK
61	P3518-23	OU4-TS-12-080724	SAM	08/15/24 16:44		Kareem	OK
62	P3596-01	918-J-WS-081324	SAM	08/15/24 16:49		Kareem	OK
63	P3596-01DUP	918-J-WS-081324DU	DUP	08/15/24 16:53		Kareem	OK
64	P3596-01L	918-J-WS-081324L	SD	08/15/24 16:57		Kareem	OK
65	P3596-01MS	918-J-WS-081324MS	MS	08/15/24 17:02		Kareem	OK
66	P3596-01MSD	918-J-WS-081324MS	MSD	08/15/24 17:06		Kareem	OK
67	P3596-01A	918-J-WS-081324A	PS	08/15/24 17:10	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
68	CCV06	CCV06	CCV	08/15/24 17:14		Kareem	OK
69	CCB06	CCB06	CCB	08/15/24 17:18		Kareem	OK
70	P3596-02	918-J-WS-081324-FD	SAM	08/15/24 17:23		Kareem	OK
71	P3518-25	OU4-TS-13-080724	SAM	08/15/24 17:27		Kareem	OK
72	P3518-27	OU4-TS-14-080724	SAM	08/15/24 17:31		Kareem	OK
73	P3518-29	OU4-TS-15-080724	SAM	08/15/24 17:35		Kareem	OK
74	P3549-01	SU-04-080924	SAM	08/15/24 17:40	NOT USE - MS,MSD,SD,A missing	Kareem	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132039

Review By	jaswal	Review On	8/22/2024 9:20:07 AM
Supervise By	Mohan	Supervise On	8/22/2024 9:20:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP81838,MP81845,MP81843,MP81842,MP81840,MP81839,MP81851
ICV Standard	MP81846
CCV Standard	MP81849
ICSA Standard	MP81847,MP81848
CRI Standard	
LCS Standard	
Chk Standard	MP81852,MP81853

75	P3549-01DUP	SU-04-080924DUP	DUP	08/15/24 17:44	NOT USE - MS,MSD,SD,A missing	Kareem	Not Ok
76	P3520-01	OU4-CF-08-080624	SAM	08/15/24 17:48		Kareem	OK
77	P3520-03	OU4-CF-09-080624	SAM	08/15/24 17:53		Kareem	OK
78	P3520-05	OU4-CF-10-080624	SAM	08/15/24 17:57		Kareem	OK
79	P3520-07	OU4-CF-11-080624	SAM	08/15/24 18:01		Kareem	OK
80	CCV07	CCV07	CCV	08/15/24 18:11		Kareem	OK
81	CCB07	CCB07	CCB	08/15/24 18:16		Kareem	OK
82	P3520-09	OU4-CF-12-080624	SAM	08/15/24 18:20		Kareem	OK
83	P3520-11	OU4-CF-13-080624	SAM	08/15/24 18:24		Kareem	OK
84	P3520-13	OU4-CF-14-080624	SAM	08/15/24 18:29		Kareem	OK
85	P3520-15	OU4-VSL-01-080624	SAM	08/15/24 18:33		Kareem	OK
86	P3520-17	OU4-57S-01-080624	SAM	08/15/24 18:37		Kareem	OK
87	P3386-02	EFF-WW	SAM	08/15/24 18:41		Kareem	OK
88	P3386-02DUP	EFF-WWDUP	DUP	08/15/24 18:46		Kareem	OK
89	P3386-02L	EFF-WWL	SD	08/15/24 18:50		Kareem	OK
90	P3386-02MS	EFF-WWMS	MS	08/15/24 18:54		Kareem	OK
91	P3386-02MSD	EFF-WWMSD	MSD	08/15/24 18:59		Kareem	OK
92	CCV08	CCV08	CCV	08/15/24 19:03		Kareem	OK
93	CCB08	CCB08	CCB	08/15/24 19:07		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132039

Review By	jaswal	Review On	8/22/2024 9:20:07 AM
Supervise By	Mohan	Supervise On	8/22/2024 9:20:54 AM

STD. NAME	STD REF.#
ICAL Standard	MP81838,MP81845,MP81843,MP81842,MP81840,MP81839,MP81851
ICV Standard	MP81846
CCV Standard	MP81849
ICSA Standard	MP81847,MP81848
CRI Standard	
LCS Standard	
Chk Standard	MP81852,MP81853

94	P3386-02A	EFF-WWA	PS	08/15/24 19:11	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
95	PB162768BL	PB162768BL	MB	08/15/24 19:16		Kareem	OK
96	PB162768BS	PB162768BS	LCS	08/15/24 19:20	0.1 ML OF M6001 AND M6010 WERE ADDED TO 10 ML OF THE SAMPLE	Kareem	OK
97	CCV09	CCV09	CCV	08/15/24 19:35	CCV fail for K,Na	Kareem	OK
98	CCB09	CCB09	CCB	08/15/24 19:40		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132142

Review By	Jaswal	Review On	8/24/2024 8:00:48 AM
Supervise By	Mohan	Supervise On	8/24/2024 8:02:39 AM
STD. NAME	STD REF.#		
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986		
ICV Standard	MP81993		
CCV Standard	MP81996		
ICSA Standard	MP81994,MP81995		
CRI Standard	MP81992,MP81998,MP81999		
LCS Standard			
Chk Standard	MP82001,MP82000		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	08/23/24 09:27		Kareem	OK
2	S1	S1	CAL2	08/23/24 09:31		Kareem	OK
3	S2	S2	CAL3	08/23/24 09:36		Kareem	OK
4	S3	S3	CAL4	08/23/24 09:40		Kareem	OK
5	S4	S4	CAL5	08/23/24 09:44		Kareem	OK
6	S5	S5	CAL6	08/23/24 09:48		Kareem	OK
7	ICV01	ICV01	ICV	08/23/24 09:52	ICV fail for B,Mo,Sr,Sn,Ti (6010D)	Kareem	OK
8	LLICV01	LLICV01	LLICV	08/23/24 09:56		Kareem	OK
9	ICB01	ICB01	ICB	08/23/24 10:01		Kareem	OK
10	CRI01	CRI01	CRDL	08/23/24 10:05		Kareem	OK
11	ICSA01	ICSA01	ICSA	08/23/24 10:09		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	08/23/24 10:14	ICSAB fail for B,Mo,Sn	Kareem	OK
13	CCV01	CCV01	CCV	08/23/24 10:22		Kareem	OK
14	CCB01	CCB01	CCB	08/23/24 10:27		Kareem	OK
15	LR1	LR1	HIGH STD	08/23/24 10:31		Kareem	OK
16	P3636-01	OILY-DEBRIS	SAM	08/23/24 10:42		Kareem	OK
17	P3669-03	1027	SAM	08/23/24 10:46		Kareem	OK
18	PB162682TB	PB162682TB	MB	08/23/24 10:50		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132142

Review By	Jaswal	Review On	8/24/2024 8:00:48 AM
Supervise By	Mohan	Supervise On	8/24/2024 8:02:39 AM
STD. NAME	STD REF.#		
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986		
ICV Standard	MP81993		
CCV Standard	MP81996		
ICSA Standard	MP81994,MP81995		
CRI Standard	MP81992,MP81998,MP81999		
LCS Standard			
Chk Standard	MP82001,MP82000		

19	LR2	LR2	HIGH STD	08/23/24 10:56		Kareem	OK
20	PB162724BL	PB162724BL	MB	08/23/24 11:01		Kareem	OK
21	PB162724BS	PB162724BS	LCS	08/23/24 11:06	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
22	PB162726BL	PB162726BL	MB	08/23/24 11:10	NOT USE - ICV fail for B,Mo,Sr,Sn,Ti (6010D)	Kareem	Not Ok
23	P3669-05	20430	SAM	08/23/24 11:14		Kareem	OK
24	CCV02	CCV02	CCV	08/23/24 11:19		Kareem	OK
25	CCB02	CCB02	CCB	08/23/24 11:23		Kareem	OK
26	PB162726BS	PB162726BS	LCS	08/23/24 11:28	NOT USE - ICV fail for B,Mo,Sr,Sn,Ti (6010D)	Kareem	Not Ok
27	PB162556TBRE	PB162556TBRE	MB	08/23/24 11:32	NOT USE	Kareem	Not Ok
28	P3461-04	P3461-01 - CHRT-286	SAM	08/23/24 11:36		Kareem	OK
29	P3669-01	20430	SAM	08/23/24 11:40		Kareem	OK
30	PB162869BL	PB162869BL	MB	08/23/24 11:45		Kareem	OK
31	PB162869BS	PB162869BS	LCS	08/23/24 11:49	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
32	P3684-03	D3646	SAM	08/23/24 11:53		Kareem	OK
33	P3684-04	D3638	SAM	08/23/24 11:58		Kareem	OK
34	P3684-06	D3564	SAM	08/23/24 12:02		Kareem	OK
35	P3685-02	Stock Pile-1	SAM	08/23/24 12:07		Kareem	OK
36	CCV03	CCV03	CCV	08/23/24 12:11		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132142

Review By	Jaswal	Review On	8/24/2024 8:00:48 AM
Supervise By	Mohan	Supervise On	8/24/2024 8:02:39 AM

STD. NAME	STD REF.#
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986
ICV Standard	MP81993
CCV Standard	MP81996
ICSA Standard	MP81994,MP81995
CRI Standard	MP81992,MP81998,MP81999
LCS Standard	
Chk Standard	MP82001,MP82000

37	CCB03	CCB03	CCB	08/23/24 12:15		Kareem	OK
38	P3704-03A	LNG-Soil PileA	PS	08/23/24 12:20	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	Not Ok
39	PB162890TB	PB162890TB	MB	08/23/24 12:24	NOT USE	Kareem	Not Ok
40	P3580-08	922-K1-SD-0.5-1-0811	SAM	08/23/24 12:41	NOT USE	Kareem	Not Ok
41	P3689-01	WC-1	SAM	08/23/24 12:45		Kareem	OK
42	P3694-02	Kingsland-Driveway	SAM	08/23/24 12:49		Kareem	OK
43	P3699-01DL	001-CONTAINER-WC	SAM	08/23/24 12:54	Straight 25x for all elements	Kareem	OK
44	P3699-02	002-CARPENTER-WC	SAM	08/23/24 12:58		Kareem	OK
45	P3650-01	RW7-SP100-20240811	SAM	08/23/24 13:03		Kareem	OK
46	PB162907BL	PB162907BL	MB	08/23/24 13:07		Kareem	OK
47	P3650-04	RW7-SP303-20240811	SAM	08/23/24 13:11		Kareem	OK
48	CCV04	CCV04	CCV	08/23/24 13:16		Kareem	OK
49	CCB04	CCB04	CCB	08/23/24 13:20		Kareem	OK
50	P3685-04	Stock Pile-2	SAM	08/23/24 13:24		Kareem	OK
51	P3685-01	Stock Pile-1	SAM	08/23/24 13:29		Kareem	OK
52	PB162907BS	PB162907BS	LCS	08/23/24 13:33	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
53	P3685-03	Stock Pile-2	SAM	08/23/24 13:37		Kareem	OK
54	P3704-03	LNG-Soil Pile	SAM	08/23/24 13:41	Only for Fax	Kareem	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132142

Review By	Jaswal	Review On	8/24/2024 8:00:48 AM
Supervise By	Mohan	Supervise On	8/24/2024 8:02:39 AM

STD. NAME	STD REF.#
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986
ICV Standard	MP81993
CCV Standard	MP81996
ICSA Standard	MP81994,MP81995
CRI Standard	MP81992,MP81998,MP81999
LCS Standard	
Chk Standard	MP82001,MP82000

55	P3704-03L	LNG-Soil PileL	SD	08/23/24 13:46	SD fail for many elements	Kareem	Not Ok
56	P3704-03DUP	LNG-Soil PileDUP	DUP	08/23/24 13:50	SD fail for many elements	Kareem	Not Ok
57	P3704-03MS	LNG-Soil PileMS	MS	08/23/24 13:55	SD fail for many elements	Kareem	Not Ok
58	P3704-03MSD	LNG-Soil PileMSD	MSD	08/23/24 13:59	SD fail for many elements	Kareem	Not Ok
59	P3685-04DUP	Stock Pile-2DUP	DUP	08/23/24 14:03		Kareem	OK
60	CCV05	CCV05	CCV	08/23/24 14:08		Kareem	OK
61	CCB05	CCB05	CCB	08/23/24 14:12		Kareem	OK
62	PB162556TB	PB162556TB	MB	08/23/24 14:16		Kareem	OK
63	P3685-04L	Stock Pile-2L	SD	08/23/24 14:21		Kareem	OK
64	P3669-01DUP	20430DUP	DUP	08/23/24 14:25		Kareem	OK
65	P3669-01L	20430L	SD	08/23/24 14:29		Kareem	OK
66	P3669-01MS	20430MS	MS	08/23/24 14:34		Kareem	OK
67	P3669-01MSD	20430MSD	MSD	08/23/24 14:38		Kareem	OK
68	P3669-01A	20430A	PS	08/23/24 14:42	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
69	P3686-02	DNB-47-COMPOSITE	SAM	08/23/24 14:46		Kareem	OK
70	P3685-04MS	Stock Pile-2MS	MS	08/23/24 14:50		Kareem	OK
71	P3685-04MSD	Stock Pile-2MSD	MSD	08/23/24 14:54		Kareem	OK
72	CCV06	CCV06	CCV	08/23/24 14:59		Kareem	OK
73	CCB06	CCB06	CCB	08/23/24 15:03		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132142

Review By	Jaswal	Review On	8/24/2024 8:00:48 AM
Supervise By	Mohan	Supervise On	8/24/2024 8:02:39 AM
STD. NAME	STD REF.#		
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986		
ICV Standard	MP81993		
CCV Standard	MP81996		
ICSA Standard	MP81994,MP81995		
CRI Standard	MP81992,MP81998,MP81999		
LCS Standard			
Chk Standard	MP82001,MP82000		

74	P3685-04A	Stock Pile-2A	PS	08/23/24 15:07	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
75	P3684-08	T3650	SAM	08/23/24 15:11		Kareem	OK
76	P3684-09DL	D3643DL	SAM	08/23/24 15:16	Straight 25x for all elements	Kareem	OK
77	P3684-10	D3637	SAM	08/23/24 15:20		Kareem	OK
78	PB162871TB	PB162871TB	MB	08/23/24 15:24		Kareem	OK
79	PB162873TB	PB162873TB	MB	08/23/24 15:29		Kareem	OK
80	PB162889BS	PB162889BS	LCS	08/23/24 15:38	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
81	PB162889BL	PB162889BL	MB	08/23/24 15:43		Kareem	OK
82	P3669-03DUP	1027DUP	DUP	08/23/24 15:47		Kareem	OK
83	P3669-03L	1027L	SD	08/23/24 15:51		Kareem	OK
84	CCV07	CCV07	CCV	08/23/24 15:56		Kareem	OK
85	CCB07	CCB07	CCB	08/23/24 16:00		Kareem	OK
86	P3669-03MS	1027MS	MS	08/23/24 16:04		Kareem	OK
87	P3669-03MSD	1027MSD	MSD	08/23/24 16:08		Kareem	OK
88	P3669-03A	1027A	PS	08/23/24 16:13	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
89	PB162615BL	PB162615BL	MB	08/23/24 16:17		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132142

Review By	Jaswal	Review On	8/24/2024 8:00:48 AM
Supervise By	Mohan	Supervise On	8/24/2024 8:02:39 AM

STD. NAME	STD REF.#
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986
ICV Standard	MP81993
CCV Standard	MP81996
ICSA Standard	MP81994,MP81995
CRI Standard	MP81992,MP81998,MP81999
LCS Standard	
Chk Standard	MP82001,MP82000

90	PB162615BS	PB162615BS	LCS	08/23/24 16:21	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
91	PB162625BL	PB162625BL	MB	08/23/24 16:25	fail for TI	Kareem	Not Ok
92	PB162870BL	PB162870BL	MB	08/23/24 16:30		Kareem	OK
93	PB162870BS	PB162870BS	LCS	08/23/24 16:34	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
94	PB162625BS	PB162625BS	LCS	08/23/24 16:38	fail for TI	Kareem	Not Ok
95	PB162766BL	PB162766BL	MB	08/23/24 16:43		Kareem	OK
96	CCV08	CCV08	CCV	08/23/24 16:47		Kareem	OK
97	CCB08	CCB08	CCB	08/23/24 16:51		Kareem	OK
98	PB162766BS	PB162766BS	LCS	08/23/24 16:55	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
99	PB162789BL	PB162789BL	MB	08/23/24 17:00		Kareem	OK
100	PB162789BS	PB162789BS	LCS	08/23/24 17:04	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
101	PB162765BL	PB162765BL	MB	08/23/24 17:08		Kareem	OK
102	PB162765BS	PB162765BS	LCS	08/23/24 17:12	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
103	PB162830BL	PB162830BL	MB	08/23/24 17:17		Kareem	OK
104	PB162830BS	PB162830BS	LCS	08/23/24 17:26	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132142

Review By	Jaswal	Review On	8/24/2024 8:00:48 AM
Supervise By	Mohan	Supervise On	8/24/2024 8:02:39 AM
STD. NAME	STD REF.#		
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986		
ICV Standard	MP81993		
CCV Standard	MP81996		
ICSA Standard	MP81994,MP81995		
CRI Standard	MP81992,MP81998,MP81999		
LCS Standard			
Chk Standard	MP82001,MP82000		

105	PB162818BL	PB162818BL	MB	08/23/24 17:31		Kareem	OK
106	PB162818BS	PB162818BS	LCS	08/23/24 17:35	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
107	PB162683TB	PB162683TB	MB	08/23/24 17:46		Kareem	OK
108	CCV09	CCV09	CCV	08/23/24 18:27	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
109	CCB09	CCB09	CCB	08/23/24 18:31		Kareem	OK
110	PB162725BL	PB162725BL	MB	08/23/24 18:36		Kareem	OK
111	PB162725BS	PB162725BS	LCS	08/23/24 18:40	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
112	PB162806TB	PB162806TB	MB	08/23/24 18:44		Kareem	OK
113	P3573-12	EB-01-080924	SAM	08/23/24 18:49		Kareem	OK
114	P3475-05	S-858-K1-SO-1.0-1.5	SAM	08/23/24 18:53	NOT USE	Kareem	Not Ok
115	P3475-05DUP	S-858-K1-SO-1.0-1.5	DUP	08/23/24 18:57	NOT USE	Kareem	Not Ok
116	P3475-05L	S-858-K1-SO-1.0-1.5	SD	08/23/24 19:01	NOT USE	Kareem	Not Ok
117	P3475-05MS	S-858-K1-SO-1.0-1.5	MS	08/23/24 19:06	NOT USE	Kareem	Not Ok
118	P3475-05MSD	S-858-K1-SO-1.0-1.5	MSD	08/23/24 19:10	NOT USE	Kareem	Not Ok
119	P3475-05A	S-858-K1-SO-1.0-1.5	PS	08/23/24 19:14	NOT USE	Kareem	Not Ok
120	CCV10	CCV10	CCV	08/23/24 19:23		Kareem	OK
121	CCB10	CCB10	CCB	08/23/24 19:27		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132142

Review By	Jaswal	Review On	8/24/2024 8:00:48 AM
Supervise By	Mohan	Supervise On	8/24/2024 8:02:39 AM

STD. NAME	STD REF.#
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986
ICV Standard	MP81993
CCV Standard	MP81996
ICSA Standard	MP81994,MP81995
CRI Standard	MP81992,MP81998,MP81999
LCS Standard	
Chk Standard	MP82001,MP82000

122	P3526-01	SP-1-COMP	SAM	08/23/24 19:31		Kareem	OK
123	P3526-01DUP	SP-1-COMPDUP	DUP	08/23/24 19:35		Kareem	OK
124	P3526-01L	SP-1-COMPL	SD	08/23/24 19:39		Kareem	OK
125	P3526-01MS	SP-1-COMPMS	MS	08/23/24 19:44		Kareem	OK
126	P3526-01MSD	SP-1-COMPMSD	MSD	08/23/24 19:48		Kareem	OK
127	P3526-01A	SP-1-COMPA	PS	08/23/24 19:52	0.1ml of m6100 and m6010 were added to 10 ml of sample	Kareem	OK
128	P3573-01	S-858-KI-SO-2-2.5-08	SAM	08/23/24 19:56		Kareem	OK
129	P3573-02	S-859-KI-SO-2-2.5-08	SAM	08/23/24 20:00		Kareem	OK
130	P3573-03	S-859-KI-SO-2-2.5-08	SAM	08/23/24 20:04		Kareem	OK
131	P3573-04	S-860-KI-SO-2-2.5-08	SAM	08/23/24 20:08		Kareem	OK
132	CCV11	CCV11	CCV	08/23/24 20:47		Kareem	OK
133	CCB11	CCB11	CCB	08/23/24 20:51		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132156

Review By	Jaswal	Review On	8/27/2024 8:04:13 AM
Supervise By	Mohan	Supervise On	8/27/2024 8:05:04 AM

STD. NAME	STD REF.#
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986
ICV Standard	MP81993
CCV Standard	MP81996
ICSA Standard	MP81994,MP81995
CRI Standard	MP81992,MP81998,MP81999
LCS Standard	
Chk Standard	MP82001,MP82000

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	08/24/24 09:57		Kareem	OK
2	S1	S1	CAL2	08/24/24 10:01		Kareem	OK
3	S2	S2	CAL3	08/24/24 10:05		Kareem	OK
4	S3	S3	CAL4	08/24/24 10:10		Kareem	OK
5	S4	S4	CAL5	08/24/24 10:14		Kareem	OK
6	S5	S5	CAL6	08/24/24 10:18		Kareem	OK
7	ICV01	ICV01	ICV	08/24/24 10:27	ICV fail for Na,Sr(95-105)%	Kareem	OK
8	LLICV01	LLICV01	LLICV	08/24/24 10:49		Kareem	OK
9	ICB01	ICB01	ICB	08/24/24 10:53		Kareem	OK
10	CRI01	CRI01	CRDL	08/24/24 11:04		Kareem	OK
11	ICSA01	ICSA01	ICSA	08/24/24 11:08		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	08/24/24 11:33		Kareem	OK
13	ICSADL	ICSADL	ICSA	08/24/24 11:37		Kareem	OK
14	ICSABDL	ICSABDL	ICSAB	08/24/24 11:42		Kareem	OK
15	CCV01	CCV01	CCV	08/24/24 11:46		Kareem	OK
16	CCB01	CCB01	CCB	08/24/24 11:50		Kareem	OK
17	LR1	LR1	HIGH STD	08/24/24 11:55		Kareem	OK
18	PB162726BL	PB162726BL	MB	08/24/24 12:05		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132156

Review By	Jaswal	Review On	8/27/2024 8:04:13 AM
Supervise By	Mohan	Supervise On	8/27/2024 8:05:04 AM
STD. NAME	STD REF.#		
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986		
ICV Standard	MP81993		
CCV Standard	MP81996		
ICSA Standard	MP81994,MP81995		
CRI Standard	MP81992,MP81998,MP81999		
LCS Standard			
Chk Standard	MP82001,MP82000		

19	PB162726BS	PB162726BS	LCS	08/24/24 12:09	0.1 ML M6001 AND M6010 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
20	LR2	LR2	HIGH STD	08/24/24 12:13		Kareem	OK
21	PB162625BL	PB162625BL	MB	08/24/24 12:19		Kareem	OK
22	PB162625BS	PB162625BS	LCS	08/24/24 12:23	0.1 ML M6001 AND M6010 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
23	P3622-01	028-OUTFALL	SAM	08/24/24 12:27		Kareem	OK
24	P3687-05	001-WILLETS-PT-BL	SAM	08/24/24 12:31		Kareem	OK
25	P3687-06	002-35TH-AVE(AUG)	SAM	08/24/24 12:36		Kareem	OK
26	P3687-06DUP	002-35TH-AVE(AUG)	DUP	08/24/24 12:40		Kareem	OK
27	CCV02	CCV02	CCV	08/24/24 12:49		Kareem	OK
28	CCB02	CCB02	CCB	08/24/24 12:53		Kareem	OK
29	P3687-06L	002-35TH-AVE(AUG)	SD	08/24/24 12:57		Kareem	OK
30	P3687-06MS	002-35TH-AVE(AUG)	MS	08/24/24 13:01		Kareem	OK
31	P3687-06MSD	002-35TH-AVE(AUG)	MSD	08/24/24 13:05		Kareem	OK
32	P3687-06A	002-35TH-AVE(AUG)	PS	08/24/24 13:09	0.1 ML M6001 AND M6010 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
33	PB162912BL	PB162912BL	MB	08/24/24 13:14		Kareem	OK
34	PB162912BS	PB162912BS	LCS	08/24/24 13:18	0.1 ML M6001 AND M6010 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132156

Review By	Jaswal	Review On	8/27/2024 8:04:13 AM
Supervise By	Mohan	Supervise On	8/27/2024 8:05:04 AM
STD. NAME	STD REF.#		
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986		
ICV Standard	MP81993		
CCV Standard	MP81996		
ICSA Standard	MP81994,MP81995		
CRI Standard	MP81992,MP81998,MP81999		
LCS Standard			
Chk Standard	MP82001,MP82000		

35	P3645-01	914-J-WS-081524	SAM	08/24/24 13:22		Kareem	OK
36	P3645-02	916-J-WS-081524	SAM	08/24/24 13:26		Kareem	OK
37	P3656-01	MW-1	SAM	08/24/24 13:31		Kareem	OK
38	P3656-02	MW-1	SAM	08/24/24 13:35		Kareem	OK
39	CCV03	CCV03	CCV	08/24/24 13:44		Kareem	OK
40	CCB03	CCB03	CCB	08/24/24 13:48		Kareem	OK
41	P3656-03	MW-3	SAM	08/24/24 13:52		Kareem	OK
42	P3656-04	MW-3	SAM	08/24/24 13:57		Kareem	OK
43	P3656-05	MW-4	SAM	08/24/24 14:01		Kareem	OK
44	P3656-06	MW-4	SAM	08/24/24 14:06		Kareem	OK
45	P3656-07	MW-2	SAM	08/24/24 14:10		Kareem	OK
46	P3656-08	MW-2	SAM	08/24/24 14:14		Kareem	OK
47	P3657-01	917-J-WS-081624	SAM	08/24/24 14:19		Kareem	OK
48	P3657-01DUP	917-J-WS-081624DUP	DUP	08/24/24 14:23		Kareem	OK
49	P3657-01L	917-J-WS-081624L	SD	08/24/24 14:27		Kareem	OK
50	P3657-01MS	917-J-WS-081624MS	MS	08/24/24 14:31		Kareem	OK
51	CCV04	CCV04	CCV	08/24/24 14:36		Kareem	OK
52	CCB04	CCB04	CCB	08/24/24 14:40		Kareem	OK
53	P3657-01MSD	917-J-WS-081624MSD	MSD	08/24/24 14:44		Kareem	OK
54	P3657-01A	917-J-WS-081624A	PS	08/24/24 14:48		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132156

Review By	Jaswal	Review On	8/27/2024 8:04:13 AM
Supervise By	Mohan	Supervise On	8/27/2024 8:05:04 AM
STD. NAME	STD REF.#		
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986		
ICV Standard	MP81993		
CCV Standard	MP81996		
ICSA Standard	MP81994,MP81995		
CRI Standard	MP81992,MP81998,MP81999		
LCS Standard			
Chk Standard	MP82001,MP82000		

55	PB162819BL	PB162819BL	MB	08/24/24 14:52		Kareem	OK
56	PB162819BS	PB162819BS	LCS	08/24/24 14:56	0.1 ML M6001 AND M6010 WERE ADDED TO 10 ML OF SAMPLE	Kareem	OK
57	P3579-02	VCPS-B1-081224-0-1	SAM	08/24/24 15:01		Kareem	OK
58	P3580-01	926-K1-SD-0-0.5-0811	SAM	08/24/24 15:05	Method change on log in page	Kareem	Not Ok
59	P3580-02	926-K1-SD-0.5-1-0811	SAM	08/24/24 15:09	Method change on log in page	Kareem	Not Ok
60	P3597-01	SP-1-COMP	SAM	08/24/24 15:16		Kareem	OK
61	P3600-02	VCPS-B2-081324-0-1	SAM	08/24/24 15:21		Kareem	OK
62	CCV05	CCV05	CCV	08/24/24 15:25		Kareem	OK
63	CCB05	CCB05	CCB	08/24/24 15:29		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132157

Review By	Jaswal	Review On	8/27/2024 8:04:22 AM
Supervise By	Mohan	Supervise On	8/27/2024 8:05:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986		
ICV Standard	MP81993		
CCV Standard	MP81996		
ICSA Standard	MP81994,MP81995		
CRI Standard	MP81992,MP81998,MP81999		
LCS Standard			
Chk Standard	MP82001,MP82000		

Sr#	SampleId	ClientID	QcType	Date	Comment	Operator	Status
1	S0	S0	CAL1	08/25/24 18:22		Kareem	OK
2	S1	S1	CAL2	08/25/24 18:26		Kareem	OK
3	S2	S2	CAL3	08/25/24 18:31		Kareem	OK
4	S3	S3	CAL4	08/25/24 18:35		Kareem	OK
5	S4	S4	CAL5	08/25/24 18:39		Kareem	OK
6	S5	S5	CAL6	08/25/24 18:43		Kareem	OK
7	ICV01	ICV01	ICV	08/25/24 18:47		Kareem	OK
8	LLICV01	LLICV01	LLICV	08/25/24 19:17		Kareem	OK
9	ICB01	ICB01	ICB	08/25/24 19:22		Kareem	OK
10	CRI01	CRI01	CRDL	08/25/24 19:26		Kareem	OK
11	ICSA01	ICSA01	ICSA	08/25/24 19:36		Kareem	OK
12	ICSAB01	ICSAB01	ICSAB	08/25/24 19:40		Kareem	OK
13	CCV01	CCV01	CCV	08/25/24 19:44		Kareem	OK
14	CCB01	CCB01	CCB	08/25/24 19:48		Kareem	OK
15	LR1	LR1	HIGH STD	08/25/24 19:53		Kareem	OK
16	P3600-04	VCPS-B3-081324-0-1	SAM	08/25/24 20:04		Kareem	OK
17	P3580-03	925-K1-SD-0-0.5-081	SAM	08/25/24 20:08	Method changed on login page	Kareem	Not Ok
18	P3580-04	925-K1-SD-0.5-1-081	SAM	08/25/24 20:12	Method changed on login page	Kareem	Not Ok

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132157

Review By	Jaswal	Review On	8/27/2024 8:04:22 AM
Supervise By	Mohan	Supervise On	8/27/2024 8:05:11 AM

STD. NAME	STD REF.#
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986
ICV Standard	MP81993
CCV Standard	MP81996
ICSA Standard	MP81994,MP81995
CRI Standard	MP81992,MP81998,MP81999
LCS Standard	
Chk Standard	MP82001,MP82000

19	P3580-05	923-K1-SD-0-0.5-0811	SAM	08/25/24 20:16	Method changed on login page	Kareem	Not Ok
20	P3580-06	923-K1-SD-0.5-1-0811	SAM	08/25/24 20:21	Method changed on login page	Kareem	Not Ok
21	P3580-07	922-K1-SD-0-0.5-0811	SAM	08/25/24 20:25	Method changed on login page	Kareem	Not Ok
22	P3600-06	VCPS-B3-081324-10-	SAM	08/25/24 20:29		Kareem	OK
23	P3600-06DUP	VCPS-B3-081324-10-	DUP	08/25/24 20:33		Kareem	OK
24	CCV02	CCV02	CCV	08/25/24 20:37		Kareem	OK
25	CCB02	CCB02	CCB	08/25/24 20:41		Kareem	OK
26	P3600-06L	VCPS-B3-081324-10-	SD	08/25/24 20:46		Kareem	OK
27	P3600-06MS	VCPS-B3-081324-10-	MS	08/25/24 20:50		Kareem	OK
28	P3600-06MSD	VCPS-B3-081324-10-	MSD	08/25/24 20:54		Kareem	OK
29	P3600-06A	VCPS-B3-081324-10-	PS	08/25/24 20:58	0.1 of m6001 and m6010 were added to 10 ml of the sample	Kareem	OK
30	P3640-02	SP-1-COMP	SAM	08/25/24 21:02		Kareem	OK
31	P3640-02DUP	SP-1-COMPDUP	DUP	08/25/24 21:07		Kareem	OK
32	P3640-02L	SP-1-COMPL	SD	08/25/24 21:11		Kareem	OK
33	P3640-02MS	SP-1-COMPMS	MS	08/25/24 21:15		Kareem	OK
34	P3640-02MSD	SP-1-COMPMSD	MSD	08/25/24 21:20		Kareem	OK
35	P3640-02A	SP-1-COMPA	PS	08/25/24 21:24	0.1 of m6001 and m6010 were added to 10 ml of the sample	Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132157

Review By	Jaswal	Review On	8/27/2024 8:04:22 AM
Supervise By	Mohan	Supervise On	8/27/2024 8:05:11 AM
STD. NAME	STD REF.#		
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986		
ICV Standard	MP81993		
CCV Standard	MP81996		
ICSA Standard	MP81994,MP81995		
CRI Standard	MP81992,MP81998,MP81999		
LCS Standard			
Chk Standard	MP82001,MP82000		

36	CCV03	CCV03	CCV	08/25/24 21:28		Kareem	OK
37	CCB03	CCB03	CCB	08/25/24 21:37		Kareem	OK
38	LR2	LR2	HIGH STD	08/25/24 21:42		Kareem	OK
39	P3607-01	NB-08-08142024	SAM	08/25/24 21:47		Kareem	OK
40	P3617-07	VCPS-B4-081424-0-1	SAM	08/25/24 21:51		Kareem	OK
41	P3617-08	VCPS-B4-081424-10-	SAM	08/25/24 21:56		Kareem	OK
42	P3617-12	VCPS-B5-081424-0-1	SAM	08/25/24 22:00		Kareem	OK
43	P3617-13	VCPS-B5-081424-10-	SAM	08/25/24 22:04		Kareem	OK
44	P3634-01	TR-06-08152024	SAM	08/25/24 22:08		Kareem	OK
45	P3637-01	ETGI - 370	SAM	08/25/24 22:12		Kareem	OK
46	P3652-01	LIVONIA-SOIL	SAM	08/25/24 22:16	Zn high	Kareem	Dilution
47	P3652-01DUP	LIVONIA-SOILDUP	DUP	08/25/24 22:21	NOT USE	Kareem	Not Ok
48	CCV04	CCV04	CCV	08/25/24 22:26		Kareem	OK
49	CCB04	CCB04	CCB	08/25/24 22:30		Kareem	OK
50	P3652-01L	LIVONIA-SOILL	SD	08/25/24 22:34	NOT USE	Kareem	Not Ok
51	P3652-01MS	LIVONIA-SOILMS	MS	08/25/24 22:38	NOT USE	Kareem	Not Ok
52	P3652-01MSD	LIVONIA-SOILMSD	MSD	08/25/24 22:43	NOT USE	Kareem	Not Ok
53	P3652-01A	LIVONIA-SOILA	PS	08/25/24 22:47	NOT USE	Kareem	Not Ok
54	P3637-02	ETGI - 370	SAM	08/25/24 22:52		Kareem	OK
55	P3637-02DUP	ETGI - 370DUP	DUP	08/25/24 22:56		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132157

Review By	Jaswal	Review On	8/27/2024 8:04:22 AM
Supervise By	Mohan	Supervise On	8/27/2024 8:05:11 AM

STD. NAME	STD REF.#
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986
ICV Standard	MP81993
CCV Standard	MP81996
ICSA Standard	MP81994,MP81995
CRI Standard	MP81992,MP81998,MP81999
LCS Standard	
Chk Standard	MP82001,MP82000

56	P3637-02L	ETGI - 370L	SD	08/25/24 23:01		Kareem	OK
57	P3637-02MS	ETGI - 370MS	MS	08/25/24 23:05		Kareem	OK
58	P3637-02MSD	ETGI - 370MSD	MSD	08/25/24 23:09		Kareem	OK
59	P3637-02A	ETGI - 370A	PS	08/25/24 23:14	0.1 of m6001 and m6010 were added to 10 ml of the sample	Kareem	OK
60	CCV05	CCV05	CCV	08/25/24 23:28		Kareem	OK
61	CCB05	CCB05	CCB	08/25/24 23:33		Kareem	OK
62	P3652-03	P3652-01 - LIVONIA-3	SAM	08/25/24 23:37		Kareem	OK
63	P3609-01	915-J-WS-081424	SAM	08/25/24 23:41		Kareem	OK
64	P3609-01DUP	915-J-WS-081424DUP	DUP	08/25/24 23:46		Kareem	OK
65	P3609-01L	915-J-WS-081424L	SD	08/25/24 23:50		Kareem	OK
66	P3609-01MS	915-J-WS-081424MS	MS	08/25/24 23:54		Kareem	OK
67	P3609-01MSD	915-J-WS-081424MSD	MSD	08/25/24 23:58		Kareem	OK
68	P3609-01A	915-J-WS-081424A	PS	08/26/24 00:02		Kareem	OK
69	P3609-02	920-J-WS-081424	SAM	08/26/24 00:07		Kareem	OK
70	P3617-02	Field Blank	SAM	08/26/24 00:11		Kareem	OK
71	P3617-03	VCPS-B2-081424-6W	SAM	08/26/24 00:15		Kareem	OK
72	CCV06	CCV06	CCV	08/26/24 00:39		Kareem	OK
73	CCB06	CCB06	CCB	08/26/24 00:44		Kareem	OK
74	P3617-04	VCPS-B2-081424-6W	SAM	08/26/24 00:48		Kareem	OK

Instrument ID: P4

Daily Analysis Runlog For Sequence/QC Batch ID # LB132157

Review By	Jaswal	Review On	8/27/2024 8:04:22 AM
Supervise By	Mohan	Supervise On	8/27/2024 8:05:11 AM

STD. NAME	STD REF.#
ICAL Standard	MP81985,MP81990,MP81989,MP81988,MP81987,MP81986
ICV Standard	MP81993
CCV Standard	MP81996
ICSA Standard	MP81994,MP81995
CRI Standard	MP81992,MP81998,MP81999
LCS Standard	
Chk Standard	MP82001,MP82000

75	P3617-14	VCPS-B2-081424-6W	SAM	08/26/24 00:52		Kareem	OK
76	P3617-15	VCPS-B2-081424-6W	SAM	08/26/24 00:56		Kareem	OK
77	P3619-01	917-J-WS-081424	SAM	08/26/24 01:01	Method changed on login page	Kareem	Not Ok
78	P3688-01	MIXED-DEMO	SAM	08/26/24 01:05		Kareem	OK
79	P3689-02	WC-1	SAM	08/26/24 01:09		Kareem	OK
80	P3694-01	Kingsland-Driveway	SAM	08/26/24 01:14		Kareem	OK
81	P3698-01DL	HD-01-8-21-2024DL	SAM	08/26/24 01:18		Kareem	OK
82	P3703-01	OR-03-8-21-2024	SAM	08/26/24 01:22		Kareem	OK
83	P3704-01	LNG-Soil Pile	SAM	08/26/24 01:26	Cr, Fe high	Kareem	Dilution
84	CCV07	CCV07	CCV	08/26/24 01:31		Kareem	OK
85	CCB07	CCB07	CCB	08/26/24 01:35		Kareem	OK
86	P3704-01DUP	LNG-Soil PileDUP	DUP	08/26/24 01:39	Cr, Fe high	Kareem	Dilution
87	P3704-01L	LNG-Soil PileL	SD	08/26/24 01:44		Kareem	OK
88	P3704-01MS	LNG-Soil PileMS	MS	08/26/24 01:48	Cr, Fe high	Kareem	Dilution
89	P3704-01MSD	LNG-Soil PileMSD	MSD	08/26/24 01:53	Cr, Fe high	Kareem	Dilution
90	P3704-01A	LNG-Soil PileA	PS	08/26/24 02:02	0.1 of m6001 and m6010 were added to 10 ml of the sample	Kareem	Dilution
91	CCV08	CCV08	CCV	08/26/24 02:07		Kareem	OK
92	CCB08	CCB08	CCB	08/26/24 02:11		Kareem	OK