



284 Sheffield Street, Mountainside, New Jersey 07092, Phone : 908 789 8900,
Fax : 908 789 8922

Metals

- 2b -

CRDL STANDARD FOR AA & ICP

Client: JACOBS Engineering Group, Inc.

SDG No.: P3573

Contract: JACO05

Lab Code: CHEM

Case No.: P3573

SAS No.: P3573

Initial Calibration Source: _____

Continuing Calibration Source: _____

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Acceptance Window (%R)	M	Analysis Date	Analysis Time	Run Number
CRI01	Silver	12.3	10.0	123	40 - 160	P	08/15/2024	12:38	LB132039
CRI01	Silver	11.6	10.0	116	40 - 160	P	08/23/2024	10:05	LB132142
CRI01	Silver	10.8	10.0	108	40 - 160	P	08/24/2024	11:04	LB132156
CRI01	Silver	12.5	10.0	125	40 - 160	P	08/25/2024	19:26	LB132157