

Metals

- 4 -

INTERFERENCE CHECK SAMPLE

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3573
Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3573 **SAS No.:** P3573
ICS Source: EPA **Instrument ID:** P4

Sample ID	Analyte	Result ug/L	True Value ug/L	% Recovery	Low Limit (ug/L)	High Limit (ug/L)	Analysis Date	Analysis Time	Run Number
ICSA01	Silver	2.10			-10	10	08/15/2024	12:42	LB132039
ICSAB01	Silver	221	201	110	170	232	08/15/2024	12:56	LB132039
ICSA01	Silver	2.05			-10	10	08/23/2024	10:09	LB132142
ICSAB01	Silver	229	201	114	170	232	08/23/2024	10:14	LB132142
ICSA01	Silver	2.05			-10	10	08/24/2024	11:08	LB132156
ICSAB01	Silver	230	201	114	170	232	08/24/2024	11:33	LB132156
ICSA01	Silver	3.22			-10	10	08/25/2024	19:36	LB132157
ICSAB01	Silver	231	201	115	170	232	08/25/2024	19:40	LB132157