

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

**Client:** JACOBS Engineering Group, Inc.      **Level:** LOW      **SDG No.:** P3573  
**Contract:** JACO05      **Lab Code:** CHEM      **Case No.:** P3573      **SAS No.:** P3573  
**Matrix:** Water      **Sample ID:** P3596-01      **Client ID:** 918-J-WS-081324DUP  
**Percent Solids for Sample:** NA      **Duplicate ID** P3596-01DUP      **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | C | Duplicate<br>Result | C | RPD | Qual | M |
|---------|-------|---------------------|------------------|---|---------------------|---|-----|------|---|
| Silver  | ug/L  | 20                  | 5.00             | U | 5.00                | U |     |      | P |

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

**Client:** JACOBS Engineering Group, Inc. **Level:** LOW **SDG No.:** P3573  
**Contract:** JACO05 **Lab Code:** CHEM **Case No.:** P3573 **SAS No.:** P3573  
**Matrix:** Water **Sample ID:** P3596-01MS **Client ID:** 918-J-WS-081324MSD  
**Percent Solids for Sample:** NA **Duplicate ID** P3596-01MSD **Percent Solids for Spike Sample:** NA

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | C | Duplicate<br>Result | C | RPD | Qual | M |
|---------|-------|---------------------|------------------|---|---------------------|---|-----|------|---|
| Silver  | ug/L  | 20                  | 34.7             |   | 36.7                |   | 6   |      | P |

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

**Client:** JACOBS Engineering Group, Inc.      **Level:** LOW      **SDG No.:** P3573  
**Contract:** JACO05      **Lab Code:** CHEM      **Case No.:** P3573      **SAS No.:** P3573  
**Matrix:** Solid      **Sample ID:** P3600-06      **Client ID:** VCPS-B3-081324-10-18DUP  
**Percent Solids for Sample:** 80.3      **Duplicate ID** P3600-06DUP      **Percent Solids for Spike Sample:** 80.3

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | C | Duplicate<br>Result | C | RPD | Qual | M |
|---------|-------|---------------------|------------------|---|---------------------|---|-----|------|---|
| Silver  | mg/Kg | 20                  | 0.45             | J | 0.58                |   | 24  |      | P |

**Metals**

- 6 -

**DUPLICATE SAMPLE SUMMARY**

|                                   |                                       |                     |                   |                                         |                                |
|-----------------------------------|---------------------------------------|---------------------|-------------------|-----------------------------------------|--------------------------------|
| <b>Client:</b>                    | <u>JACOBS Engineering Group, Inc.</u> | <b>Level:</b>       | <u>LOW</u>        | <b>SDG No.:</b>                         | <u>P3573</u>                   |
| <b>Contract:</b>                  | <u>JACO05</u>                         | <b>Lab Code:</b>    | <u>CHEM</u>       | <b>Case No.:</b>                        | <u>P3573</u>                   |
| <b>Matrix:</b>                    | <u>Solid</u>                          | <b>Sample ID:</b>   | <u>P3600-06MS</u> | <b>Client ID:</b>                       | <u>VCPS-B3-081324-10-18MSD</u> |
| <b>Percent Solids for Sample:</b> | 80.3                                  | <b>Duplicate ID</b> | P3600-06MSD       | <b>Percent Solids for Spike Sample:</b> | 80.3                           |

| Analyte | Units | Acceptance<br>Limit | Sample<br>Result | C | Duplicate<br>Result | C | RPD | Qual | M |
|---------|-------|---------------------|------------------|---|---------------------|---|-----|------|---|
| Silver  | mg/Kg | 20                  | 4.84             |   | 4.75                |   | 2   |      | P |