

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081524\
 Data File : BP021515.D
 Acq On : 15 Aug 2024 15:37
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 15 16:12:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	20.000	20.000	0.0	105	0.00
2	1,4-Dioxane	40.000	39.721	0.7	104	0.00
3	Pyridine	40.000	39.855	0.4	104	0.00
4	n-Nitrosodimethylamine	40.000	40.827	-2.1	108	0.00
5 S	2-Fluorophenol	80.000	81.986	-2.5	108	0.00
6	Aniline	40.000	39.293	1.8	103	0.00
7 S	Phenol-d6	80.000	81.542	-1.9	107	0.00
8	2-Chlorophenol	40.000	41.119	-2.8	108	0.00
9	Benzaldehyde	40.000	38.463	3.8	106	0.00
10 C	Phenol	40.000	40.255	-0.6	105	0.00
11	bis(2-Chloroethyl)ether	40.000	39.109	2.2	104	0.00
12	1,3-Dichlorobenzene	40.000	39.534	1.2	106	0.00
13 C	1,4-Dichlorobenzene	40.000	39.379	1.6	105	0.00
14	1,2-Dichlorobenzene	40.000	39.319	1.7	106	0.00
15	Benzyl Alcohol	40.000	40.558	-1.4	106	0.00
16	2,2'-oxybis(1-Chloropropane	40.000	38.883	2.8	105	0.00
17	2-Methylphenol	40.000	40.526	-1.3	106	0.00
18	Hexachloroethane	40.000	40.569	-1.4	108	0.00
19 P	n-Nitroso-di-n-propylamine	40.000	38.203	4.5	102	0.00
20	3+4-Methylphenols	40.000	41.196	-3.0	107	0.00
21 I	Naphthalene-d8	20.000	20.000	0.0	107	0.00
22	Acetophenone	40.000	39.083	2.3	105	0.00
23 S	Nitrobenzene-d5	80.000	90.589	-13.2	114	0.00
24	Nitrobenzene	40.000	44.076	-10.2	112	0.00
25	Isophorone	40.000	38.545	3.6	103	0.00
26 C	2-Nitrophenol	40.000	45.564	-13.9	132	0.00
27	2,4-Dimethylphenol	40.000	40.514	-1.3	108	0.00
28	bis(2-Chloroethoxy)methane	40.000	39.035	2.4	106	0.00
29 C	2,4-Dichlorophenol	40.000	43.498	-8.7	114	0.00
30	1,2,4-Trichlorobenzene	40.000	39.703	0.7	106	0.00
31	Naphthalene	40.000	39.804	0.5	107	0.00
32	Benzoic acid	40.000	49.992	-25.0	155	0.00
33	4-Chloroaniline	40.000	40.717	-1.8	108	0.00
34 C	Hexachlorobutadiene	40.000	39.188	2.0	105	0.00
35	Caprolactam	40.000	43.743	-9.4	114	0.00
36 C	4-Chloro-3-methylphenol	40.000	43.006	-7.5	111	0.00
37	2-Methylnaphthalene	40.000	39.324	1.7	106	-0.01
38	1-Methylnaphthalene	40.000	39.539	1.2	106	0.00
39 I	Acenaphthene-d10	20.000	20.000	0.0	109	-0.01
40	1,2,4,5-Tetrachlorobenzene	40.000	39.477	1.3	106	0.00
41 P	Hexachlorocyclopentadiene	40.000	39.224	1.9	103	0.00
42 S	2,4,6-Tribromophenol	80.000	81.327	-1.7	104	0.00
43 C	2,4,6-Trichlorophenol	40.000	43.465	-8.7	112	0.00
44	2,4,5-Trichlorophenol	40.000	45.100	-12.8	116	0.00
45 S	2-Fluorobiphenyl	80.000	79.374	0.8	107	0.00
46	1,1'-Biphenyl	40.000	39.694	0.8	107	-0.01
47	2-Chloronaphthalene	40.000	39.608	1.0	106	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081524\
 Data File : BP021515.D
 Acq On : 15 Aug 2024 15:37
 Operator : RC/JU
 Sample : SSTDCCC040
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_P
 LabSampleId :
 SSTDCCC040

Quant Time: Aug 15 16:12:42 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Aug 14 00:00:30 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
48	2-Nitroaniline	40.000	44.587	-11.5	130	0.00
49	Acenaphthylene	40.000	40.538	-1.3	109	0.00
50	Dimethylphthalate	40.000	40.806	-2.0	110	0.00
51	2,6-Dinitrotoluene	40.000	44.191	-10.5	126	-0.02
52 C	Acenaphthene	40.000	40.528	-1.3	109	0.00
53	3-Nitroaniline	40.000	45.344	-13.4	128	-0.02
54 P	2,4-Dinitrophenol	40.000	48.146	-20.4	147	0.00
55	Dibenzofuran	40.000	40.462	-1.2	109	0.00
56 P	4-Nitrophenol	40.000	47.162	-17.9	135	0.00
57	2,4-Dinitrotoluene	40.000	45.473	-13.7	132	0.00
58	Fluorene	40.000	39.875	0.3	108	0.00
59	2,3,4,6-Tetrachlorophenol	40.000	44.289	-10.7	113	-0.01
60	Diethylphthalate	40.000	41.012	-2.5	110	-0.01
61	4-Chlorophenyl-phenylether	40.000	39.744	0.6	108	-0.01
62	4-Nitroaniline	40.000	46.527	-16.3	127	-0.02
63	Azobenzene	40.000	39.977	0.1	107	-0.02
64 I	Phenanthrene-d10	20.000	20.000	0.0	111	0.00
65	4,6-Dinitro-2-methylphenol	40.000	45.318	-13.3	139	0.00
66 c	n-Nitrosodiphenylamine	40.000	39.253	1.9	108	-0.02
67	4-Bromophenyl-phenylether	40.000	36.394	9.0	103	-0.02
68	Hexachlorobenzene	40.000	38.761	3.1	107	-0.01
69	Atrazine	40.000	35.940	10.2	96	-0.02
70 C	Pentachlorophenol	40.000	42.785	-7.0	119	0.00
71	Phenanthrene	40.000	40.008	-0.0	111	0.00
72	Anthracene	40.000	39.833	0.4	110	-0.01
73	Carbazole	40.000	41.090	-2.7	112	-0.02
74	Di-n-butylphthalate	40.000	40.677	-1.7	110	0.00
75 C	Fluoranthene	40.000	41.674	-4.2	114	0.00
76 I	Chrysene-d12	20.000	20.000	0.0	117	0.00
77	Benzydine	40.000	39.007	2.5	112	-0.01
78	Pyrene	40.000	38.402	4.0	114	0.00
79 S	Terphenyl-d14	80.000	74.981	6.3	112	-0.01
80	Butylbenzylphthalate	40.000	38.009	5.0	115	-0.01
81	Benzo(a)anthracene	40.000	39.464	1.3	114	0.00
82	3,3'-Dichlorobenzidine	40.000	42.721	-6.8	117	-0.02
83	Chrysene	40.000	39.453	1.4	114	0.00
84	Bis(2-ethylhexyl)phthalate	40.000	41.242	-3.1	115	-0.01
85 c	Di-n-octyl phthalate	40.000	38.250	4.4	114	-0.02
86 I	Perylene-d12	20.000	20.000	0.0	118	0.00
87	Indeno(1,2,3-cd)pyrene	40.000	40.793	-2.0	118	-0.02
88	Benzo(b)fluoranthene	40.000	40.252	-0.6	119	0.00
89	Benzo(k)fluoranthene	40.000	39.736	0.7	116	-0.01
90 C	Benzo(a)pyrene	40.000	40.447	-1.1	118	-0.01
91	Dibenzo(a,h)anthracene	40.000	40.485	-1.2	117	0.00
92	Benzo(g,h,i)perylene	40.000	40.874	-2.2	119	0.02

Data Path : Z:\svoasrv\HPCHEM1\BNA_P\Data\BP081524\
Data File : BP021515.D
Acq On : 15 Aug 2024 15:37
Operator : RC/JU
Sample : SSTDCCC040
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_P
LabSampleId :
SSTDCCC040

Quant Time: Aug 15 16:12:42 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_P\Methods\8270E-BP081324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Aug 14 00:00:30 2024
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)
----------	--------	-------	------	-------	----------

(#) = Out of Range SPCC's out = 0 CCC's out = 0