

Report of Analysis

Client:	JACOBS Engineering Group, Inc.	Date Collected:	
Project:	Former Schlumberger Site Princeton NJ	Date Received:	
Client Sample ID:	PB162688BS	SDG No.:	P3573
Lab Sample ID:	PB162688BS	Matrix:	SOIL
Analytical Method:	8270E	% Solid:	100
Sample Wt/Vol:	30.01 Units: g	Final Vol:	1000 uL
Soil Aliquot Vol:	uL	Test:	SVOCMS Group2
Extraction Type :	Decanted : N	Level :	LOW
Injection Volume :	GPC Factor : 1.0	GPC Cleanup :	N PH :
Prep Method :	SW3541		

File ID/Qc Batch:	Dilution:	Prep Date	Date Analyzed	Prep Batch ID
BF139075.D	1	08/13/24 08:30	08/19/24 15:47	PB162688

CAS Number	Parameter	Conc.	Qualifier	MDL	LOQ / CRQL	Units(Dry Weight)
TARGETS						
56-55-3	Benzo(a)anthracene	1600		80.7	170	ug/Kg
50-32-8	Benzo(a)pyrene	1900		93.0	170	ug/Kg
53-70-3	Dibenzo(a,h)anthracene	1500		81.2	170	ug/Kg
SURROGATES						
4165-60-0	Nitrobenzene-d5	87.6		30 (18) - 130 (107)	88%	SPK: 100
321-60-8	2-Fluorobiphenyl	87.7		30 (20) - 130 (109)	88%	SPK: 100
1718-51-0	Terphenyl-d14	106		30 (10) - 130 (105)	106%	SPK: 100
INTERNAL STANDARDS						
3855-82-1	1,4-Dichlorobenzene-d4	53200	6.834			
1146-65-2	Naphthalene-d8	219000	8.116			
15067-26-2	Acenaphthene-d10	122000	9.869			
1517-22-2	Phenanthrene-d10	202000	11.357			
1719-03-5	Chrysene-d12	98700	14.004			
1520-96-3	Perylene-d12	90200	15.468			

U = Not Detected

LOQ = Limit of Quantitation

MDL = Method Detection Limit

LOD = Limit of Detection

E = Value Exceeds Calibration Range

Q = indicates LCS control criteria did not meet requirements

M = MS/MSD acceptance criteria did not meet requirements

J = Estimated Value

B = Analyte Found in Associated Method Blank

N = Presumptive Evidence of a Compound

* = Values outside of QC limits

D = Dilution

() = Laboratory InHouse Limit

A = Aldol-Condensation Reaction Products