

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF073024

Review By	yogesh	Review On	7/31/2024 6:39:45 AM
Supervise By	mohammad	Supervise On	7/31/2024 8:08:34 AM
SubDirectory	BF073024	HP Acquire Method	BNA_F
		HP Processing Method	bf073024
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557		
CCC	SP6553		
Internal Standard/PEM	S12036 10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF138679.D	30 Jul 2024 12:24	RC/JU	Ok
2	SSTDICC2.5	BF138680.D	30 Jul 2024 12:54	RC/JU	Ok
3	SSTDICC005	BF138681.D	30 Jul 2024 13:25	RC/JU	Ok
4	SSTDICC010	BF138682.D	30 Jul 2024 13:56	RC/JU	Ok,M
5	SSTDICC020	BF138683.D	30 Jul 2024 14:25	RC/JU	Ok
6	SSTDICCC040	BF138684.D	30 Jul 2024 14:56	RC/JU	Ok
7	SSTDICC050	BF138685.D	30 Jul 2024 15:27	RC/JU	Ok
8	SSTDICC060	BF138686.D	30 Jul 2024 15:58	RC/JU	Ok
9	SSTDICC080	BF138687.D	30 Jul 2024 16:29	RC/JU	Ok
10	SSTDICV040	BF138688.D	30 Jul 2024 17:55	RC/JU	Ok
11	PB162188BL	BF138689.D	30 Jul 2024 18:26	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF081524

Review By	Jagrut	Review On	8/16/2024 10:58:33 AM
Supervise By	mohammad	Supervise On	8/17/2024 2:51:27 AM
SubDirectory	BF081524	HP Acquire Method	BNA_F
		HP Processing Method	bf073024
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557		
CCC	SP6553		
Internal Standard/PEM	S12041,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF139011.D	15 Aug 2024 09:17	RC/JU	Ok
2	SSTDCCC040	BF139012.D	15 Aug 2024 09:47	RC/JU	Ok,M
3	PB162727BL	BF139013.D	15 Aug 2024 10:16	RC/JU	Ok
4	PB162590TB	BF139014.D	15 Aug 2024 10:46	RC/JU	Ok
5	PB162727BS	BF139015.D	15 Aug 2024 11:15	RC/JU	Ok,M
6	PB162682TB	BF139016.D	15 Aug 2024 11:45	RC/JU	Ok
7	PB162588TB	BF139017.D	15 Aug 2024 12:15	RC/JU	Ok
8	PB162683TB	BF139018.D	15 Aug 2024 12:45	RC/JU	Ok
9	PB162621TB	BF139019.D	15 Aug 2024 13:16	RC/JU	Ok
10	P3574-01	BF139020.D	15 Aug 2024 13:51	RC/JU	Ok
11	P3574-01MS	BF139021.D	15 Aug 2024 14:22	RC/JU	Ok,M
12	P3574-01MSD	BF139022.D	15 Aug 2024 14:52	RC/JU	Ok,M
13	P3574-02	BF139023.D	15 Aug 2024 15:23	RC/JU	Ok
14	DFTPP	BF139024.D	15 Aug 2024 16:00	RC/JU	Ok
15	SSTDCCC040	BF139025.D	15 Aug 2024 16:31	RC/JU	Ok,M
16	PB162688BL	BF139026.D	15 Aug 2024 17:03	RC/JU	Ok
17	P3580-03	BF139027.D	15 Aug 2024 17:37	RC/JU	Ok
18	P3580-04	BF139028.D	15 Aug 2024 18:08	RC/JU	Ok
19	P3580-05	BF139029.D	15 Aug 2024 18:40	RC/JU	Ok
20	P3580-06	BF139030.D	15 Aug 2024 19:11	RC/JU	Ok
21	P3580-07	BF139031.D	15 Aug 2024 19:41	RC/JU	Ok

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF081524

Review By	Jagrut	Review On	8/16/2024 10:58:33 AM
Supervise By	mohammad	Supervise On	8/17/2024 2:51:27 AM
SubDirectory	BF081524	HP Acquire Method	BNA_F
		HP Processing Method	bf073024
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557		
CCC	SP6553		
Internal Standard/PEM	S12041,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	P3580-08	BF139032.D	15 Aug 2024 20:12	RC/JU	Ok
23	P3600-04	BF139033.D	15 Aug 2024 20:44	RC/JU	Ok,M
24	P3475-01	BF139034.D	15 Aug 2024 21:14	RC/JU	Ok
25	P3573-07	BF139035.D	15 Aug 2024 21:45	RC/JU	Ok
26	P3573-09	BF139036.D	15 Aug 2024 22:16	RC/JU	Ok
27	P3573-05	BF139037.D	15 Aug 2024 22:47	RC/JU	Ok
28	P3573-05MS	BF139038.D	15 Aug 2024 23:17	RC/JU	Ok,M
29	P3573-05MSD	BF139039.D	15 Aug 2024 23:47	RC/JU	Ok,M
30	P3485-01	BF139040.D	16 Aug 2024 00:18	RC/JU	Ok
31	P3485-01MS	BF139041.D	16 Aug 2024 00:48	RC/JU	Ok,M
32	P3485-01MSD	BF139042.D	16 Aug 2024 01:18	RC/JU	Ok,M
33	P3483-01	BF139043.D	16 Aug 2024 01:49	RC/JU	Ok,M
34	P3597-01	BF139044.D	16 Aug 2024 02:19	RC/JU	Ok,M
35	P3548-06RE	BF139045.D	16 Aug 2024 02:49	RC/JU	Confirms
36	P3548-12RE	BF139046.D	16 Aug 2024 03:20	RC/JU	Confirms
37	P3526-01	BF139047.D	16 Aug 2024 03:50	RC/JU	Ok,M

M : Manual Integration

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF081924

Review By	Jagrut	Review On	8/20/2024 8:53:23 AM
Supervise By	mohammad	Supervise On	8/21/2024 1:24:27 AM
SubDirectory	BF081924	HP Acquire Method	BNA_F
		HP Processing Method	bf073024
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557		
CCC	SP6553		
Internal Standard/PEM	S12041,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF139064.D	19 Aug 2024 08:41	RC/JU	Ok
2	SSTDCCC040	BF139065.D	19 Aug 2024 09:41	RC/JU	Ok,M
3	PB162549BL	BF139066.D	19 Aug 2024 10:10	RC/JU	Ok
4	PB162553BL	BF139067.D	19 Aug 2024 10:40	RC/JU	Ok
5	P3390-03	BF139068.D	19 Aug 2024 11:16	RC/JU	Ok,M
6	P3390-03	BF139069.D	19 Aug 2024 11:46	RC/JU	Ok,M
7	P3390-09	BF139070.D	19 Aug 2024 12:32	RC/JU	Ok,M
8	P3390-09	BF139071.D	19 Aug 2024 13:02	RC/JU	Ok,M
9	PB162612BL	BF139072.D	19 Aug 2024 14:17	RC/JU	Ok
10	PB162612BS	BF139073.D	19 Aug 2024 14:46	RC/JU	Ok,M
11	PB162549BS	BF139074.D	19 Aug 2024 15:17	RC/JU	Ok,M
12	PB162688BS	BF139075.D	19 Aug 2024 15:47	RC/JU	Ok,M
13	SSTDCCC040	BF139076.D	19 Aug 2024 16:18	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF082024

Review By	Jagrut	Review On	8/21/2024 10:35:25 AM
Supervise By	mohammad	Supervise On	8/22/2024 7:52:09 AM
SubDirectory	BF082024	HP Acquire Method	BNA_F
		HP Processing Method	bf082024
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588		
CCC	SP6591		
Internal Standard/PEM	S12041,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF139077.D	20 Aug 2024 16:39	RC/JU	Ok
2	SSTDICC2.5	BF139078.D	20 Aug 2024 17:10	RC/JU	Ok
3	SSTDICC005	BF139079.D	20 Aug 2024 17:40	RC/JU	Ok,M
4	SSTDICC010	BF139080.D	20 Aug 2024 18:10	RC/JU	Ok
5	SSTDICC020	BF139081.D	20 Aug 2024 18:40	RC/JU	Ok
6	SSTDICCC040	BF139082.D	20 Aug 2024 19:10	RC/JU	Ok,M
7	SSTDICC050	BF139083.D	20 Aug 2024 19:41	RC/JU	Ok
8	SSTDICC060	BF139084.D	20 Aug 2024 20:11	RC/JU	Ok,M
9	SSTDICC080	BF139085.D	20 Aug 2024 20:41	RC/JU	Ok,M
10	SSTDICV040	BF139086.D	20 Aug 2024 21:41	RC/JU	Ok
11	PB162687BL	BF139087.D	20 Aug 2024 22:10	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF082224

Review By	yogesh	Review On	8/23/2024 6:48:24 AM
Supervise By	mohammad	Supervise On	8/23/2024 6:56:56 AM
SubDirectory	BF082224	HP Acquire Method	BNA_F
		HP Processing Method	bf082024
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588		
CCC	SP6591		
Internal Standard/PEM	S12041,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BF139112.D	22 Aug 2024 09:55	RC/JU	Ok
2	SSTDCCC040	BF139113.D	22 Aug 2024 10:25	RC/JU	Ok
3	PB162782BL	BF139114.D	22 Aug 2024 10:54	RC/JU	Not Ok
4	PB162687BS	BF139115.D	22 Aug 2024 11:23	RC/JU	Ok,M
5	PB162687BSD	BF139116.D	22 Aug 2024 11:53	RC/JU	Ok,M
6	P3641-04	BF139117.D	22 Aug 2024 12:22	RC/JU	Ok,M
7	P3663-03	BF139118.D	22 Aug 2024 12:52	RC/JU	Ok,M
8	P3641-02	BF139119.D	22 Aug 2024 13:22	RC/JU	Ok,M
9	P3640-01	BF139120.D	22 Aug 2024 13:51	RC/JU	Not Ok
10	P3641-06	BF139121.D	22 Aug 2024 14:20	RC/JU	Ok,M
11	P3641-08	BF139122.D	22 Aug 2024 14:50	RC/JU	Ok,M
12	P3641-09	BF139123.D	22 Aug 2024 15:21	RC/JU	Ok,M
13	P3641-10	BF139124.D	22 Aug 2024 15:51	RC/JU	Ok,M
14	P3641-01	BF139125.D	22 Aug 2024 16:21	RC/JU	Ok,M
15	P3663-09	BF139126.D	22 Aug 2024 16:51	RC/JU	Ok,M
16	P3663-11	BF139127.D	22 Aug 2024 17:21	RC/JU	Ok,M
17	P3663-12	BF139128.D	22 Aug 2024 17:51	RC/JU	Ok,M
18	P3641-03	BF139129.D	22 Aug 2024 18:21	RC/JU	Dilution
19	P3660-03DL	BF139130.D	22 Aug 2024 18:50	RC/JU	Ok,M
20	P3660-04DL	BF139131.D	22 Aug 2024 19:20	RC/JU	Ok,M
21	P3663-10DL	BF139132.D	22 Aug 2024 19:50	RC/JU	Ok,M

Instrument ID: **BNA_F**

Daily Analysis Runlog For Sequence/QC Batch ID # BF082224

Review By	yogesh	Review On	8/23/2024 6:48:24 AM
Supervise By	mohammad	Supervise On	8/23/2024 6:56:56 AM
SubDirectory	BF082224	HP Acquire Method	BNA_F
		HP Processing Method	bf082024
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588		
CCC	SP6591		
Internal Standard/PEM	S12041,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	P3641-05	BF139133.D	22 Aug 2024 20:20	RC/JU	Ok,M
23	P3641-07	BF139134.D	22 Aug 2024 20:50	RC/JU	Ok,M
24	P3671-12	BF139135.D	22 Aug 2024 21:20	RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP081324

Review By	yogesh	Review On	8/14/2024 5:49:15 AM
Supervise By	mohammad	Supervise On	8/14/2024 6:23:14 AM
SubDirectory	BP081324	HP Acquire Method	BNA_P
		HP Processing Method	BP081324
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588		
CCC	SP6591		
Internal Standard/PEM	S12041,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP021480.D	13 Aug 2024 11:49	RC/JU	Ok
2	SSTDICC2.5	BP021481.D	13 Aug 2024 12:32	RC/JU	Ok
3	SSTDICC005	BP021482.D	13 Aug 2024 13:15	RC/JU	Ok,M
4	SSTDICC010	BP021483.D	13 Aug 2024 13:59	RC/JU	Ok,M
5	SSTDICC020	BP021484.D	13 Aug 2024 14:42	RC/JU	Ok,M
6	SSTDICCC040	BP021485.D	13 Aug 2024 15:26	RC/JU	Ok,M
7	SSTDICC050	BP021486.D	13 Aug 2024 16:09	RC/JU	Ok,M
8	SSTDICC060	BP021487.D	13 Aug 2024 16:53	RC/JU	Ok,M
9	SSTDICC080	BP021488.D	13 Aug 2024 17:36	RC/JU	Ok,M
10	SSTDICV040	BP021489.D	13 Aug 2024 19:03	RC/JU	Ok,M
11	PB162612TB	BP021490.D	13 Aug 2024 19:46	RC/JU	Ok

M : Manual Integration

Instrument ID: **BNA_P**

Daily Analysis Runlog For Sequence/QC Batch ID # BP081524

Review By	Jagrut	Review On	8/16/2024 9:58:34 AM
Supervise By	mohammad	Supervise On	8/17/2024 2:51:46 AM
SubDirectory	BP081524	HP Acquire Method	BNA_P
		HP Processing Method	BP081324
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588		
CCC	SP6591		
Internal Standard/PEM	S12041,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

Sr#	SampleId	Data File Name	Date-Time	Operator	Status
1	DFTPP	BP021505.D	15 Aug 2024 08:28	RC/JU	Ok
2	SSTDCCC040	BP021506.D	15 Aug 2024 09:10	RC/JU	Ok,M
3	PB162549BL	BP021507.D	15 Aug 2024 09:52	RC/JU	Ok
4	PB162553BL	BP021508.D	15 Aug 2024 10:34	RC/JU	Ok
5	P3390-03	BP021509.D	15 Aug 2024 11:18	RC/JU	Ok
6	P3390-03	BP021510.D	15 Aug 2024 12:00	RC/JU	Ok,M
7	P3390-09	BP021511.D	15 Aug 2024 12:43	RC/JU	Ok,M
8	P3390-09	BP021512.D	15 Aug 2024 13:26	RC/JU	Ok
9	SSTDCCC040	BP021513.D	15 Aug 2024 14:09	RC/JU	Ok,M
10	DFTPP	BP021514.D	15 Aug 2024 14:54	RC/JU	Ok
11	SSTDCCC040	BP021515.D	15 Aug 2024 15:37	RC/JU	Ok,M
12	PB162688BL	BP021516.D	15 Aug 2024 16:19	RC/JU	Not Ok
13	P3573-12	BP021517.D	15 Aug 2024 17:09	RC/JU	Ok
14	P3582-07	BP021518.D	15 Aug 2024 17:53	RC/JU	Ok
15	P3582-02	BP021519.D	15 Aug 2024 18:37	RC/JU	Ok
16	P3582-04	BP021520.D	15 Aug 2024 19:20	RC/JU	Ok
17	P3582-03	BP021521.D	15 Aug 2024 20:03	RC/JU	Ok
18	P3573-06	BP021522.D	15 Aug 2024 20:47	RC/JU	Ok
19	P3573-08	BP021523.D	15 Aug 2024 21:30	RC/JU	Ok
20	P3573-10	BP021524.D	15 Aug 2024 22:14	RC/JU	Ok
21	P3607-01	BP021525.D	15 Aug 2024 22:57	RC/JU	Ok,M

Instrument ID: **BNA_P**

Daily Analysis Runlog For Sequence/QC Batch ID # BP081524

Review By	Jagrut	Review On	8/16/2024 9:58:34 AM
Supervise By	mohammad	Supervise On	8/17/2024 2:51:46 AM
SubDirectory	BP081524	HP Acquire Method	BNA_P
		HP Processing Method	BP081324
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588		
CCC	SP6591		
Internal Standard/PEM	S12041,10ul/1000ul sample		
ICV/I.BLK	SP6559		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

22	P3607-01MS	BP021526.D	15 Aug 2024 23:41	RC/JU	Ok,M
23	P3607-01MSD	BP021527.D	16 Aug 2024 00:24	RC/JU	Ok,M
24	P3579-02	BP021528.D	16 Aug 2024 01:07	RC/JU	Ok,M
25	P3598-01	BP021529.D	16 Aug 2024 01:51	RC/JU	Ok
26	P3598-03	BP021530.D	16 Aug 2024 02:34	RC/JU	Ok

M : Manual Integration