

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF073024

Review By	yogesh	Review On	7/31/2024 6:39:45 AM
Supervise By	mohammad	Supervise On	7/31/2024 8:08:34 AM
SubDirectory	BF073024	HP Acquire Method	BNA_F
		HP Processing Method	bf073024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12036 10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF138679.D	30 Jul 2024 12:24		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF138680.D	30 Jul 2024 12:54		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF138681.D	30 Jul 2024 13:25	Compound#32,41,54,56,65,70 removed from 5 ppm	RC/JU	Ok
4	SSTDICC010	SSTDICC010	BF138682.D	30 Jul 2024 13:56		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BF138683.D	30 Jul 2024 14:25	Comopund#41 Kept on LR	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF138684.D	30 Jul 2024 14:56	The calibration is good for 8270 DOD & 625.1	RC/JU	Ok
7	SSTDICC050	SSTDICC050	BF138685.D	30 Jul 2024 15:27		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF138686.D	30 Jul 2024 15:58	Compound#9 removed from 60 ppm	RC/JU	Ok
9	SSTDICC080	SSTDICC080	BF138687.D	30 Jul 2024 16:29	Compound#9,77 removed from 80 ppm	RC/JU	Ok
10	SSTDICV040	ICVBF073024	BF138688.D	30 Jul 2024 17:55		RC/JU	Ok
11	PB162188BL	PB162188BL	BF138689.D	30 Jul 2024 18:26		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF081524

Review By	Jagrut	Review On	8/16/2024 10:58:33 AM
Supervise By	mohammad	Supervise On	8/17/2024 2:51:27 AM
SubDirectory	BF081524	HP Acquire Method	BNA_F
		HP Processing Method	bf073024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557
CCC	SP6553
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF139011.D	15 Aug 2024 09:17		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF139012.D	15 Aug 2024 09:47		RC/JU	Ok,M
3	PB162727BL	PB162727BL	BF139013.D	15 Aug 2024 10:16		RC/JU	Ok
4	PB162590TB	PB162590TB	BF139014.D	15 Aug 2024 10:46		RC/JU	Ok
5	PB162727BS	PB162727BS	BF139015.D	15 Aug 2024 11:15		RC/JU	Ok,M
6	PB162682TB	PB162682TB	BF139016.D	15 Aug 2024 11:45		RC/JU	Ok
7	PB162588TB	PB162588TB	BF139017.D	15 Aug 2024 12:15		RC/JU	Ok
8	PB162683TB	PB162683TB	BF139018.D	15 Aug 2024 12:45		RC/JU	Ok
9	PB162621TB	PB162621TB	BF139019.D	15 Aug 2024 13:16		RC/JU	Ok
10	P3574-01	WC-S-06-202408	BF139020.D	15 Aug 2024 13:51		RC/JU	Ok
11	P3574-01MS	WC-S-06-202408MS	BF139021.D	15 Aug 2024 14:22		RC/JU	Ok,M
12	P3574-01MSD	WC-S-06-202408MSD	BF139022.D	15 Aug 2024 14:52		RC/JU	Ok,M
13	P3574-02	WC-W-06-202408	BF139023.D	15 Aug 2024 15:23		RC/JU	Ok
14	DFTPP	DFTPP	BF139024.D	15 Aug 2024 16:00		RC/JU	Ok
15	SSTDCCC040	SSTDCCC040	BF139025.D	15 Aug 2024 16:31		RC/JU	Ok,M
16	PB162688BL	PB162688BL	BF139026.D	15 Aug 2024 17:03		RC/JU	Ok
17	P3580-03	925-K1-SD-0-0.5-0812	BF139027.D	15 Aug 2024 17:37		RC/JU	Ok
18	P3580-04	925-K1-SD-0.5-1-0812	BF139028.D	15 Aug 2024 18:08		RC/JU	Ok

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF081524

Review By	Jagrut	Review On	8/16/2024 10:58:33 AM		
Supervise By	mohammad	Supervise On	8/17/2024 2:51:27 AM		
SubDirectory	BF081524	HP Acquire Method	BNA_F	HP Processing Method	bf073024
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557				
CCC	SP6553				
Internal Standard/PEM	S12041,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P3580-05	923-K1-SD-0-0.5-08122	BF139029.D	15 Aug 2024 18:40		RC/JU	Ok
20	P3580-06	923-K1-SD-0.5-1-08122	BF139030.D	15 Aug 2024 19:11		RC/JU	Ok
21	P3580-07	922-K1-SD-0-0.5-08122	BF139031.D	15 Aug 2024 19:41		RC/JU	Ok
22	P3580-08	922-K1-SD-0.5-1-08122	BF139032.D	15 Aug 2024 20:12		RC/JU	Ok
23	P3600-04	VCPS-B3-081324-0-10	BF139033.D	15 Aug 2024 20:44		RC/JU	Ok,M
24	P3475-01	S-866-N-SO-0-0.5-0809	BF139034.D	15 Aug 2024 21:14		RC/JU	Ok
25	P3573-07	S-856-J-SO-1.5-2-0809	BF139035.D	15 Aug 2024 21:45		RC/JU	Ok
26	P3573-09	S-856-J-SO-3-3.5-0809	BF139036.D	15 Aug 2024 22:16		RC/JU	Ok
27	P3573-05	S-856-J-SO-1-1.5-0809	BF139037.D	15 Aug 2024 22:47		RC/JU	Ok
28	P3573-05MS	S-856-J-SO-1-1.5-0809	BF139038.D	15 Aug 2024 23:17		RC/JU	Ok,M
29	P3573-05MSD	S-856-J-SO-1-1.5-0809	BF139039.D	15 Aug 2024 23:47		RC/JU	Ok,M
30	P3485-01	C-SOIL	BF139040.D	16 Aug 2024 00:18		RC/JU	Ok
31	P3485-01MS	C-SOILMS	BF139041.D	16 Aug 2024 00:48		RC/JU	Ok,M
32	P3485-01MSD	C-SOILMSD	BF139042.D	16 Aug 2024 01:18		RC/JU	Ok,M
33	P3483-01	A-TOPSOIL	BF139043.D	16 Aug 2024 01:49	Parameter list is not updated	RC/JU	Ok,M
34	P3597-01	SP-1-COMP	BF139044.D	16 Aug 2024 02:19		RC/JU	Ok,M
35	P3548-06RE	343RE	BF139045.D	16 Aug 2024 02:49	Fax is already given	RC/JU	Confirms
36	P3548-12RE	IN-SITU-SOILRE	BF139046.D	16 Aug 2024 03:20	Fax is already given	RC/JU	Confirms
37	P3526-01	SP-1-COMP	BF139047.D	16 Aug 2024 03:50		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF081524

Review By	Jagrut	Review On	8/16/2024 10:58:33 AM		
Supervise By	mohammad	Supervise On	8/17/2024 2:51:27 AM		
SubDirectory	BF081524	HP Acquire Method	BNA_F	HP Processing Method	bf073024
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557				
CCC	SP6553				
Internal Standard/PEM	S12041,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF081924

Review By	Jagrut	Review On	8/20/2024 8:53:23 AM		
Supervise By	mohammad	Supervise On	8/21/2024 1:24:27 AM		
SubDirectory	BF081924	HP Acquire Method	BNA_F	HP Processing Method	bf073024
STD. NAME		STD REF.#			
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6550,SP6551,SP6552,SP6553,SP6554,SP6555,SP6556,SP6557				
CCC	SP6553				
Internal Standard/PEM	S12041,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF139064.D	19 Aug 2024 08:41		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF139065.D	19 Aug 2024 09:41		RC/JU	Ok,M
3	PB162549BL	PB162549BL	BF139066.D	19 Aug 2024 10:10		RC/JU	Ok
4	PB162553BL	PB162553BL	BF139067.D	19 Aug 2024 10:40		RC/JU	Ok
5	P3390-03	MDL-SOIL-03-QT3-202	BF139068.D	19 Aug 2024 11:16	MDL-SOIL-4 ppm	RC/JU	Ok,M
6	P3390-03	MDL-SOIL-03-QT3-202	BF139069.D	19 Aug 2024 11:46	MDL-SOIL-8 ppm	RC/JU	Ok,M
7	P3390-09	MDL-WATER-03-QT3-2	BF139070.D	19 Aug 2024 12:32	MDL-WATER-4 ppm	RC/JU	Ok,M
8	P3390-09	MDL-WATER-03-QT3-2	BF139071.D	19 Aug 2024 13:02	MDL-WATER-8 ppm	RC/JU	Ok,M
9	PB162612BL	PB162612BL	BF139072.D	19 Aug 2024 14:17		RC/JU	Ok
10	PB162612BS	PB162612BS	BF139073.D	19 Aug 2024 14:46		RC/JU	Ok,M
11	PB162549BS	PB162549BS	BF139074.D	19 Aug 2024 15:17		RC/JU	Ok,M
12	PB162688BS	PB162688BS	BF139075.D	19 Aug 2024 15:47		RC/JU	Ok,M
13	SSTDCCC040	SSTDCCC040EC	BF139076.D	19 Aug 2024 16:18		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF082024

Review By	Jagrut	Review On	8/21/2024 10:35:25 AM
Supervise By	mohammad	Supervise On	8/22/2024 7:52:09 AM
SubDirectory	BF082024	HP Acquire Method	BNA_F
		HP Processing Method	bf082024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588
CCC	SP6591
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF139077.D	20 Aug 2024 16:39		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BF139078.D	20 Aug 2024 17:10		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BF139079.D	20 Aug 2024 17:40	Compound #9,32,54,56,65 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BF139080.D	20 Aug 2024 18:10	Com#26 and 48 Fail for %RE Criteria in the Calibration.	RC/JU	Ok
5	SSTDICC020	SSTDICC020	BF139081.D	20 Aug 2024 18:40	Compound #26,32,40,48,51,53,57,62,92 Kept on LR and Compound #54,65 Kept on QR	RC/JU	Ok
6	SSTDICCC040	SSTDICCC040	BF139082.D	20 Aug 2024 19:10	The Calibration is Good For 8270 E NON-DOD except com#26 and 48 and good for 625.1 Method for all compounds.	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BF139083.D	20 Aug 2024 19:41		RC/JU	Ok
8	SSTDICC060	SSTDICC060	BF139084.D	20 Aug 2024 20:11		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BF139085.D	20 Aug 2024 20:41	Compound #9 removed from 80ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBF082024	BF139086.D	20 Aug 2024 21:41		RC/JU	Ok
11	PB162687BL	PB162687BL	BF139087.D	20 Aug 2024 22:10		RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF082224

Review By	yogesh	Review On	8/23/2024 6:48:24 AM
Supervise By	mohammad	Supervise On	8/23/2024 6:56:56 AM
SubDirectory	BF082224	HP Acquire Method	BNA_F
		HP Processing Method	bf082024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588
CCC	SP6591
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BF139112.D	22 Aug 2024 09:55		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BF139113.D	22 Aug 2024 10:25		RC/JU	Ok
3	PB162782BL	PB162782BL	BF139114.D	22 Aug 2024 10:54	Analyzed for contamination check	RC/JU	Not Ok
4	PB162687BS	PB162687BS	BF139115.D	22 Aug 2024 11:23		RC/JU	Ok,M
5	PB162687BSD	PB162687BSD	BF139116.D	22 Aug 2024 11:53		RC/JU	Ok,M
6	P3641-04	S-908-K1-0.5-1.0-0815	BF139117.D	22 Aug 2024 12:22		RC/JU	Ok,M
7	P3663-03	914-J-SD-0-0.5-081624	BF139118.D	22 Aug 2024 12:52		RC/JU	Ok,M
8	P3641-02	S-907-K1-0.5-1.0-0815	BF139119.D	22 Aug 2024 13:22		RC/JU	Ok,M
9	P3640-01	SP-1-COMP	BF139120.D	22 Aug 2024 13:51	Com#26 and 48 Fail for %RE Criteria in the Calibration.	RC/JU	Not Ok
10	P3641-06	S-909-K1-0.5-1.0-0815	BF139121.D	22 Aug 2024 14:20	SIM Analysis not required	RC/JU	Ok,M
11	P3641-08	S-910-K1-0-0.5-081524	BF139122.D	22 Aug 2024 14:50	FD,Sim Analysis Not Required	RC/JU	Ok,M
12	P3641-09	S-910-K1-0.5-1.0-0815	BF139123.D	22 Aug 2024 15:21	Sim Analysis not required	RC/JU	Ok,M
13	P3641-10	S-910-K1-0.5-1.0-0815	BF139124.D	22 Aug 2024 15:51	Sim Analysis not required	RC/JU	Ok,M
14	P3641-01	S-907-K1-0-0.5-081524	BF139125.D	22 Aug 2024 16:21		RC/JU	Ok,M
15	P3663-09	915-J-SD-0-0.5-081624	BF139126.D	22 Aug 2024 16:51	SIM Analysis not required	RC/JU	Ok,M
16	P3663-11	919-J-SD-0-0.5-1-0816	BF139127.D	22 Aug 2024 17:21	SIM Analysis not required	RC/JU	Ok,M
17	P3663-12	919-J-SD-0.5-1-081624	BF139128.D	22 Aug 2024 17:51		RC/JU	Ok,M

Instrument ID: BNA_F

Daily Analysis Runlog For Sequence/QC Batch ID # BF082224

Review By	yogesh	Review On	8/23/2024 6:48:24 AM		
Supervise By	mohammad	Supervise On	8/23/2024 6:56:56 AM		
SubDirectory	BF082224	HP Acquire Method	BNA_F	HP Processing Method	bf082024
STD. NAME	STD REF.#				
Tune/Reschk	SP6573				
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588				
CCC	SP6591				
Internal Standard/PEM	S12041,10ul/1000ul sample				
ICV/I.BLK	SP6559				
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

18	P3641-03	S-908-K1-0-0.5-081524	BF139129.D	22 Aug 2024 18:21	Need 5X Dilution, SIM Analysis not required	RC/JU	Dilution
19	P3660-03DL	S-902-J-0-0.5-081624D	BF139130.D	22 Aug 2024 18:50		RC/JU	Ok,M
20	P3660-04DL	S-902-J-0.5-1-081624D	BF139131.D	22 Aug 2024 19:20		RC/JU	Ok,M
21	P3663-10DL	915-J-SD-0.5-1-081624	BF139132.D	22 Aug 2024 19:50		RC/JU	Ok,M
22	P3641-05	S-909-K1-0-0.5-081524	BF139133.D	22 Aug 2024 20:20	SIM Analysis not required	RC/JU	Ok,M
23	P3641-07	S-910-K1-0-0.5-081524	BF139134.D	22 Aug 2024 20:50	Sim Analysis not required	RC/JU	Ok,M
24	P3671-12	S-903-J-SO-0.5-1.0-08	BF139135.D	22 Aug 2024 21:20		RC/JU	Ok,M

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP081324

Review By	yogesh	Review On	8/14/2024 5:49:15 AM
Supervise By	mohammad	Supervise On	8/14/2024 6:23:14 AM
SubDirectory	BP081324	HP Acquire Method	BNA_P
		HP Processing Method	BP081324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588
CCC	SP6591
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP021480.D	13 Aug 2024 11:49		RC/JU	Ok
2	SSTDICC2.5	SSTDICC2.5	BP021481.D	13 Aug 2024 12:32		RC/JU	Ok
3	SSTDICC005	SSTDICC005	BP021482.D	13 Aug 2024 13:15	Compound #32,54,56,65,70,77,85 removed from 5ppm	RC/JU	Ok,M
4	SSTDICC010	SSTDICC010	BP021483.D	13 Aug 2024 13:59		RC/JU	Ok,M
5	SSTDICC020	SSTDICC020	BP021484.D	13 Aug 2024 14:42	Compound #48,51,53,56,57,62,80,85 Kept on LR and Compound #26,32,54,65,77 Kept on QR	RC/JU	Ok,M
6	SSTDICCC040	SSTDICCC040	BP021485.D	13 Aug 2024 15:26	The Calibration is Good For 8270 DOD except com#32,77 and good for 8270E	RC/JU	Ok,M
7	SSTDICC050	SSTDICC050	BP021486.D	13 Aug 2024 16:09		RC/JU	Ok,M
8	SSTDICC060	SSTDICC060	BP021487.D	13 Aug 2024 16:53		RC/JU	Ok,M
9	SSTDICC080	SSTDICC080	BP021488.D	13 Aug 2024 17:36	Compound #9 removed from 80ppm	RC/JU	Ok,M
10	SSTDICV040	ICVBP081324	BP021489.D	13 Aug 2024 19:03	Com#32 and 77 failed in ICV for 8270 DOD Criteria	RC/JU	Ok,M
11	PB162612TB	PB162612TB	BP021490.D	13 Aug 2024 19:46	Already analyzed with ok status	RC/JU	Ok

M : Manual Integration

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP081524

Review By	Jagrut	Review On	8/16/2024 9:58:34 AM
Supervise By	mohammad	Supervise On	8/17/2024 2:51:46 AM
SubDirectory	BP081524	HP Acquire Method	BNA_P
		HP Processing Method	BP081324

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588
CCC	SP6591
Internal Standard/PEM	S12041,10ul/1000ul sample
ICV/I.BLK	SP6559
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BP021505.D	15 Aug 2024 08:28		RC/JU	Ok
2	SSTDCCC040	SSTDCCC040	BP021506.D	15 Aug 2024 09:10		RC/JU	Ok,M
3	PB162549BL	PB162549BL	BP021507.D	15 Aug 2024 09:52		RC/JU	Ok
4	PB162553BL	PB162553BL	BP021508.D	15 Aug 2024 10:34		RC/JU	Ok
5	P3390-03	MDL-SOIL-03-QT3-202	BP021509.D	15 Aug 2024 11:18	MDL-SOIL-4 ppm	RC/JU	Ok
6	P3390-03	MDL-SOIL-03-QT3-202	BP021510.D	15 Aug 2024 12:00	MDL-SOIL-8 ppm	RC/JU	Ok,M
7	P3390-09	MDL-WATER-03-QT3-2	BP021511.D	15 Aug 2024 12:43	MDL-WATER-4 ppm	RC/JU	Ok,M
8	P3390-09	MDL-WATER-03-QT3-2	BP021512.D	15 Aug 2024 13:26	MDL-WATER-8 ppm	RC/JU	Ok
9	SSTDCCC040	SSTDCCC040EC	BP021513.D	15 Aug 2024 14:09		RC/JU	Ok,M
10	DFTPP	DFTPP	BP021514.D	15 Aug 2024 14:54		RC/JU	Ok
11	SSTDCCC040	SSTDCCC040	BP021515.D	15 Aug 2024 15:37		RC/JU	Ok,M
12	PB162688BL	PB162688BL	BP021516.D	15 Aug 2024 16:19	Not Used	RC/JU	Not Ok
13	P3573-12	EB-01-080924	BP021517.D	15 Aug 2024 17:09		RC/JU	Ok
14	P3582-07	EB-01-081224	BP021518.D	15 Aug 2024 17:53		RC/JU	Ok
15	P3582-02	S-855-J-SO-1.5-2-0812	BP021519.D	15 Aug 2024 18:37		RC/JU	Ok
16	P3582-04	S-855-J-SO-3-3.5-0812	BP021520.D	15 Aug 2024 19:20		RC/JU	Ok
17	P3582-03	S-855-J-SO-2-2.5-0812	BP021521.D	15 Aug 2024 20:03		RC/JU	Ok
18	P3573-06	S-856-J-SO-1-1.5-0809	BP021522.D	15 Aug 2024 20:47		RC/JU	Ok

Instrument ID: BNA_P

Daily Analysis Runlog For Sequence/QC Batch ID # BP081524

Review By	Jagrut	Review On	8/16/2024 9:58:34 AM		
Supervise By	mohammad	Supervise On	8/17/2024 2:51:46 AM		
SubDirectory	BP081524	HP Acquire Method	BNA_P	HP Processing Method	BP081324
STD. NAME		STD REF.#			
Tune/Reschk		SP6573			
Initial Calibration Stds		SP6595,SP6594,SP6593,SP6592,SP6591,SP6590,SP6589,SP6588			
CCC		SP6591			
Internal Standard/PEM		S12041,10ul/1000ul sample			
ICV/I.BLK		SP6559			
Surrogate Standard					
MS/MSD Standard					
LCS Standard					

19	P3573-08	S-856-J-SO-2-2.5-0809	BP021523.D	15 Aug 2024 21:30		RC/JU	Ok
20	P3573-10	S-856-J-SO-4-4.5-0809	BP021524.D	15 Aug 2024 22:14		RC/JU	Ok
21	P3607-01	NB-08-08142024	BP021525.D	15 Aug 2024 22:57		RC/JU	Ok,M
22	P3607-01MS	NB-08-08142024MS	BP021526.D	15 Aug 2024 23:41		RC/JU	Ok,M
23	P3607-01MSD	NB-08-08142024MSD	BP021527.D	16 Aug 2024 00:24		RC/JU	Ok,M
24	P3579-02	VCPS-B1-081224-0-16	BP021528.D	16 Aug 2024 01:07		RC/JU	Ok,M
25	P3598-01	JC-03-081324-A	BP021529.D	16 Aug 2024 01:51		RC/JU	Ok
26	P3598-03	JC-03-081324-B	BP021530.D	16 Aug 2024 02:34		RC/JU	Ok

M : Manual Integration