

## Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P3573

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270E DataFile: BF139115.D

| Lab Sample ID | Parameter             | Spike | Result | Unit | Rec | RPD | Qual | RPD  | Limits  |           | RPD |
|---------------|-----------------------|-------|--------|------|-----|-----|------|------|---------|-----------|-----|
|               |                       |       |        |      |     |     |      | Qual | Low     | High      |     |
| PB162687BS    | Benzo(a)anthracene    | 50    | 48.7   | ug/L | 97  |     |      |      | 70 (62) | 130 (107) |     |
|               | Benzo(a)pyrene        | 50    | 52.9   | ug/L | 106 |     |      |      | 70 (72) | 130 (131) |     |
|               | Dibenz(a,h)anthracene | 50    | 55.1   | ug/L | 110 |     |      |      | 70 (78) | 130 (115) |     |