

Laboratory Control Sample/Laboratory Control Sample Duplicate Summary

SW-846

SDG No.: P3573

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270E DataFile: BF139116.D

Lab Sample ID	Parameter	Spike	Result	Unit	Rec	RPD	Qual	RPD	Limits		
								Qual	Low	High	RPD
PB162687BSD	Benzo(a)anthracene	50	49.0	ug/L	98	1			70 (62)	130 (107)	20 (20)
	Benzo(a)pyrene	50	52.1	ug/L	104	2			70 (72)	130 (131)	20 (20)
	Dibenz(a,h)anthracene	50	53.4	ug/L	107	3			70 (78)	130 (115)	20 (20)