

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3573 SAS No.: P3573 SDG NO.: P3573

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/19/2024

Lab File ID: BF139065.D Time Analyzed: 09:41

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	49558	6.839	204011	8.12	113932	9.88
UPPER LIMIT	99116	7.339	408022	8.622	227864	10.375
LOWER LIMIT	24779	6.339	102006	7.622	56966	9.375
EPA SAMPLE NO.						
01 PB162688BS	53216	6.83	219351	8.12	122326	9.87

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.