

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3573 SAS No.: P3573 SDG NO.: P3573

EPA Sample No.: SSTDICCC040 Date Analyzed: 08/20/2024

Lab File ID: BF139082.D Time Analyzed: 19:10

Instrument ID: BNA_F GC Column: DB-UI ID: 0.18 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	116386	6.899	419218	8.18	240364	9.93
UPPER LIMIT	232772	7.399	838436	8.681	480728	10.434
LOWER LIMIT	58193	6.399	209609	7.681	120182	9.434
EPA SAMPLE NO.						
01 PB162687BL	90299	6.89	335821	8.18	200816	9.93

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.