

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3573 SAS No.: P3573 SDG NO.: P3573

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/15/2024

Lab File ID: BP021515.D Time Analyzed: 15:37

Instrument ID: BNA_P GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	352497	7.805	1470750	10.59	969870	14.45
UPPER LIMIT	704994	8.305	2941500	11.093	1939740	14.945
LOWER LIMIT	176249	7.305	735375	10.093	484935	13.945
EPA SAMPLE NO.						
01 EB-01-080924	407321	7.80	1612740	10.58	1030570	14.44
02 S-856-J-SO-1-1.5-080924-FD	407787	7.80	1624320	10.59	1041840	14.45
03 S-856-J-SO-2-2.5-080924	387937	7.80	1532010	10.59	962989	14.44
04 S-856-J-SO-4-4.5-080924	358136	7.80	1355590	10.59	799134	14.44

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.