

Metals
- 5b -
POST DIGEST SPIKE SUMMARY

Client: JACOBS Engineering Group, Inc. **SDG No.:** P3596
Contract: JACO05 **Lab Code:** CHEM **Case No.:** P3596 **SAS No.:** P3596
Matrix: Water **Level:** LOW **Client ID:** 918-J-WS-081324-FDA
Sample ID: P3596-02 **Spiked ID:** P3596-02A

Analyte	Units	Acceptance Limit %R	Spiked Result	C	Sample Result	C	Spike Added	% Recovery	Qual	M
Silver	ug/L	75 - 125	376		1.00	U	500	75		P