

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN081024\
 Data File : BN033341.D
 Acq On : 10 Aug 2024 17:54
 Operator : MA/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN081024

Quant Time: Aug 12 01:13:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN081024.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Aug 12 01:09:50 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	88	0.00
2	1,4-Dioxane	0.400	0.458	-14.5	105	0.00
3	n-Nitrosodimethylamine	0.400	0.427	-6.7	96	0.00
4 S	2-Fluorophenol	0.400	0.337	15.8	73	0.00
5 S	Phenol-d6	0.400	0.329	17.8	72	0.00
6	bis(2-Chloroethyl)ether	0.400	0.446	-11.5	100	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	89	0.00
8 S	Nitrobenzene-d5	0.400	0.420	-5.0	96	0.00
9	Naphthalene	0.400	0.427	-6.7	98	0.00
10	Hexachlorobutadiene	0.400	0.419	-4.7	94	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.413	-3.2	95	0.00
12	2-Methylnaphthalene	0.400	0.435	-8.7	101	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	89	0.00
14 S	2,4,6-Tribromophenol	0.400	0.324	19.0	77	0.00
15 S	2-Fluorobiphenyl	0.400	0.443	-10.7	99	0.00
16	Acenaphthylene	0.400	0.437	-9.2	105	0.00
17	Acenaphthene	0.400	0.431	-7.7	101	0.00
18	Fluorene	0.400	0.422	-5.5	99	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	89	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.379	5.3	98	-0.01
21	4-Bromophenyl-phenylether	0.400	0.424	-6.0	100	0.00
22	Hexachlorobenzene	0.400	0.436	-9.0	100	0.00
23	Atrazine	0.400	0.378	5.5	93	0.00
24	Pentachlorophenol	0.400	0.334	16.5	88	0.00
25	Phenanthrene	0.400	0.448	-12.0	104	0.00
26	Anthracene	0.400	0.442	-10.5	105	0.00
27 SURR	Fluoranthene-d10	0.400	0.400	0.0	94	0.00
28	Fluoranthene	0.400	0.419	-4.7	99	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	92	0.00
30	Pyrene	0.400	0.429	-7.2	99	0.00
31 S	Terphenyl-d14	0.400	0.462	-15.5	106	0.00
32	Benzo(a)anthracene	0.400	0.427	-6.7	105	0.00
33	Chrysene	0.400	0.457	-14.2	105	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.362	9.5	96	0.00
35 I	Perylene-d12	0.400	0.400	0.0	87	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.428	-7.0	103	0.00
37	Benzo(b)fluoranthene	0.400	0.429	-7.2	99	0.00
38	Benzo(k)fluoranthene	0.400	0.458	-14.5	103	0.00
39 C	Benzo(a)pyrene	0.400	0.447	-11.7	107	0.00
40	Dibenzo(a,h)anthracene	0.400	0.428	-7.0	103	0.00
41	Benzo(g,h,i)perylene	0.400	0.394	1.5	93	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN081024