

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BN081724

SequenceID : BN081724

	NA	NO	YES
1. Chromatograms Labeled/Compounds Identified. (Field samples and Method Blanks)	_____	_____	_____✓_____
2. GC/MS Tuning Specifications. DFTPP Meet Criteria Criteria (NOTE THAT THERE ARE DIFFERENT CRITERIA FOR NY ASP CLP, CLP AND NJ)	_____	_____	_____✓_____
3. GC/MS Tuning Frequency - Performed every 24 hours for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
4. GC/MS Calibration - Initial Calibration performed within 30 days before sample analysis and continuing calibration performed within 24 hours of sample analysis for 600 series and 12 hours for 8000 series	_____	_____	_____✓_____
5. GC/MS Calibration Met:	_____	_____	_____✓_____
a. Initial calibration Meet Criteria	_____	_____	_____✓_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____	_____✓_____
b. Continuous Calibration(CCC) Meet Criteria	_____	_____✓_____	_____
If not met, list those compounds and their recoveries which fall outside the acceptable range.	_____	_____✓_____	_____
Compounds #24,34 are biased high in the SSTDCCC(BN033428.D) If any samples are found with hit of these compounds they will be re-analyzed with a properly passing CCC.	_____	_____✓_____	_____
6. Blank Contamination - If yes, list compounds and concentrations in each blank:	_____	_____✓_____	_____
a. B/N Fraction	_____	_____✓_____	_____
d. Acid Fraction	_____	_____✓_____	_____

7. Surrogate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable ranges.

_____ ✓ _____

a. B/N Fraction Surrogate Fail in P3580-04

d. Acid Fraction

8. Matrix Spike/Matrix Spike Duplicate Recoveries Meet Criteria

If not met, list those compounds and their recoveries which fall outside the acceptable range.

_____ ✓ _____

a. B/N Fraction 1,4-Dioxane being poor recovery compound, was biased low in P3440-02MS/ P3440-03MSD and no corrective action is needed.

d. Acid Fraction

9. Internal Standard Area/Retention Time Shift Meet Criteria

Comments:

_____ ✓ _____

10. Extraction Holding Time Met

If not met, list number of days exceeded for each sample:

_____ ✓ _____

11. Analysis Holding Time Met

If not met, list number of days exceeded for each sample:

_____ ✓ _____

ADDITIONAL COMMENTS:

Jagrut
Analyst

08/16/2024
Date

NEW JERSEY LAB ID#:20012 : NEW YORK LAB ID#: 11376

GC/MS SEMI-VOLATILE ANALYSIS CONFORMANCE/NON-CONFORMANCE SUMMARY

CHEMTECH PROJECT NUMBER: BN081724

SequenceID : BN081724 NA NO YES