

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN081024

Review By	yogesh	Review On	8/13/2024 5:11:28 AM
Supervise By	mohammad	Supervise On	8/13/2024 5:17:38 AM
SubDirectory	BN081024	HP Acquire Method	BNA_N, 8270_HP Processing Method bn081024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597
CCC	SP6601
Internal Standard/PEM	SP6527
ICV/I.BLK	SP6548
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN033333.D	10 Aug 2024 12:56		MA/JU	Ok
2	SSTDICC0.1	SSTDICC0.1	BN033334.D	10 Aug 2024 13:40		MA/JU	Ok
3	SSTDICC0.2	SSTDICC0.2	BN033335.D	10 Aug 2024 14:16		MA/JU	Ok
4	SSTDICCC0.4	SSTDICCC0.4	BN033336.D	10 Aug 2024 14:53	Method is Good For DOD	MA/JU	Ok
5	SSTDICC0.8	SSTDICC0.8	BN033337.D	10 Aug 2024 15:29		MA/JU	Ok
6	SSTDICC1.6	SSTDICC1.6	BN033338.D	10 Aug 2024 16:05		MA/JU	Ok
7	SSTDICC3.2	SSTDICC3.2	BN033339.D	10 Aug 2024 16:41		MA/JU	Ok
8	SSTDICC5.0	SSTDICC5.0	BN033340.D	10 Aug 2024 17:17		MA/JU	Ok
9	SSTDICV0.4	ICVBN081024	BN033341.D	10 Aug 2024 17:54		MA/JU	Ok,M
10	PB162604BL	PB162604BL	BN033342.D	10 Aug 2024 19:06		MA/JU	Ok

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN081724

Review By	yogesh	Review On	8/17/2024 4:59:53 AM
Supervise By	mohammad	Supervise On	8/21/2024 1:23:52 AM
SubDirectory	BN081724	HP Acquire Method	BNA_N, 8270_HP Processing Method BN081024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597
CCC	SP6601
Internal Standard/PEM	SP6527
ICV/I.BLK	SP6548
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleId	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN033427.D	16 Aug 2024 12:55		MA/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN033428.D	16 Aug 2024 13:34		MA/JU	Ok
3	PB162692BL	PB162692BL	BN033429.D	16 Aug 2024 14:10		MA/JU	Ok
4	P3440-02MS	923-K1-WS-080124MS	BN033430.D	16 Aug 2024 14:50		MA/JU	Ok
5	P3440-03MSD	923-K1-WS-080124MS	BN033431.D	16 Aug 2024 15:26		MA/JU	Ok
6	P3430-01	927-K1-WPO-0-0.25-07	BN033432.D	16 Aug 2024 16:02		MA/JU	Ok,M
7	P3430-02	927-K1-WPO-0-0.25-07	BN033433.D	16 Aug 2024 16:39		MA/JU	Ok,M
8	P3430-03	926-K1-WPO-0-0.25-07	BN033434.D	16 Aug 2024 17:15		MA/JU	Ok
9	P3430-04	931-K1-WPO-0-0.25-07	BN033435.D	16 Aug 2024 17:51		MA/JU	Ok
10	P3430-05	925-K1-WPO-0-0.25-07	BN033436.D	16 Aug 2024 18:27		MA/JU	Ok
11	P3450-01	921-J-WPO-0.25-08012	BN033437.D	16 Aug 2024 19:03		MA/JU	Ok,M
12	P3450-02	923-K1-WPO-0.25-0801	BN033438.D	16 Aug 2024 19:39		MA/JU	Ok,M
13	P3450-03	922-K1-WPO-0.25-0801	BN033439.D	16 Aug 2024 20:15		MA/JU	Ok,M
14	P3466-01	924-K1-WPO-0.25-0802	BN033440.D	16 Aug 2024 20:52		MA/JU	Ok,M
15	P3466-04	932-K1-WPO-0.25-0802	BN033441.D	16 Aug 2024 21:28		MA/JU	Ok
16	P3466-05	919-J-WPO-0.25-08022	BN033442.D	16 Aug 2024 22:04		MA/JU	Ok,M
17	P3620-01	915-J-WPO-0.25-08142	BN033443.D	16 Aug 2024 22:40		MA/JU	Ok
18	P3620-02	920-J-WPO-0.25-08142	BN033444.D	16 Aug 2024 23:16		MA/JU	Ok

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN081724

Review By	yogesh	Review On	8/17/2024 4:59:53 AM
Supervise By	mohammad	Supervise On	8/21/2024 1:23:52 AM
SubDirectory	BN081724	HP Acquire Method	BNA_N, 8270_HP Processing Method BN081024
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597		
CCC	SP6601		
Internal Standard/PEM	SP6527		
ICV/I.BLK	SP6548		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

19	SSTDCCC0.4	SSTDCCC0.4EC	BN033445.D	16 Aug 2024 23:53		MA/JU	Ok
20	DFTPP	DFTPP	BN033446.D	17 Aug 2024 01:08		MA/JU	Ok
21	SSTDCCC0.4	SSTDCCC0.4	BN033447.D	17 Aug 2024 01:48		MA/JU	Ok,M
22	PB162758BL	PB162758BL	BN033448.D	17 Aug 2024 02:24		MA/JU	Ok
23	P3601-02	918-J-WPO-0.25-08132	BN033449.D	17 Aug 2024 03:01		MA/JU	Ok
24	P3580-02	926-KI-SD-0.5-1-08122	BN033450.D	17 Aug 2024 03:37		MA/JU	Ok,M
25	P3580-04	925-KI-SD-0.5-1-08122	BN033451.D	17 Aug 2024 04:14	Surrogate Fail	MA/JU	ReRun
26	P3580-08	922-KI-SD-0.5-1-08122	BN033452.D	17 Aug 2024 04:49		MA/JU	Ok
27	P3580-03	925-KI-SD-0.5-08122	BN033453.D	17 Aug 2024 05:26		MA/JU	Ok,M
28	P3609-02	920-J-WS-081424	BN033454.D	17 Aug 2024 06:02		MA/JU	Ok
29	P3580-06	923-KI-SD-0.5-1-08122	BN033455.D	17 Aug 2024 06:38		MA/JU	Ok,M
30	P3580-07	922-KI-SD-0.5-08122	BN033456.D	17 Aug 2024 07:15		MA/JU	Ok,M
31	P3580-05	923-KI-SD-0.5-08122	BN033457.D	17 Aug 2024 07:51		MA/JU	Ok,M
32	P3596-01	918-J-WS-081324	BN033458.D	17 Aug 2024 08:27		MA/JU	Ok,M
33	P3596-02	918-J-WS-081324-FD	BN033459.D	17 Aug 2024 09:04		MA/JU	Ok,M
34	P3601-01	918-J-WPO-0.25-08132	BN033460.D	17 Aug 2024 09:40		MA/JU	Ok
35	PB162758BS	PB162758BS	BN033461.D	17 Aug 2024 10:16		MA/JU	Ok,M
36	PB162758BSD	PB162758BSD	BN033462.D	17 Aug 2024 10:52		MA/JU	Ok,M
37	SSTDCCC0.4	SSTDCCC0.4EC	BN033463.D	17 Aug 2024 11:28		MA/JU	Ok,M

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN081724

Review By	yogesh	Review On	8/17/2024 4:59:53 AM
Supervise By	mohammad	Supervise On	8/21/2024 1:23:52 AM
SubDirectory	BN081724	HP Acquire Method	BNA_N, 8270_HP Processing Method BN081024
STD. NAME	STD REF.#		
Tune/Reschk	SP6573		
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597		
CCC	SP6601		
Internal Standard/PEM	SP6527		
ICV/I.BLK	SP6548		
Surrogate Standard			
MS/MSD Standard			
LCS Standard			

M : Manual Integration

Instrument ID: BNA_N

Daily Analysis Runlog For Sequence/QC Batch ID # BN081824

Review By	Jagrut	Review On	8/19/2024 9:18:56 AM
Supervise By	mohammad	Supervise On	8/21/2024 1:24:00 AM
SubDirectory	BN081824	HP Acquire Method	BNA_N, 8270_HP Processing Method BN081024

STD. NAME	STD REF.#
Tune/Reschk	SP6573
Initial Calibration Stds	SP6603,SP6602,SP6601,SP6600,SP6599,SP6598,SP6597
CCC	SP6601
Internal Standard/PEM	SP6527
ICV/I.BLK	SP6548
Surrogate Standard	
MS/MSD Standard	
LCS Standard	

Sr#	SampleID	ClientID	Data File Name	Date-Time	Comment	Operator	Status
1	DFTPP	DFTPP	BN033464.D	17 Aug 2024 12:05		MA/JU	Ok
2	SSTDCCC0.4	SSTDCCC0.4	BN033465.D	17 Aug 2024 12:44		MA/JU	Ok,M
3	PB162723BL	PB162723BL	BN033466.D	17 Aug 2024 13:56		MA/JU	Ok
4	PB162723BS	PB162723BS	BN033467.D	17 Aug 2024 14:32		MA/JU	Ok,M
5	PB162723BSD	PB162723BSD	BN033468.D	17 Aug 2024 15:09		MA/JU	Ok,M
6	PB162692BS	PB162692BS	BN033469.D	17 Aug 2024 15:45		MA/JU	Ok,M
7	P3619-01	917-J-WS-081424	BN033470.D	17 Aug 2024 16:21	Sample Removed From login Page	MA/JU	Not Ok
8	P3609-01	915-J-WS-081424	BN033471.D	17 Aug 2024 16:57		MA/JU	Ok,M
9	P3580-01	926-KI-SD-0-0.5-08122	BN033472.D	17 Aug 2024 17:33	Need 5X Dilution	MA/JU	Dilution
10	P3580-01MS	926-KI-SD-0-0.5-08122	BN033473.D	17 Aug 2024 18:09	Recovery Fail for Many Compound	MA/JU	Ok,M
11	P3580-01MSD	926-KI-SD-0-0.5-08122	BN033474.D	17 Aug 2024 18:45	Recovery Fail for Many Compound	MA/JU	Ok,M
12	SSTDCCC0.4	SSTDCCC0.4EC	BN033475.D	17 Aug 2024 19:22		MA/JU	Ok,M

M : Manual Integration