

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3596 SAS No.: P3596 SDG NO.: P3596

EPA Sample No.: SSTDCCC040 Date Analyzed: 08/21/2024

Lab File ID: BM047290.D Time Analyzed: 09:44

Instrument ID: BNA_M GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	282815	7.357	1065920	10.11	694208	14.02
UPPER LIMIT	565630	7.857	2131840	10.61	1388420	14.516
LOWER LIMIT	141408	6.857	532960	9.61	347104	13.516
EPA SAMPLE NO.						
01 918-J-WS-081324	183659	7.35	677846	10.11	451328	14.01
02 918-J-WS-081324-FD	190222	7.35	707486	10.11	485993	14.02

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.