

VOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH Contract: JACO05
 Lab Code: CHEM Case No.: P3596 SAS No.: P3596 SDG NO.: P3596
 Lab File ID: VN083295.D Date Analyzed: 08/14/2024
 Instrument ID: MSVOA_N Time Analyzed: 12:41
 GC Column: RXI-624 ID: 0.25 (mm) Heated Purge: (Y/N) N

	IS4 AREA #	RT #				
12 HOUR STD	122669	13.794				
UPPER LIMIT	245338	14.294				
LOWER LIMIT	61334.5	13.294				
EPA SAMPLE NO.						
918-J-WS-081324	116254	13.79				
918-J-WS-081324-FD	107123	13.79				
TB-01-081324	110441	13.79				
VN0814WBL01	128230	13.79				
VN0814WBS01	117869	13.79				
VN0814WBSD01	107371	13.79				

IS4 = 1,4-Dichlorobenzene-d4

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT LOWER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.