

Surrogate Summary

SDG No.: **P3596**

Client: **JACOBS Engineering Group, Inc.**

Analytical Method: **SW8260-Low**

Lab Sample ID	Client ID	Parameter	Spike	Result	RecoveryQual	Limits	
						Low	High
P3596-01	918-J-WS-081324	1,2-Dichloroethane-d4	50	59.1	118	70 (74)	130 (125)
		Dibromofluoromethane	50	54.5	109	70 (75)	130 (124)
		Toluene-d8	50	53.9	108	70 (86)	130 (113)
		4-Bromofluorobenzene	50	57.9	116	70 (77)	130 (121)
P3596-02	918-J-WS-081324-FD	1,2-Dichloroethane-d4	50	59.4	119	70 (74)	130 (125)
		Dibromofluoromethane	50	54.6	109	70 (75)	130 (124)
		Toluene-d8	50	54.3	109	70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.0	112	70 (77)	130 (121)
P3596-03	TB-01-081324	1,2-Dichloroethane-d4	50	57.1	114	70 (74)	130 (125)
		Dibromofluoromethane	50	53.8	108	70 (75)	130 (124)
		Toluene-d8	50	53.5	107	70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.3	113	70 (77)	130 (121)
VN0814WBL01	VN0814WBL01	1,2-Dichloroethane-d4	50	55.6	111	70 (74)	130 (125)
		Dibromofluoromethane	50	53.1	106	70 (75)	130 (124)
		Toluene-d8	50	54.2	108	70 (86)	130 (113)
		4-Bromofluorobenzene	50	56.3	113	70 (77)	130 (121)
VN0814WBS01	VN0814WBS01	1,2-Dichloroethane-d4	50	56.3	113	70 (74)	130 (125)
		Dibromofluoromethane	50	54.0	108	70 (75)	130 (124)
		Toluene-d8	50	53.1	106	70 (86)	130 (113)
		4-Bromofluorobenzene	50	52.3	105	70 (77)	130 (121)
VN0814WBSD0	VN0814WBSD01	1,2-Dichloroethane-d4	50	58.7	117	70 (74)	130 (125)
		Dibromofluoromethane	50	54.0	108	70 (75)	130 (124)
		Toluene-d8	50	53.0	106	70 (86)	130 (113)
		4-Bromofluorobenzene	50	53.0	106	70 (77)	130 (121)

