

Surrogate Summary

SW-846

SDG No.: P3609

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270-Modified

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P3609-01	915-J-WS-081424	2-Methylnaphthalene-d10	0.4	0.28	69		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.37	91		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.35	88		30 (27)	130 (123)
		2-Fluorobiphenyl	0.4	0.33	81		30 (34)	130 (132)
		Terphenyl-d14	0.4	0.65	163	*	30 (35)	130 (157)
P3609-02	920-J-WS-081424	2-Methylnaphthalene-d10	0.4	0.27	67		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.35	88		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.33	83		30 (27)	130 (123)
		2-Fluorobiphenyl	0.4	0.29	73		30 (34)	130 (132)
		Terphenyl-d14	0.4	0.49	122		30 (35)	130 (157)
PB162758BL	PB162758BL	2-Methylnaphthalene-d10	0.4	0.30	76		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.31	77		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.39	96		30 (27)	130 (123)
		2-Fluorobiphenyl	0.4	0.35	88		30 (34)	130 (132)
		Terphenyl-d14	0.4	0.44	109		30 (35)	130 (157)
PB162758BS	PB162758BS	2-Methylnaphthalene-d10	0.4	0.33	82		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.32	80		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.40	100		30 (27)	130 (123)
		2-Fluorobiphenyl	0.4	0.38	94		30 (34)	130 (132)
		Terphenyl-d14	0.4	0.39	97		30 (35)	130 (157)
PB162758BSD	PB162758BSD	2-Methylnaphthalene-d10	0.4	0.32	81		30 (20)	150 (139)
		Fluoranthene-d10	0.4	0.32	79		30 (30)	150 (150)
		Nitrobenzene-d5	0.4	0.40	100		30 (27)	130 (123)
		2-Fluorobiphenyl	0.4	0.38	95		30 (34)	130 (132)
		Terphenyl-d14	0.4	0.39	98		30 (35)	130 (157)