

8B

SEMIVOLATILE INTERNAL STANDARD AREA AND RT SUMMARY

Lab Name: CHEMTECH

Lab Code: CHEM Case No.: P3609 SAS No.: P3609 SDG NO.: P3609

EPA Sample No.: SSTDCCC0.4 Date Analyzed: 08/17/2024

Lab File ID: BN033447.D Time Analyzed: 01:48

Instrument ID: BNA_N GC Column: ZB-GR ID: 0.25 (mm)

	IS1 (DCB) AREA #	RT #	IS2 (NPT) AREA #	RT #	IS3 (ANT) AREA #	RT #
12 HOUR STD	8031	7.566	21798	10.33	11010	14.20
UPPER LIMIT	16062	8.066	43596	10.825	22020	14.699
LOWER LIMIT	4015.5	7.066	10899	9.825	5505	13.699
EPA SAMPLE NO.						
01 PB162758BL	7577	7.57	19459	10.33	7966	14.20
02 PB162758BS	10423	7.56	28103	10.33	13134	14.19
03 PB162758BSD	10549	7.56	28597	10.33	13216	14.19
04 920-J-WS-081424	7674	7.56	20880	10.33	10506	14.20

IS1 (DCB) = 1,4-Dichlorobenzene-d4

IS2 (NPT) = Naphthalene-d8

IS3 (ANT) = Acenaphthene-d10

AREA UPPER LIMIT = +100% of internal standard area

AREA LOWER LIMIT = -50% of internal standard area

RT UPPER LIMIT = +0.50 minutes of internal standard RT

RT UPPER LIMIT = -0.50 minutes of internal standard RT

Column used to flag values outside QC limits with an asterisk.

* Values outside of QC limits.