

Surrogate Summary

SW-846

SDG No.: P3609

Client: JACOBS Engineering Group, Inc.

Analytical Method: 8270E

Lab Sample ID	Client ID	Parameter	Spike (PPM)	Result (PPM)	Recovery (%)	Qual	Limits (%)	
							Low	High
P3609-01	915-J-WS-081424	2-Fluorophenol	150	55.5	37		15 (10)	110 (139)
		Phenol-d6	150	34.7	23		15 (10)	110 (134)
		Nitrobenzene-d5	100	68.0	68		30 (49)	130 (133)
		2-Fluorobiphenyl	100	70.3	70		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	155	103		15 (44)	110 (137)
		Terphenyl-d14	100	86.4	86		30 (48)	130 (125)
P3609-02	920-J-WS-081424	2-Fluorophenol	150	57.3	38		15 (10)	110 (139)
		Phenol-d6	150	34.2	23		15 (10)	110 (134)
		Nitrobenzene-d5	100	77.0	77		30 (49)	130 (133)
		2-Fluorobiphenyl	100	80.0	80		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	175	117	*	15 (44)	110 (137)
		Terphenyl-d14	100	92.1	92		30 (48)	130 (125)
PB162757BL	PB162757BL	2-Fluorophenol	150	120	80		15 (10)	110 (139)
		Phenol-d6	150	111	74		15 (10)	110 (134)
		Nitrobenzene-d5	100	75.0	75		30 (49)	130 (133)
		2-Fluorobiphenyl	100	79.1	79		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	131	88		15 (44)	110 (137)
		Terphenyl-d14	100	96.4	96		30 (48)	130 (125)
PB162757BS	PB162757BS	2-Fluorophenol	150	134	89		15 (10)	110 (139)
		Phenol-d6	150	131	87		15 (10)	110 (134)
		Nitrobenzene-d5	100	78.7	79		30 (49)	130 (133)
		2-Fluorobiphenyl	100	80.7	81		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	141	94		15 (44)	110 (137)
		Terphenyl-d14	100	95.4	95		30 (48)	130 (125)
PB162757BSD	PB162757BSD	2-Fluorophenol	150	135	90		15 (10)	110 (139)
		Phenol-d6	150	131	87		15 (10)	110 (134)
		Nitrobenzene-d5	100	78.4	78		30 (49)	130 (133)
		2-Fluorobiphenyl	100	81.7	82		30 (52)	130 (132)
		2,4,6-Tribromophenol	150	142	95		15 (44)	110 (137)
		Terphenyl-d14	100	94.7	95		30 (48)	130 (125)